

Delivery receipt



Customer no.
150053065

Number of report
305877840
Please reference on inquiries

Date of receiving Delivery receipt date
/ / 03/05/2021

Consignee, 150053065
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: VIC BIALOUS
PO: 002080509
710 RANOCAS RD
MOUNT HOLLY NJ 08060-5642

Customer, 150053065
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
PO: 002080509
710 RANOCAS RD
MOUNT HOLLY NJ 08060-5642

Your order

Date of order:
02/19/2021
Purch.ord.no.:
002080509

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	Simulator,CU34 110V elec ctr	4540083	1020916813	DDUJS3-0116
001	Simulator, temperature probe	4412025	1024735085	DDSF2-0433

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150053065
Number of report 305877840
Date of receiving 02/20/2021
Please reference on inquiries

Customer
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
710 RANOCAS RD
MOUNT HOLLY NJ 08060-5642

Consignee
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
710 RANOCAS RD
MOUNT HOLLY NJ 08060-5642

Your order

Date of order:
02/19/2021
Your reference:
002080509

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 144398605
		COE008387
		Eq.No. : 1020916813 SN: DDUJS3-0116
		Material: 4540083 Simulator,CU34 110V elec ctrl w/o tmeter
		Inv.No. :
		Location:
		COE008387
0.0	H	R001 Repair time - normal hours
1	EA	4412034 Hose pump-sim w/sim-cuvette hose 2" KIT

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Westampton PD 710 Rancocas Rd., Westampton, NJ 08060	2. Contact: Lt. Bialous 3. Phone Number: 609-267-3000	4. Date: 2/14/2020 5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0025	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
I received a call from the Westampton PD, advising that during a breath test, an error code was displayed on the screen, "error code 112".

10. Comments:
The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

****Unable to download data due to error code, (code #112), please contact to download remotely or send back with original data to be downloaded properly. Thank you.**

**Tpr. I Dave Napolitano #7237
609-203-8361
LPP7237@GW.NJSP.ORG**

Seq. File #953 was last data line from Alcotest Inquiry System.

Last Known Sequential File #: 953


- ☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- ☒ 12. The above Instrument/Component is placed out of service pending further evaluation.
- ☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dave Napolitano #7237
Name & Badge Number (Print)

 **7237** **2.14.20**
Signature Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Westampton PD 710 Rancocas Rd., Westampton, NJ 08060		2. Contact: Lt. Bialous 3. Phone Number: 609-267-3000	4. Date: 7/30/2020 5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0025	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	
9. Reason for Service: The alcotest instrument was returned from outside evaluation and placed back in service.			
10. Comments: See Draeger Return and Repair form.			
Last Known Sequential File #: N/A			
<input type="checkbox"/> 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.			
<input type="checkbox"/> 12. The above Instrument/Component is placed out of service pending further evaluation.			
<input checked="" type="checkbox"/> 13. The above Instrument/Component is placed back in service.			
<p>"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."</p>			
Tpr. I Dave Napolitano #7237 Name & Badge Number (Print)		Signature 	Date 7-30-20

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Westampton PD, Westampton, NJ

Date Received: 04/09/2020

Date given to service: 04/09/2020

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR NK-0025

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1.5 ^{hr}	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and geared engine.

FUEL CELL = ARNC-0321

CAL W/QC AND OPS CHECK

Service Technician 

Date: 07/21/2020

Delivery receipt



Customer no.
150053065

Number of report
304963862

Please reference on inquiries

Date of receiving Delivery receipt date
/ / 07/22/2020

Consignee, 150053065

**TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
PO#ARNK-0025
710 RANOCAS ROAD
MOUNT HOLLY NJ 08060-5642**

Customer, 150053065

**TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
PO#ARNK-0025
710 RANOCAS ROAD
MOUNT HOLLY NJ 08060-5642**

Your order

Date of order:
02/24/2020
Purch.ord.no.:
ARNK-0025

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030222432	ARNK-0025
	REPAIR/ARNK-0025			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no.
150053065

Number of report
304963862

Date of receiving

Please reference on inquiries

Customer

TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
710 RANCOCAS ROAD
MOUNT HOLLY NJ 08060-5642

Consignee

TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
710 RANCOCAS ROAD
MOUNT HOLLY NJ 08060-5642

Your order

Date of order:
02/24/2020

Your reference:
ARNK-0025

NL-Text

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142602807
		REPAIR/ARNK-0025
		Eq.No. : 1030222432 SN: ARNK-0025
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ARNK-0025
	1.5 H	R001 Repair time - normal hours
	1 EA	6808455 Sensor Alcotest B
	1 EA	6808374 Geared Engine, A7110/A9510

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Westampton Police Department 710 Rancocas Road Westampton, NJ 08060	2. Contact: Sgt. Ent 3. Phone Number: 609-267-3000	4. Date: 02-14-2014 5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0025	7. Simulator Component Serial Number:	8. Temperature Probe Component Serial Number:

9. Reason for Service:
Complaint received of instrument experiencing repeated control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 586

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

A/Sgt. M. Goncalves #6040
Name & Badge Number (Print)

A/Sgt. M. Goncalves #6040 2/14/14
Signature Date

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Westampton Twp P.D., Westampton, NJ

Date Received: 03/18/14

Date given to service: 03/18/14

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR NK-0025

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: Oct. 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFA-1575

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 04/03/14

Packing Slip

Customer no Order No Order date
150053065 10691708 04/03/2014

Packing slip no. Ship date
80848738 04/03/2014

Please reference on inquiries

ship to

150053065
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF
BURLINGTON COUNTY
710 RANOCAS ROAD
MOUNT HOLLY, NJ 08060
USA

bill to

Your Purchase Order Number

SGT. ENT
Any questions? Please contact:
CHARLES ROBINSON II

Your sales person

Don Pouliot
Phone: 973-398-3228

Customer

TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF
BURLINGTON COUNTY
710 RANOCAS ROAD
MOUNT HOLLY, NJ 08060
USA

Delivery terms

FA
FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARNK-0025 RECEIVED 03/18/2014			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Draeger Safety Diagnostics Inc.
Accounting Address:
101 Technology Drive
Pittsburgh, PA 15275
Tel: (412) 788-5537
Fax: (412) 788-5598

Remit Wire Transfers To:
Citizens Bank
Acct. # 6209426615
Acct. Name: Draeger Safety Diagnostics Inc.
Transit Number: 036076150
Federal ID Number: 84-1600159

REMIT TO:
Draeger Safety Diagnostics Inc.
P.O. Box 200337
Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053065 10691708 04/03/2014

Packing slip no. Ship date
80848738 04/03/2014

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Westampton Police Department 710 Rancocas Road Westampton, NJ 08060		2. Contact: Sgt. Ent	4. Date: 04-14-2014
		3. Phone Number: 609-267-3000	5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0025	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.

10. Comments:

See attached packing slip and Drager Return and Repair Form.

Last Known Sequential File #: N/A

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☐ 12. The above Instrument/Component is placed out of service pending further evaluation.

☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

A/Sgt. M. Goncalves #6040
Name & Badge Number (Print)

Signature

Date

Packing Slip

Customer no Order No Order date
150053065 10353260 07/08/2009

Packing slip no. Ship date
80454006 07/08/2009

Please reference on inquiries

ship to

150053065
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF
710 RANOCAS ROAD
BURLINGTON COUNTY
MOUNT HOLLY, NJ 08060
USA

bill to

Your Purchase Order Number

7110 WARRANTY
Any questions? Please contact:
MARA HERRERA

Customer

TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF
710 RANOCAS ROAD
BURLINGTON COUNTY
MOUNT HOLLY, NJ 08060
USA

Your sales person

Don Pouliot
Phone: 973-398-3228

Delivery terms

FA
FREIGHT ALLOWED

Page 1 / 2

Number of packages: 1

Item#	sh. Quant.	Part no. Description			
		Waybill number: 30806095200000029740981622022012 Ship via: Fed Ex P1			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARNK-0025			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
Accounting Address:
101 Technology Drive
Pittsburgh, PA 15275
Tel: (412) 788-5537
Fax: (412) 788-5598

Remit Wire Transfers To:
Citizens Bank
Acct. # 6209426615
Acct. Name: Draeger Safety Diagnostics Inc.
Transit Number: 036076150
Federal ID Number: 84-1600159

REMIT TO:
Draeger Safety Diagnostics Inc.
P.O. Box 200337
Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053065 10353260 07/08/2009

Packing slip no. Ship date
80454006 07/08/2009

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part. no. Description			
0020	1 EA	6808486 PLATES, 7110 Ordered / Back ordered : 0 / 0 EA			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Return & Repair Form:

Customer Information:

Company Name: WEST HAMPTON B: _____ S: _____
 Date Received: 7-6 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR NK 0025
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: <u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w/WHV</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C w/WHV</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C w/WHV</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C w/WHV</u>

Repair Notes:

REPLACE FC + 1 PLATE
CAL w/DC & OPS Check

Service Technician: [Signature]

Date: 7-08-09

DRAEGER SAFETY DIAGNOSTICS, INC.

CERTIFICATE

I, Dori L. Mansur Ratka, Esq., an attorney-at-law duly admitted, inter alia, to the bar of the State of New Jersey, under penalty of perjury under the laws of the United States of America and pursuant to 28 U.S.C. § 1746, hereby certify to the appropriate court officers of the State of New Jersey, on behalf of my client, Draeger Safety Diagnostics, Inc., a Delaware corporation ("Draeger"), and not in my individual capacity, that:

1. With respect to the various requests for records reflecting repairs performed on a particular device, the Repair & Return Form constitutes Draeger's record of the actual repair work conducted on the Alcotest® 7110 MK IIIC device.

2. The Alcotest® 7110 MK IIIC Checklist, Configuration - Calibration - QC, constitutes Draeger's internal record of the re-calibration and quality check conducted by Draeger on the device following the completion of the repair work. This Checklist is a standardized, step-by-step list of calibration and quality tasks undertaken by Draeger in order to support the "Certificate of Accuracy" issued with the repaired device.

3. The instrument print-outs, including the small pieces of paper titled "Control Check" or "Breath Test" and the "Calibrating Unit, New Standard Solution Report", are documents generated during the calibration and quality check, and are used for internal verification of the completion of the various steps on the above Checklist.

4. Once the device is recalibrated, which Draeger understands to be the standard practice of the State of New Jersey upon receipt of a device repaired by Draeger, the calibration and quality checks performed on the device while in Draeger's possession, including the various attached print outs, are rendered immaterial and superseded by the subsequent State calibration process.

IN WITNESS WHEREOF, the undersigned has executed this Certificate on the 25th day of October, 2011 in the Borough of Telford, Commonwealth of Pennsylvania.



Dori L. Mansur Ratka, Esq.
on behalf of
Draeger Safety Diagnostics, Inc.



Delivery receipt



Customer no.
150053065

Number of report
306528808
Please reference on inquiries

Date of receiving Delivery receipt date
/ / 12/13/2021

Consignee, 150053065

TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: SGT. RYAN BIERI
PO: DDUJS3-0117
710 RANCOCAS RD
MOUNT HOLLY NJ 08060-5642

Customer, 150053065

TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: SGT. RYAN BIERI
PO: DDUJS3-0117
710 RANCOCAS RD
MOUNT HOLLY NJ 08060-5642

Your order

Date of order:
10/26/2021
Purch.ord.no.:
DDUJS3-0117

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	Simulator,CU34 110V elec ctr	4540083	1020916815	DDUJS3-0117
001	Simulator, temperature probe	4412025	1021847980	DDWAP2-116

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150053065 Number of report 306528808 Date of receiving 10/26/2021

Please reference on inquiries

Customer
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
710 RANOCAS RD
MOUNT HOLLY NJ 08060-5642

Consignee
TOWNSHIP OF WESTAMPTON POLICE DEPT
ATTN: CHIEF STEPHEN ENT
710 RANOCAS RD
MOUNT HOLLY NJ 08060-5642

Your order

Date of order:
10/26/2021
Your reference:
DDUJS3-0117

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 145682982
		CALIBRATION DDUJS3-0117
		Eq.No. : 1020916815 SN: DDUJS3-0117
		Material: 4540083 Simulator,CU34 110V elec ctrl w/o tmeter
		Inv.No. :
		Location:
		CALIBRATION DDUJS3-0117
		CALIBRATION DDUJS3-0117
0.0	H	R001 Repair time - normal hours
1	EA	4412034 Hose pump-sim w/sim-cuvette hose 2" KIT
<hr/>		
PAPERWORK		