

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Westampton Police Department 710 Rancocas Road Westampton, NJ 08060	2. Contact: Sgt. Ent 3. Phone Number: 609-267-3000	4. Date: 02-14-2014 5. County: Burlington
--	---	--

6. Alcotest Instrument Serial Number: ARNK-0025	7. Simulator Component Serial Number:	8. Temperature Probe Component Serial Number:
---	---------------------------------------	---

9. Reason for Service:
Complaint received of instrument experiencing repeated control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 586

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

A/Sgt. M. Goncalves #6040
Name & Badge Number (Print)
A/Sgt. Ent #6040
Signature
2/14/14
Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Westampton Twp P.D., Westampton, NJ

Date Received: 03/18/14

Date given to service: 03/18/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NK-0025

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: Oct. 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFA-1575

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 04/03/14

Packing Slip

Customer no Order No Order date
 150053065 10691708 04/03/2014

Packing slip no. Ship date
 80848738 04/03/2014

Please reference on inquiries

ship to
 150053065
 TOWNSHIP OF WESTAMPTON POLICE DEPT
 ATTN: CHIEF
 BURLINGTON COUNTY
 710 RANCOCAS ROAD
 MOUNT HOLLY, NJ 08060
 USA

bill to
 [Empty box for bill to information]

Your Purchase Order Number
 SGT. ENT
 Any questions? Please contact:
 CHARLES ROBINSON II

Customer
 TOWNSHIP OF WESTAMPTON POLICE DEPT
 ATTN: CHIEF
 BURLINGTON COUNTY
 710 RANCOCAS ROAD
 MOUNT HOLLY, NJ 08060
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110:ARNK-0025 RECEIVED 03/18/2014
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053065 10691708 04/03/2014

Packing slip no. Ship date
80848738 04/03/2014

Please reference an inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6908455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6908486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Westampton Police Department 710 Rancocas Road Westampton, NJ 08060	2. Contact: Sgt. Ent 3. Phone Number: 609-267-3000	4. Date: 04-14-2014 5. County: Burlington
--	---	--

6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARNK-0025</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>
--	--	--

9. Reason for Service:
The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.


10. Comments:
See attached packing slip and Drager Return and Repair Form.

Last Known Sequential File #: N/A

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

A/Sgt. M. Goncalves #6040
Name & Badge Number (Print)


Signature

4/14/2014
Date

Packing Slip

Customer no Order No Order date
 150053065 10353260 07/08/2009

Packing slip no. Shp date
 80454006 07/08/2009

Please reference on inquiries

ship to
 150053065
 TOWNSHIP OF WESTAMPTON POLICE DEPT
 ATTN: CHIEF
 710 RANCOCAS ROAD
 BURLINGTON COUNTY
 MOUNT HOLLY, NJ 08060
 USA

bill to



Your Purchase Order Number

7110 WARRANTY
 Any questions? Please contact:
 MARA HERRERA

Customer
 TOWNSHIP OF WESTAMPTON POLICE DEPT
 ATTN: CHIEF
 710 RANCOCAS ROAD
 BURLINGTON COUNTY
 MOUNT HOLLY, NJ 08060
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Number of packages: 1

Item#	sh. Quant.	Part no. Description
		Waybill number: 3080609520000029740981622022012 Ship via: Fed Ex P1
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARNK-0025
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053065 10353260 07/08/2009

Packing slip no. Ship date
80454006 07/08/2009

Please reference on Inquiries

Page 2 / 2

Item#	sh. Quant.	Par. no. Description			
0020	1 EA	6808486 PLATES,7110 Ordered / Back ordered : 0 / 0 EA			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Return & Repair Form:

Customer Information:

Company Name: WEST HAMPTON B: _____ S: _____
 Date Received: 7-6 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR NK 0025
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/WHV</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C W/WHV</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/WHV</u>
<u>MP Lubo</u>	<u>Lubo</u>	<u>.5</u>	<u>N/C W/WHV</u>

Repair Notes:

REPLACE FC + 1 PLATE
CAL w/DC & OPS CHECK

Service Technician: [Signature]

Date: 7-08-09

DRAEGER SAFETY DIAGNOSTICS, INC.

CERTIFICATE

I, Dori L. Mansur Ratka, Esq., an attorney-at-law duly admitted, inter alia, to the bar of the State of New Jersey, under penalty of perjury under the laws of the United States of America and pursuant to 28 U.S.C. § 1746, hereby certify to the appropriate court officers of the State of New Jersey, on behalf of my client, Draeger Safety Diagnostics, Inc., a Delaware corporation ("Draeger"), and not in my individual capacity, that:


1. With respect to the various requests for records reflecting repairs performed on a particular device, the Repair & Return Form constitutes Draeger's record of the actual repair work conducted on the Alcotest® 7110 MK IIC device.

2. The Alcotest® 7110 MK IIC Checklist, Configuration - Calibration - QC, constitutes Draeger's internal record of the re-calibration and quality check conducted by Draeger on the device following the completion of the repair work. This Checklist is a standardized, step-by-step list of calibration and quality tasks undertaken by Draeger in order to support the "Certificate of Accuracy" issued with the repaired device.

3. The instrument print-outs, including the small pieces of paper titled "Control Check" or "Breath Test" and the "Calibrating Unit, New Standard Solution Report", are documents generated during the calibration and quality check, and are used for internal verification of the completion of the various steps on the above Checklist.

4. Once the device is recalibrated, which Draeger understands to be the standard practice of the State of New Jersey upon receipt of a device repaired by Draeger, the calibration and quality checks performed on the device while in Draeger's possession, including the various attached print outs, are rendered immaterial and superseded by the subsequent State calibration process.

IN WITNESS WHEREOF, the undersigned has executed this Certificate on the 25th day of October, 2011 in the Borough of Telford, Commonwealth of Pennsylvania.



Dori L. Mansur Ratka, Esq.
on behalf of
Draeger Safety Diagnostics, Inc.