

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00352	Calib. Date: 08/22/2011
Certification File No.:	00353	Calib. No.: 00016
Linearity File No.:	00354	Cert. Date: 08/22/2011
Solution File No.:	00367	Cert. No.: 00011
Sequential File No.:	00367	Lin. Date: 08/22/2011
		Lin. No.: 00011
		Soln. Date: 12/14/2011
		Soln. No.: 00092
		File Date: 12/14/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 05/23/2013
Solution Control Lot:	11E089		Bottle No.: 1315

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:14S		
Control 1 EC	0.099%	21:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15S		
Control 2 EC	0.099%	21:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:17S		
Control 3 EC	0.099%	21:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_

Badge No.: 16142

Date: 12/14/2011

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00352 Calib. Date: 08/22/2011 Calib. No.: 00016  
Certification File No.: 00353 Cert. Date: 08/22/2011 Cert. No.: 00011  
Linearity File No.: 00354 Lin. Date: 08/22/2011 Lin. No.: 00011  
Solution File No.: 00365 Soln. Date: 11/18/2011 Soln. No.: 00091  
Sequential File No.: 00365 File Date: 11/18/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 12/31/2011  
Solution Control Lot: 09L072 Bottle No.: 0220

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:56S		
Control 1 EC	0.099%	18:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	18:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:58S		
Control 2 EC	0.098%	18:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:59S		
Control 3 EC	0.099%	19:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

M1: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 11/18/2011

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00352 Calib. Date: 08/22/2011 Calib. No.: 00016  
Certification File No.: 00353 Cert. Date: 08/22/2011 Cert. No.: 00011  
Linearity File No.: 00354 Lin. Date: 08/22/2011 Lin. No.: 00011  
Solution File No.: 00359 Soln. Date: 10/19/2011 Soln. No.: 00090  
Sequential File No.: 00359 File Date: 10/19/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 12/31/2011  
Solution Control Lot: 09L072 Bottle No.: 0213

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:49D		
Control 1 EC	0.099%	21:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51D		
Control 2 EC	0.099%	21:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:52D		
Control 3 EC	0.098%	21:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 10/19/2011

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00352 Calib. Date: 08/22/2011 Calib. No.: 00016  
Certification File No.: 00353 Cert. Date: 08/22/2011 Cert. No.: 00011  
Linearity File No.: 00354 Lin. Date: 08/22/2011 Lin. No.: 00011  
Solution File No.: 00358 Soln. Date: 09/19/2011 Soln. No.: 00089  
Sequential File No.: 00358 File Date: 09/19/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 12/31/2011  
Solution Control Lot: 09L072 Bottle No.: 0218

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:38D		
Control 1 EC	0.099%	01:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:39D		
Control 2 EC	0.099%	01:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	01:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:41D		
Control 3 EC	0.098%	01:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 09/19/2011

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00329	Calib. Date:	03/07/2011
Certification File No.:	00330	Cert. Date:	03/07/2011
Linearity File No.:	00331	Lin. Date:	03/07/2011
Solution File No.:	00349	Soln. Date:	07/26/2011
Sequential File No.:	00349	File Date:	07/26/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	12/31/2011
Solution Control Lot:	09L072			Bottle No.:	0203

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:26D		
Control 1 EC	0.099%	15:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27D		
Control 2 EC	0.098%	15:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28D		
Control 3 EC	0.098%	15:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 07/26/2011

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00329	Calib. Date:	03/07/2011
Certification File No.:	00330	Cert. Date:	03/07/2011
Linearity File No.:	00331	Lin. Date:	03/07/2011
Solution File No.:	00347	Soln. Date:	06/27/2011
Sequential File No.:	00347	File Date:	06/27/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	12/31/2011
Solution Control Lot:	09L072			Bottle No.:	0202

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:07D		
Control 1 EC	0.099%	00:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08D		
Control 2 EC	0.098%	00:09D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:09D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:09D		
Control 3 EC	0.098%	00:10D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:10D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

M1: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 06/27/2011

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00329	Calib. Date:	03/07/2011
Certification File No.:	00330	Cert. Date:	03/07/2011
Linearity File No.:	00331	Lin. Date:	03/07/2011
Solution File No.:	00343	Soln. Date:	06/01/2011
Sequential File No.:	00343	File Date:	06/01/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	12/31/2011
Solution Control Lot:	09L072			Bottle No.:	0224

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:12D		
Control 1 EC	0.099%	23:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:13D		
Control 2 EC	0.098%	23:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:14D		
Control 3 EC	0.099%	23:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	23:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 06/01/2011

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
**Location:** WILLINGBORO TWP POLICE  
Calibration File No.: 00329 Calib. Date: 03/07/2011 Calib. No.: 00015  
Certification File No.: 00330 Cert. Date: 03/07/2011 Cert. No.: 00010  
Linearity File No.: 00331 Lin. Date: 03/07/2011 Lin. No.: 00010  
Solution File No.: 00337 Soln. Date: 04/03/2011 Soln. No.: 00083  
Sequential File No.: 00337 File Date: 04/03/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 12/31/2011  
Solution Control Lot: 09L072 Bottle No.: 0208

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:41D		
Control 1 EC	0.099%	23:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:43D		
Control 2 EC	0.100%	23:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:44D		
Control 3 EC	0.099%	23:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 04/03/2011



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00307	Calib. Date:	09/15/2010
Certification File No.:	00308	Cert. Date:	09/15/2010
Linearity File No.:	00309	Lin. Date:	09/15/2010
Solution File No.:	00326	Soln. Date:	02/23/2011
Sequential File No.:	00326	File Date:	02/23/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	09LO72	Expires:	12/31/2011
		Bottle No.:	0225

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:59S		
Control 1 EC	0.101%	23:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:01S		
Control 2 EC	0.101%	23:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03S		
Control 3 EC	0.101%	23:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 02/23/2011

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00307	Calib. Date:	09/15/2010
Certification File No.:	00308	Cert. Date:	09/15/2010
Linearity File No.:	00309	Lin. Date:	09/15/2010
Solution File No.:	00324	Soln. Date:	01/23/2011
Sequential File No.:	00324	File Date:	01/23/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	12/31/2011
Solution Control Lot:	09L072			Bottle No.:	0214

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:11S		
Control 1 EC	0.100%	23:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:12S		
Control 2 EC	0.100%	23:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:14S		
Control 3 EC	0.100%	23:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 01/23/2011