

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00574	Calib. Date:	03/01/2017
Certification File No.:	00575	Cert. No.:	00033
Linearity File No.:	00576	Cert. Date:	03/01/2017
Solution File No.:	00578	Lin. No.:	00026
Sequential File No.:	00578	Soln. Date:	03/29/2017
		Soln. No.:	00169
		File Date:	03/29/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	05/17/2018
Solution Control Lot:	201605D			Bottle No.:	1101

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:54S		
Control 1 EC	0.102%	07:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 2 EC	0.100%	07:56S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56S		
Control 3 EC	0.099%	07:57S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 03/29/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00560 Calib. Date: 10/13/2016 Calib. No.: 00032
Certification File No.: 00561 Cert. Date: 10/13/2016 Cert. No.: 00025
Linearity File No.: 00562 Lin. Date: 10/13/2016 Lin. No.: 00025
Solution File No.: 00573 Soln. Date: 02/06/2017 Soln. No.: 00167
Sequential File No.: 00573 File Date: 02/06/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 05/17/2018
Solution Control Lot: 201605D Bottle No.: 0780

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:11S		
Control 1 EC	0.102%	10:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S		
Control 2 EC	0.102%	10:14S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14S		
Control 3 EC	0.102%	10:15S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____

Sgt. J. Knecht

Badge No.: 16187

Date: 02/06/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00560	Calib. Date: 10/13/2016
Certification File No.:	00561	Calib. No.: 00032
Linearity File No.:	00562	Cert. Date: 10/13/2016
Solution File No.:	00570	Cert. No.: 00025
Sequential File No.:	00570	Lin. Date: 10/13/2016
		Lin. No.: 00025
		Soln. Date: 01/05/2017
		Soln. No.: 00166
		File Date: 01/05/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 05/17/2018
Solution Control Lot:	201605D		Bottle No.: 0783

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:17S		
Control 1 EC	0.101%	09:18S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19S		
Control 2 EC	0.101%	09:19S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20S		
Control 3 EC	0.101%	09:21S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____

Sgt. J. Knecht

Badge No.: 16187

Date: 01/05/2017