

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00606 Calib. Date: 01/19/2018 Calib. No.: 00035
Certification File No.: 00607 Cert. Date: 01/19/2018 Cert. No.: 00028
Linearity File No.: 00608 Lin. Date: 01/19/2018 Lin. No.: 00028
Solution File No.: 00613 Soln. Date: 03/20/2018 Soln. No.: 00184
Sequential File No.: 00613 File Date: 03/20/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 10/17/2019
Solution Control Lot: 17360 Bottle No.: 1108

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 02:09S | | |
| Control 1 EC | 0.098% | 02:10S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 02:10S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:11S | | |
| Control 2 EC | 0.096% | 02:12S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 02:12S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:12S | | |
| Control 3 EC | 0.098% | 02:13S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 02:13S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:14S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 03/20/2018

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | |
| Calibration File No.: | 00606 | Calib. Date: 01/19/2018 |
| Certification File No.: | 00607 | Calib. No.: 00035 |
| Linearity File No.: | 00608 | Cert. Date: 01/19/2018 |
| Solution File No.: | 00611 | Cert. No.: 00028 |
| Sequential File No.: | 00611 | Lin. Date: 01/19/2018 |
| | | Lin. No.: 00028 |
| | | Soln. Date: 02/17/2018 |
| | | Soln. No.: 00183 |
| | | File Date: 02/17/2018 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWE S3-0188 |
| Solution Control Lot: | 17360 | Expires: 10/17/2019 |
| | | Bottle No.: 1251 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 21:25S | | |
| Control 1 EC | 0.097% | 21:25S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 21:25S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:26S | | |
| Control 2 EC | 0.098% | 21:27S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 21:27S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:27S | | |
| Control 3 EC | 0.096% | 21:28S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 21:28S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:29S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 02/17/2018

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | |
| Calibration File No.: | 00589 | Calib. Date: 09/12/2017 |
| Certification File No.: | 00590 | Calib. No.: 00034 |
| Linearity File No.: | 00591 | Cert. Date: 09/12/2017 |
| Solution File No.: | 00605 | Cert. No.: 00027 |
| Sequential File No.: | 00605 | Lin. Date: 09/12/2017 |
| | | Lin. No.: 00027 |
| | | Soln. Date: 01/12/2018 |
| | | Soln. No.: 00181 |
| | | File Date: 01/12/2018 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWE S3-0188 |
| Solution Control Lot: | 201605D | Expires: 05/17/2018 |
| | | Bottle No.: 0297 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 06:13S | | |
| Control 1 EC | 0.104% | 06:14S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 06:14S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:15S | | |
| Control 2 EC | 0.101% | 06:16S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.102% | 06:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:16S | | |
| Control 3 EC | 0.101% | 06:17S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.104% | 06:17S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:18S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: 

Badge No.: 16142

Date: 01/12/2018