Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE			
Calibration File No.:	00632	Calib. Date:	01/07/2019	Calib. No.: 00037
Certification File No.:	00633	Cert. Date:	01/07/2019	Cert. No.: 00030
Linearity File No.:	00634	Lin. Date:	01/07/2019	Lin. No.: 00030
Solution File No.:	00638	Soln. Date:	02/05/2019	Soln. No.: 00196
Sequential File No.:	00638	File Date:	02/05/2019	
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%			Expires: 10/17/2019
Solution Control Lot:	17360			Bottle No.: 1249

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:14S		
Control 1 EC	0.100%	03:14S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15S		
Control 2 EC	0.099%	03:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17S		
Control 3 EC	0.097%	03:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT First Name: JENNIFER MI: A

Signature: Sst Jennifer 187

Badge No.: 16187

Date: 02/05/2019