

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00108	Calib. Date:	02/13/2007
Certification File No.:	00109	Cert. Date:	02/13/2007
Linearity File No.:	00110	Lin. Date:	02/13/2007
Solution File No.:	00158	Soln. Date:	12/10/2007
Sequential File No.:	00158	File Date:	12/10/2007
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	06J036	Expires:	10/10/2008
		Bottle No.:	0048

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:47S		
Control 1 EC	0.099%	15:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		
Control 2 EC	0.097%	15:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50S		
Control 3 EC	0.098%	15:50S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:

Sgt. Raymond Norfo #102

Badge No.: 102

Date: 12/10/2007

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00108	Calib. Date:	02/13/2007
Certification File No.:	00109	Cert. Date:	02/13/2007
Linearity File No.:	00110	Lin. Date:	02/13/2007
Solution File No.:	00153	Soln. Date:	11/10/2007
Sequential File No.:	00153	File Date:	11/10/2007

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	10/10/2008
Solution Control Lot:	06J036			Bottle No.:	0039

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:11S		
Control 1 EC	0.099%	17:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:12S		
Control 2 EC	0.096%	17:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14S		
Control 3 EC	0.097%	17:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____

Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 11/10/2007

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00150 Soln. Date: 10/19/2007 Soln. No.: 00031
Sequential File No.: 00150 File Date: 10/19/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0029

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:43D		
Control 1 EC	0.099%	07:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		
Control 2 EC	0.098%	07:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46D		
Control 3 EC	0.098%	07:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: 

*102

Badge No.: 102

Date: 10/19/2007

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00146 Soln. Date: 09/21/2007 Soln. No.: 00030
Sequential File No.: 00146 File Date: 09/21/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0037

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:25D		
Control 1 EC	0.099%	10:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:26D		
Control 2 EC	0.098%	10:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 3 EC	0.098%	10:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2-293 we

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: TPR. I WILL M. Cross #5342

Badge No.: 5342

Date: 09/21/2007

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00145 Soln. Date: 08/28/2007 Soln. No.: 00029
Sequential File No.: 00145 File Date: 08/28/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: O6J036 Bottle No.: 0038

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:21D		
Control 1 EC	0.098%	11:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		
Control 2 EC	0.097%	11:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 3 EC	0.097%	11:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____



Badge No.: 102

Date: 08/28/2007

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00141 Soln. Date: 07/30/2007 Soln. No.: 00028
Sequential File No.: 00141 File Date: 07/30/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0032

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:17D		
Control 1 EC	0.099%	10:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 2 EC	0.098%	10:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		
Control 3 EC	0.098%	10:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____

Raymond A. Norfo

Badge No.: 102

Date: 07/30/2007

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00138 Soln. Date: 07/07/2007 Soln. No.: 00027
Sequential File No.: 00138 File Date: 07/07/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0027

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:15D		
Control 1 EC	0.099%	09:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16D		
Control 2 EC	0.098%	09:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:17D		
Control 3 EC	0.098%	09:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____



Badge No.: 102

Date: 07/07/2007

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00134 Soln. Date: 06/14/2007 Soln. No.: 00026
Sequential File No.: 00134 File Date: 06/14/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0033

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:46D		
Control 1 EC	0.099%	09:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		
Control 2 EC	0.098%	09:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		
Control 3 EC	0.097%	09:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____



Badge No.: 102

Date: 06/14/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00128 Soln. Date: 05/16/2007 Soln. No.: 00025
Sequential File No.: 00128 File Date: 05/16/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0028

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:57D		
Control 1 EC	0.098%	08:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		
Control 2 EC	0.097%	08:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59D		
Control 3 EC	0.098%	09:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 05/16/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
Solution File No.: 00119 Soln. Date: 04/10/2007 Soln. No.: 00023
Sequential File No.: 00119 File Date: 04/10/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 10/10/2008
Solution Control Lot: 06J036 Bottle No.: 0043

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:59D		
Control 1 EC	0.099%	09:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01D		
Control 2 EC	0.098%	09:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02D		
Control 3 EC	0.097%	09:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 04/10/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
 Location: WILLINGBORO TWP POLICE
 Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004
 Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003
 Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003
 Solution File No.: 00114 Soln. Date: 03/13/2007 Soln. No.: 00022
 Sequential File No.: 00114 File Date: 03/13/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
 Control Solution %: 0.100% Expires: 10/10/2008
 Solution Control Lot: 06J036 Bottle No.: 0030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:14S		
Control 1 EC	0.100%	08:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		
Control 2 EC	0.098%	08:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		
Control 3 EC	0.097%	08:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 03/13/2007

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00085	Calib. Date: 10/31/2006
Certification File No.:	00086	Cert. No.: 00003
Linearity File No.:	00087	Cert. Date: 10/31/2006
Solution File No.:	00105	Lin. No.: 00002
Sequential File No.:	00105	Lin. Date: 10/31/2006
		Soln. No.: 00020
		Soln. Date: 01/20/2007
		File Date: 01/20/2007
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	05D017	Expires: 04/14/2007
		Bottle No.: 0247

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:11S		
Control 1 EC	0.098%	14:11S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:12S		
Control 2 EC	0.096%	14:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:13S		
Control 3 EC	0.096%	14:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____

Sgt. Raymond Norfo #102

Badge No.: 102

Date: 01/20/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00085 Calib. Date: 10/31/2006 Calib. No.: 00003
Certification File No.: 00086 Cert. Date: 10/31/2006 Cert. No.: 00002
Linearity File No.: 00087 Lin. Date: 10/31/2006 Lin. No.: 00002
Solution File No.: 00103 Soln. Date: 01/15/2007 Soln. No.: 00019
Sequential File No.: 00103 File Date: 01/15/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0225

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:59S		
Control 1 EC	0.098%	09:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		
Control 2 EC	0.096%	09:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02S		
Control 3 EC	0.096%	09:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 01/15/2007