

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00200	Calib. Date: 08/18/2008
Certification File No.:	00201	Calib. No.: 00008
Linearity File No.:	00202	Cert. Date: 08/18/2008
Solution File No.:	00221	Cert. No.: 00005
Sequential File No.:	00221	Lin. Date: 08/18/2008
		Lin. No.: 00005
		Soln. Date: 12/15/2008
		Soln. No.: 00049
		File Date: 12/15/2008

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWE S3-0188
Control Solution %: 0.100%		Expires: 06/09/2010
Solution Control Lot: 08F053		Bottle No.: 0462

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:48S		
Control 1 EC	0.098%	11:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:49S		
Control 2 EC	0.097%	11:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51S		
Control 3 EC	0.097%	11:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: 

Badge No.: 102

Date: 12/15/2008

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
**Location:** WILLINGBORO TWP POLICE  
**Calibration File No.:** 00200 **Calib. Date:** 08/18/2008 **Calib. No.:** 00008  
**Certification File No.:** 00201 **Cert. Date:** 08/18/2008 **Cert. No.:** 00005  
**Linearity File No.:** 00202 **Lin. Date:** 08/18/2008 **Lin. No.:** 00005  
**Solution File No.:** 00219 **Soln. Date:** 12/03/2008 **Soln. No.:** 00048  
**Sequential File No.:** 00219 **File Date:** 12/03/2008

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0188  
**Control Solution %:** 0.100% **Expires:** 06/09/2010  
**Solution Control Lot:** 08F053 **Bottle No.:** 0455

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:49S		
Control 1 EC	0.098%	12:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:51S		
Control 2 EC	0.098%	12:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52S		
Control 3 EC	0.098%	12:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:

 102

Badge No.: 102

Date: 12/03/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00200	Calib. Date:	08/18/2008
Certification File No.:	00201	Cert. Date:	08/18/2008
Linearity File No.:	00202	Lin. Date:	08/18/2008
Solution File No.:	00213	Soln. Date:	11/03/2008
Sequential File No.:	00213	File Date:	11/03/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	08F053	Expires:	06/09/2010
		Bottle No.:	0461

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:31S		
Control 1 EC	0.098%	11:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		
Control 2 EC	0.097%	11:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33S		
Control 3 EC	0.097%	11:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:

*Sgt. Raymond Norfo* <sup>102</sup>

Badge No.: 102

Date: 11/03/2008

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00200 Calib. Date: 08/18/2008 Calib. No.: 00008  
Certification File No.: 00201 Cert. Date: 08/18/2008 Cert. No.: 00005  
Linearity File No.: 00202 Lin. Date: 08/18/2008 Lin. No.: 00005  
Solution File No.: 00209 Soln. Date: 10/15/2008 Soln. No.: 00046  
Sequential File No.: 00209 File Date: 10/15/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
Control Solution %: 0.100% Expires: 06/09/2010  
Solution Control Lot: 08F053 Bottle No.: 0460

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:51D		
Control 1 EC	0.098%	10:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		
Control 2 EC	0.097%	10:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		
Control 3 EC	0.097%	10:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		

All tests within acceptable tolerance.


On this date, 1 installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

M1: A

Signature: 

Badge No.: 102

Date: 10/15/2008

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
**Location:** WILLINGBORO TWP POLICE  
**Calibration File No.:** 00200 **Calib. Date:** 08/18/2008 **Calib. No.:** 00008  
**Certification File No.:** 00201 **Cert. Date:** 08/18/2008 **Cert. No.:** 00005  
**Linearity File No.:** 00202 **Lin. Date:** 08/18/2008 **Lin. No.:** 00005  
**Solution File No.:** 00205 **Soln. Date:** 09/16/2008 **Soln. No.:** 00045  
**Sequential File No.:** 00205 **File Date:** 09/16/2008

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0188  
**Control Solution %:** 0.100% **Expires:** 06/09/2010  
**Solution Control Lot:** 08F053 **Bottle No.:** 0456

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:01D		
Control 1 EC	0.098%	10:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 2 EC	0.097%	10:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		
Control 3 EC	0.097%	10:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 09/16/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00169	Calib. Date:	01/25/2008
Certification File No.:	00170	Cert. Date:	01/25/2008
Linearity File No.:	00171	Lin. Date:	01/25/2008
Solution File No.:	00199	Soln. Date:	08/14/2008
Sequential File No.:	00199	File Date:	08/14/2008

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	10/10/2008
Solution Control Lot:	06J036			Bottle No.:	0040

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:41D		
Control 1 EC	0.099%	04:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:42D		
Control 2 EC	0.099%	04:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:43D		
Control 3 EC	0.099%	04:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: \_\_\_\_\_

Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 08/14/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00169	Calib. Date: 01/25/2008
Certification File No.:	00170	Calib. No.: 00007
Linearity File No.:	00171	Cert. Date: 01/25/2008
Solution File No.:	00198	Cert. No.: 00004
Sequential File No.:	00198	Lin. Date: 01/25/2008
		Lin. No.: 00004
		Soln. Date: 07/16/2008
		Soln. No.: 00042
		File Date: 07/16/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	06J036	Expires: 10/10/2008
		Bottle No.: 0031

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:00D		
Control 1 EC	0.099%	12:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		
Control 2 EC	0.099%	12:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		
Control 3 EC	0.099%	12:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: \_\_\_\_\_

Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 07/16/2008

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00169 Calib. Date: 01/25/2008 Calib. No.: 00007  
Certification File No.: 00170 Cert. Date: 01/25/2008 Cert. No.: 00004  
Linearity File No.: 00171 Lin. Date: 01/25/2008 Lin. No.: 00004  
Solution File No.: 00191 Soln. Date: 06/17/2008 Soln. No.: 00041  
Sequential File No.: 00191 File Date: 06/17/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
Control Solution %: 0.100% Expires: 10/10/2008  
Solution Control Lot: 06J036 Bottle No.: 0049

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39D		
Control 1 EC	0.099%	11:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40D		
Control 2 EC	0.099%	11:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41D		
Control 3 EC	0.099%	11:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

M1: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 06/17/2008



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00169 Calib. Date: 01/25/2008 Calib. No.: 00007  
Certification File No.: 00170 Cert. Date: 01/25/2008 Cert. No.: 00004  
Linearity File No.: 00171 Lin. Date: 01/25/2008 Lin. No.: 00004  
Solution File No.: 00186 Soln. Date: 05/20/2008 Soln. No.: 00040  
Sequential File No.: 00186 File Date: 05/20/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
Control Solution %: 0.100% Expires: 10/10/2008  
Solution Control Lot: 06J036 Bottle No.: 0035

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:32D		
Control 1 EC	0.097%	11:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		
Control 2 EC	0.097%	11:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		
Control 3 EC	0.098%	11:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 05/20/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00169	Calib. Date: 01/25/2008
Certification File No.:	00170	Cert. No.: 00007
Linearity File No.:	00171	Cert. Date: 01/25/2008
Solution File No.:	00184	Lin. No.: 00004
Sequential File No.:	00184	Lin. Date: 01/25/2008
		Soln. No.: 00039
		Soln. Date: 04/22/2008
		File Date: 04/22/2008

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	06J036	Expires: 10/10/2008
		Bottle No.: 0034

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:55D		
Control 1 EC	0.100%	09:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 2 EC	0.099%	09:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 3 EC	0.098%	09:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NOREO

First Name: RAYMOND

M1: A

Signature: \_\_\_\_\_

Sgt. Raymond Noreo #102

Badge No.: 102

Date: 04/22/2008

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00169 Calib. Date: 01/25/2008 Calib. No.: 00007  
Certification File No.: 00170 Cert. Date: 01/25/2008 Cert. No.: 00004  
Linearity File No.: 00171 Lin. Date: 01/25/2008 Lin. No.: 00004  
Solution File No.: 00174 Soln. Date: 02/22/2008 Soln. No.: 00037  
Sequential File No.: 00174 File Date: 02/22/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 10/10/2008  
Solution Control Lot: 06J036 Bottle No.: 0044

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:20S		
Control 1 EC	0.101%	12:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21S		
Control 2 EC	0.099%	12:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23S		
Control 3 EC	0.100%	12:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 02/22/2008

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00108 Calib. Date: 02/13/2007 Calib. No.: 00004  
Certification File No.: 00109 Cert. Date: 02/13/2007 Cert. No.: 00003  
Linearity File No.: 00110 Lin. Date: 02/13/2007 Lin. No.: 00003  
Solution File No.: 00163 Soln. Date: 01/05/2008 Soln. No.: 00034  
Sequential File No.: 00163 File Date: 01/05/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 10/10/2008  
Solution Control Lot: 06J036 Bottle No.: 0036

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:42S		
Control 1 EC	0.099%	11:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		
Control 2 EC	0.096%	11:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45S		
Control 3 EC	0.097%	11:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:

 # 102

Badge No.: 102

Date: 01/05/2008