

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00265	Calib. Date:	09/09/2009
Certification File No.:	00266	Cert. Date:	09/09/2009
Linearity File No.:	00267	Lin. Date:	09/09/2009
Solution File No.:	00285	Soln. Date:	12/22/2009
Sequential File No.:	00285	File Date:	12/22/2009

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	06/09/2010
Solution Control Lot:	08F053			Bottle No.:	0472

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:49S		
Control 1 EC	0.097%	09:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		
Control 2 EC	0.098%	09:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52S		
Control 3 EC	0.097%	09:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 12/22/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00265	Calib. Date: 09/09/2009
Certification File No.:	00266	Cert. No.: 00012
Linearity File No.:	00267	Cert. Date: 09/09/2009
Solution File No.:	00282	Lin. No.: 00007
Sequential File No.:	00282	Lin. Date: 09/09/2009
		Soln. No.: 00064
		Soln. Date: 11/24/2009
		File Date: 11/24/2009

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 06/09/2010
Solution Control Lot:	08F053		Bottle No.: 0451

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:55S		
Control 1 EC	0.098%	10:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 2 EC	0.097%	10:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		
Control 3 EC	0.097%	10:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____

Badge No.: 16142

Date: 11/24/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00265	Calib. Date: 09/09/2009
Certification File No.:	00266	Calib. No.: 00012
Linearity File No.:	00267	Cert. Date: 09/09/2009
Solution File No.:	00280	Cert. No.: 00007
Sequential File No.:	00280	Lin. Date: 09/09/2009
		Lin. No.: 00007
		Soln. Date: 10/27/2009
		Soln. No.: 00063
		File Date: 10/27/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0452

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:44S		
Control 1 EC	0.098%	07:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45S		
Control 2 EC	0.099%	07:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47S		
Control 3 EC	0.099%	07:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 10/27/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00265 Calib. Date: 09/09/2009 Calib. No.: 00012
Certification File No.: 00266 Cert. Date: 09/09/2009 Cert. No.: 00007
Linearity File No.: 00267 Lin. Date: 09/09/2009 Lin. No.: 00007
Solution File No.: 00276 Soln. Date: 10/06/2009 Soln. No.: 00062
Sequential File No.: 00276 File Date: 10/06/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0473

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:43D		
Control 1 EC	0.097%	10:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		
Control 2 EC	0.098%	10:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 3 EC	0.097%	10:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOMON

First Name: KINAMO

MI: A

Signature:  #164

Badge No.: 16164

Date: 10/06/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00241 Calib. Date: 03/12/2009 Calib. No.: 00011
Certification File No.: 00242 Cert. Date: 03/12/2009 Cert. No.: 00006
Linearity File No.: 00243 Lin. Date: 03/12/2009 Lin. No.: 00006
Solution File No.: 00262 Soln. Date: 08/23/2009 Soln. No.: 00060
Sequential File No.: 00262 File Date: 08/23/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0471

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:44D		
Control 1 EC	0.100%	10:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45D		
Control 2 EC	0.100%	10:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 3 EC	0.099%	10:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 08/23/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00241	Calib. Date:	03/12/2009
Certification File No.:	00242	Cert. Date:	03/12/2009
Linearity File No.:	00243	Lin. Date:	03/12/2009
Solution File No.:	00260	Soln. Date:	07/26/2009
Sequential File No.:	00260	File Date:	07/26/2009

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	06/09/2010
Solution Control Lot:	08F053			Bottle No.:	0465

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:57D		
Control 1 EC	0.099%	10:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 2 EC	0.099%	10:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		
Control 3 EC	0.099%	11:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 07/26/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00241	Calib. Date: 03/12/2009
Certification File No.:	00242	Calib. No.: 00011
Linearity File No.:	00243	Cert. Date: 03/12/2009
Solution File No.:	00256	Cert. No.: 00006
Sequential File No.:	00256	Lin. Date: 03/12/2009
		Lin. No.: 00006
		Soln. Date: 06/28/2009
		Soln. No.: 00058
		File Date: 06/28/2009

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0469

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:46D		
Control 1 EC	0.099%	09:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		
Control 2 EC	0.099%	09:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		
Control 3 EC	0.099%	09:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 06/28/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00241 Calib. Date: 03/12/2009 Calib. No.: 00011
Certification File No.: 00242 Cert. Date: 03/12/2009 Cert. No.: 00006
Linearity File No.: 00243 Lin. Date: 03/12/2009 Lin. No.: 00006
Solution File No.: 00250 Soln. Date: 06/05/2009 Soln. No.: 00057
Sequential File No.: 00250 File Date: 06/05/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0459

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:02D		
Control 1 EC	0.098%	18:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		
Control 2 EC	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:05D		
Control 3 EC	0.099%	18:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 06/05/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00241	Calib. Date:	03/12/2009
Certification File No.:	00242	Cert. Date:	03/12/2009
Linearity File No.:	00243	Lin. Date:	03/12/2009
Solution File No.:	00249	Soln. Date:	05/08/2009
Sequential File No.:	00249	File Date:	05/08/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	08F053	Expires:	06/09/2010
		Bottle No.:	0458

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:46D		
Control 1 EC	0.100%	10:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		
Control 2 EC	0.100%	10:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		
Control 3 EC	0.100%	10:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 05/08/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00241	Calib. Date:	03/12/2009
Certification File No.:	00242	Cert. Date:	03/12/2009
Linearity File No.:	00243	Lin. Date:	03/12/2009
Solution File No.:	00247	Soln. Date:	04/10/2009
Sequential File No.:	00247	File Date:	04/10/2009

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	06/09/2010
Solution Control Lot:	08F053			Bottle No.:	0467

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:11D		
Control 1 EC	0.101%	10:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 2 EC	0.101%	10:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		
Control 3 EC	0.101%	10:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 04/10/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00200 Calib. Date: 08/18/2008 Calib. No.: 00008
Certification File No.: 00201 Cert. Date: 08/18/2008 Cert. No.: 00005
Linearity File No.: 00202 Lin. Date: 08/18/2008 Lin. No.: 00005
Solution File No.: 00234 Soln. Date: 02/07/2009 Soln. No.: 00052
Sequential File No.: 00234 File Date: 02/07/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0457

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:49S		
Control 1 EC	0.096%	08:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		
Control 2 EC	0.097%	08:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	08:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52S		
Control 3 EC	0.096%	08:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 02/07/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00200 **Calib. Date:** 08/18/2008 **Calib. No.:** 00008
Certification File No.: 00201 **Cert. Date:** 08/18/2008 **Cert. No.:** 00005
Linearity File No.: 00202 **Lin. Date:** 08/18/2008 **Lin. No.:** 00005
Solution File No.: 00231 **Soln. Date:** 01/30/2009 **Soln. No.:** 00051
Sequential File No.: 00231 **File Date:** 01/30/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0188
Control Solution %: 0.100% **Expires:** 06/09/2010
Solution Control Lot: 08F053 **Bottle No.:** 0463

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:29S		
Control 1 EC	0.096%	09:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		
Control 2 EC	0.097%	09:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	09:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32S		
Control 3 EC	0.097%	09:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	09:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 01/30/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00200 Calib. Date: 08/18/2008 Calib. No.: 00008
Certification File No.: 00201 Cert. Date: 08/18/2008 Cert. No.: 00005
Linearity File No.: 00202 Lin. Date: 08/18/2008 Lin. No.: 00005
Solution File No.: 00225 Soln. Date: 01/02/2009 Soln. No.: 00050
Sequential File No.: 00225 File Date: 01/02/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0466

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:32S		
Control 1 EC	0.096%	10:33S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		
Control 2 EC	0.096%	10:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S		
Control 3 EC	0.096%	10:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	10:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NORFO

First Name: RAYMOND

M1: A

Signature: _____



Badge No.: 102

Date: 01/02/2009