

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00307	Calib. Date:	09/15/2010
Certification File No.:	00308	Cert. Date:	09/15/2010
Linearity File No.:	00309	Lin. Date:	09/15/2010
Solution File No.:	00321	Soln. Date:	12/28/2010
Sequential File No.:	00321	File Date:	12/28/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	09LO72	Expires:	12/31/2011
		Bottle No.:	0219

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:34S		
Control 1 EC	0.101%	09:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36S		
Control 2 EC	0.101%	09:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38S		
Control 3 EC	0.101%	09:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 12/28/2010



# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00307 Calib. Date: 09/15/2010 Calib. No.: 00014  
Certification File No.: 00308 Cert. Date: 09/15/2010 Cert. No.: 00009  
Linearity File No.: 00309 Lin. Date: 09/15/2010 Lin. No.: 00009  
Solution File No.: 00318 Soln. Date: 11/13/2010 Soln. No.: 00077  
Sequential File No.: 00318 File Date: 11/13/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
Control Solution %: 0.100% Expires: 12/31/2011  
Solution Control Lot: 09L072 Bottle No.: 0201

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:59S		
Control 1 EC	0.098%	11:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00S		
Control 2 EC	0.098%	12:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02S		
Control 3 EC	0.098%	12:02S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 11/13/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00307	Calib. Date:	09/15/2010
Certification File No.:	00308	Cert. Date:	09/15/2010
Linearity File No.:	00309	Lin. Date:	09/15/2010
Solution File No.:	00315	Soln. Date:	10/17/2010
Sequential File No.:	00315	File Date:	10/17/2010
		Calib. No.:	00014
		Cert. No.:	00009
		Lin. No.:	00009
		Soln. No.:	00076

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	12/31/2011
Solution Control Lot:	09L072			Bottle No.:	0206

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:18D	.	
Control 1 EC	0.098%	10:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		
Control 2 EC	0.098%	10:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		
Control 3 EC	0.098%	10:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 10/17/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00307	Calib. Date:	09/15/2010
Certification File No.:	00308	Cert. Date:	09/15/2010
Linearity File No.:	00309	Lin. Date:	09/15/2010
Solution File No.:	00310	Soln. Date:	09/15/2010
Sequential File No.:	00310	File Date:	09/15/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	09L072	Expires:	12/31/2011
		Bottle No.:	0212

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:10D		
Control 1 EC	0.100%	12:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.100%	12:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.100%	12:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2-389<sub>jo</sub>

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*Tpr II J. Dellanoce # 6027*

Badge No.: 6027

Date: 09/15/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00289	Calib. No.: 00013
Certification File No.:	00290	Cert. No.: 00008
Linearity File No.:	00291	Lin. No.: 00008
Solution File No.:	00303	Soln. No.: 00074
Sequential File No.:	00303	File Date: 08/31/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	09L072	Expires: 12/31/2011
		Bottle No.: 0211

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:16D		
Control 1 EC	0.100%	09:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		
Control 2 EC	0.100%	09:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		
Control 3 EC	0.100%	09:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_

Badge No.: 16142

Date: 08/31/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00289	Calib. Date: 03/16/2010
Certification File No.:	00290	Calib. No.: 00013
Linearity File No.:	00291	Cert. Date: 03/16/2010
Solution File No.:	00302	Cert. No.: 00008
Sequential File No.:	00302	Lin. Date: 03/16/2010
		Lin. No.: 00008
		Soln. Date: 08/03/2010
		Soln. No.: 00073
		File Date: 08/03/2010

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0188
Control Solution %:	0.100%		Expires: 12/31/2011
Solution Control Lot:	09L072		Bottle No.: 0216

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:45D		
Control 1 EC	0.099%	10:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 2 EC	0.100%	10:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		
Control 3 EC	0.099%	10:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 08/03/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00289	Calib. Date: 03/16/2010
Certification File No.:	00290	Calib. No.: 00013
Linearity File No.:	00291	Cert. Date: 03/16/2010
Solution File No.:	00300	Cert. No.: 00008
Sequential File No.:	00300	Lin. Date: 03/16/2010
		Lin. No.: 00008
		Soln. Date: 07/06/2010
		Soln. No.: 00072
		File Date: 07/06/2010

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWE S3-0188
Control Solution %: 0.100%		Expires: 12/31/2011
Solution Control Lot: 09L072		Bottle No.: 0217

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:58D		
Control 1 EC	0.101%	09:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		
Control 2 EC	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 3 EC	0.100%	10:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 07/06/2010



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00289	Calib. Date: 03/16/2010
Certification File No.:	00290	Calib. No.: 00013
Linearity File No.:	00291	Cert. Date: 03/16/2010
Solution File No.:	00299	Cert. No.: 00008
Sequential File No.:	00299	Lin. Date: 03/16/2010
		Lin. No.: 00008
		Soln. Date: 06/08/2010
		Soln. No.: 00071
		File Date: 06/08/2010

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	09L072	Expires: 12/31/2011
		Bottle No.: 0222

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:03D		
Control 1 EC	0.100%	10:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 2 EC	0.099%	10:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D		
Control 3 EC	0.099%	10:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 06/08/2010

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00289 Calib. Date: 03/16/2010 Calib. No.: 00013  
Certification File No.: 00290 Cert. Date: 03/16/2010 Cert. No.: 00008  
Linearity File No.: 00291 Lin. Date: 03/16/2010 Lin. No.: 00008  
Solution File No.: 00297 Soln. Date: 05/11/2010 Soln. No.: 00070  
Sequential File No.: 00297 File Date: 05/11/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
Control Solution %: 0.100% Expires: 12/31/2011  
Solution Control Lot: 09L072 Bottle No.: 0221

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:28D		
Control 1 EC	0.099%	10:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		
Control 2 EC	0.099%	10:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31D		
Control 3 EC	0.100%	10:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 05/11/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00289	Calib. Date: 03/16/2010
Certification File No.:	00290	Calib. No.: 00013
Linearity File No.:	00291	Cert. Date: 03/16/2010
Solution File No.:	00294	Cert. No.: 00008
Sequential File No.:	00294	Lin. Date: 03/16/2010
		Lin. No.: 00008
		Soln. Date: 04/13/2010
		Soln. No.: 00069
		File Date: 04/13/2010

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0188
Control Solution %:	0.100%		Expires: 06/09/2010
Solution Control Lot:	08F053		Bottle No.: 0475

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:10D		
Control 1 EC	0.100%	09:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:11D		
Control 2 EC	0.099%	09:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		
Control 3 EC	0.099%	09:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 04/13/2010

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
**Location:** WILLINGBORO TWP POLICE  
**Calibration File No.:** 00289 **Calib. Date:** 03/16/2010 **Calib. No.:** 00013  
**Certification File No.:** 00290 **Cert. Date:** 03/16/2010 **Cert. No.:** 00008  
**Linearity File No.:** 00291 **Lin. Date:** 03/16/2010 **Lin. No.:** 00008  
**Solution File No.:** 00292 **Soln. Date:** 03/16/2010 **Soln. No.:** 00068  
**Sequential File No.:** 00292 **File Date:** 03/16/2010

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0188  
**Control Solution %:** 0.100% **Expires:** 06/09/2010  
**Solution Control Lot:** 08F053 **Bottle No.:** 0468

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:41S	34.0°C	
Control 1 EC	0.100%	11:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42S	34.0°C	
Control 2 EC	0.100%	11:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S	34.0°C	
Control 3 EC	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45S	34.0°C	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP3 - 389040

**Changed By:**

Last Name: DELLANOCE First Name: JOSEPH MI: S

Signature: Joseph Dellanoce # 6027 Badge No.: 6027 Date: 03/16/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00265	Calib. Date:	09/09/2009
Certification File No.:	00266	Cert. Date:	09/09/2009
Linearity File No.:	00267	Lin. Date:	09/09/2009
Solution File No.:	00288	Soln. Date:	02/20/2010
Sequential File No.:	00288	File Date:	02/20/2010

Calib. No.:	00012
Cert. No.:	00007
Lin. No.:	00007
Soln. No.:	00067

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	06/09/2010
Solution Control Lot:	08F053			Bottle No.:	0474

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:26S		
Control 1 EC	0.098%	12:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27S		
Control 2 EC	0.098%	12:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		
Control 3 EC	0.098%	12:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 02/20/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00265	Calib. Date:	09/09/2009
Certification File No.:	00266	Cert. Date:	09/09/2009
Linearity File No.:	00267	Lin. Date:	09/09/2009
Solution File No.:	00286	Soln. Date:	01/19/2010
Sequential File No.:	00286	File Date:	01/19/2010

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	06/09/2010
Solution Control Lot:	08F053			Bottle No.:	0454

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:36S		
Control 1 EC	0.097%	11:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37S		
Control 2 EC	0.098%	11:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39S		
Control 3 EC	0.098%	11:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 01/19/2010