

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00006	Calib. Date: 11/03/2005
Certification File No.:	00007	Calib. No.: 00002
Linearity File No.:	00008	Cert. Date: 11/03/2005
Solution File No.:	00009	Cert. No.: 00001
Sequential File No.:	00009	Lin. Date: 11/03/2005
		Lin. No.: 00001
		Soln. Date: 11/03/2005
		Soln. No.: 00002
		File Date: 11/03/2005

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 04/14/2007
Solution Control Lot:	05D017		Bottle No.: 0158

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:08S		
Control 1 EC	0.100%	11:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		
Control 2 EC	0.100%	11:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10S		
Control 3 EC	0.100%	11:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: \_\_\_\_\_

TPR. I WILLIAM M. CROSS #5342

Badge No.: 5342

Date: 11/03/2005

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00006 Calib. Date: 11/03/2005 Calib. No.: 00002  
Certification File No.: 00007 Cert. Date: 11/03/2005 Cert. No.: 00001  
Linearity File No.: 00008 Lin. Date: 11/03/2005 Lin. No.: 00001  
Solution File No.: 00011 Soln. Date: 11/13/2005 Soln. No.: 00003  
Sequential File No.: 00011 File Date: 11/13/2005

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189  
Control Solution %: 0.100% Expires: 04/14/2007  
Solution Control Lot: 05D017 Bottle No.: 0155

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:44S		
Control 1 EC	0.099%	14:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45S		
Control 2 EC	0.099%	14:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47S		
Control 3 EC	0.098%	14:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:48S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: NORFO

First Name: RAYMOND

M1: A

Signature: 

Badge No.: 102

Date: 11/13/2005

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00006	Calib. Date:	11/03/2005
Certification File No.:	00007	Cert. Date:	11/03/2005
Linearity File No.:	00008	Lin. Date:	11/03/2005
Solution File No.:	00018	Soln. Date:	12/11/2005
Sequential File No.:	00018	File Date:	12/11/2005
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	05D017	Expires:	04/14/2007
		Bottle No.:	0209

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:26S		
Control 1 EC	0.099%	09:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27S		
Control 2 EC	0.099%	09:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28S		
Control 3 EC	0.099%	09:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 12/11/2005