

ALCOTEST CHECKLIST

Municipality: STRATFORD

Alcotest Ser. #: ARNK-0039

County: CAMDEN

Date of Calibration: 12-12-2025

- ☒ 1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
- ☒ 2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDUL 53-0163
- ☒ 3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDUJ P2-030
- ☒ 4. **Digital Temperature Measuring System Report of Calibration.**
Ser. #: 240 418 745
- ☒ 5. **NIST - Traceable Digital Thermometer Readings.**
- ☒ 6. ☒ A. Alcotest 7110 Calibration Record
☒ B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
☒ C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
☒ D. **Alcotest Card** of operator/coordinator who performed tests.
☒ E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDEF P2-081
- ☒ 7. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:
- | | | |
|--|-----------------|--------------|
| <input checked="" type="checkbox"/> A. | 0.04% Solution. | <u>24230</u> |
| <input checked="" type="checkbox"/> B. | 0.08% Solution. | <u>24240</u> |
| <input checked="" type="checkbox"/> C. | 0.10% Solution. | <u>24220</u> |
| <input checked="" type="checkbox"/> D. | 0.16% Solution. | <u>24250</u> |
- ☒ 8. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
- | | | |
|--|--|---------------------|
| <input checked="" type="checkbox"/> A. | 0.04% used in Calibration/Linearity Testing | <u>DDCB-0002</u> |
| <input checked="" type="checkbox"/> B. | 0.08% used in Calibration/Linearity Testing. | <u>DDCB-0001</u> |
| <input checked="" type="checkbox"/> C. | 0.10% used in Calibration/Linearity Testing.
[Same as CU34 unit on instrument.] | |
| <input checked="" type="checkbox"/> D. | 0.16% used in Calibration/Linearity Testing | <u>DDSC 53-0013</u> |
- ☒ 9. ☒ A. New Standard Solution Report following Calibration.
☒ B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
☒ C. Certificate of Analysis 0.10% solution for same.
Lot #: 25080
☒ D. Alcotest card of operator/coordinator who completed change.

12-14-2025

Dräger

Alcotest® 7110 MKII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest 7110-C is compliant with "driller" and "reynolds" EPT with 49 FR 48884, 49 FR 48884 and 50 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

1-19-17

SERIAL NUMBER:

ARKK-0039

Dräger Safety Diagnostics, Inc.

BC

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ☒ Model: ALCOTEST CU34
☐ Model: MARK IIA
☐ X-Cal 2000 (Alcosim)
☐ Other: _____

Serial Number:

DDuLS3-0163

Certification Date:

10-8-25

Technician:

DH / QR

Re-Certification Due Date:

10-8-26

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDuJP2-030

Certification Date:

10-8-25

Next Certification Due:

10-8-26

Probe Value:

99

Draeger, Inc.

DH / QR



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4000-15138511

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by : VWR International LLC Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA, 19087

Instrument Identification:

Model: 61220-601,

S/N: 240418745

Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath	93139		
Thermistor Module	A17118	08 Aug 2024	1000494698
Temperature Probe	3039	03 Aug 2024	15-E5O3Q-20-1
Temperature Probe	5401	28 Jul 2024	4500031366
Temperature Probe	5357	28 Jul 2024	4500031372
Thermistor Module	B5C344	10 Jul 2024	1000493579
Temperature Calibration Bath	B5C478		
Temperature Calibration Bath	B93537		
Temperature Calibration Bath	B96546		
Temperature Probe	5407	10 Jan 2025	4500038066
Thermistor Module	8A070	28 Aug 2024	1000495892

Certificate Information:

Technician: 420

Procedure: CAL-06

Cal Date: 21 May 2024

Cal Due Date: 21 May 2026

Test Conditions: 64.01%RH 24.08°C 1009mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C	N.A.	N.A.		0.000	0.001	Y	-0.05	0.05	0.0087	>4:1
°C	N.A.	N.A.		24.999	25.001	Y	24.949	25.049	0.0087	>4:1
°C	N.A.	N.A.		50.000	50.001	Y	49.95	50.05	0.0087	>4:1
°C	N.A.	N.A.		99.997	100.002	Y	99.947	100.047	0.0087	>4:1

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement : (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min=As Left Nominal(Rounded) - Tolerance; Max= As Left Nominal(Rounded) + Tolerance;

Nicol Rodriguez

Nicol Rodriguez, Quality Manager

Note :

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometer change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2017 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.
International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4000-15138511

Traceable® Certificate of Calibration for Digital Thermometer

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

Issue Date : 21 May 2024

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2017 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.
International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).

Alcotest 7110 MKIII-C Calibration
NIST-Traceable Digital Thermometer Readings

Coordinator:

Tpr. I Joshua B. Hall
Name

7750
Badge No.

Location:

Stratford Police Dept.
Agency

ARNK-0039
Alcotest Serial No.

Equipment:

240 418 745
Digital NIST Temperature Measuring System Serial No.

Simulator Solution Concentration	CU-34 Simulator Serial No.	Time Simulators Started to Heat	Time Temp. Reading Obtained	Temp. Reading on NIST Traceable Thermometer
0.04%	DDCB-0002	07:30 S	08:31 S	34.0 °C
0.08%	DDCB-0001	07:30 S	08:32 S	33.9 °C
0.10%	DDWL S3-0163	07:30 S	08:33 S	34.0 °C
0.16%	PDSC S3-0013	07:30 S	08:34 S	33.8 °C

Pursuant to law and the "Chemical Breath Testing Regulations" established at N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity and consistent with the "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on Alcotest 7110 MKIII-C instruments. Pursuant to and consistent with the current "Calibration Check Procedure for Alcotest 7110", I performed a Calibration Check Procedure on the Alcotest 7110 MKIII-C instrument identified on this certificate. Pursuant to the current "Calibration Check Procedure for Alcotest 7110", I used the Digital NIST-traceable Temperature Measuring System identified on this certificate to confirm that the temperatures of the 0.10%, 0.04%, 0.08%, and 0.16% Simulator Solutions used in the respective CU-34 Simulators identified on this certificate, were 34.0 degrees Celsius \pm 0.2 degrees Celsius. I hereby certify that I truthfully recorded on this certificate the temperatures of each of the simulator solutions as shown on the Digital NIST-traceable Temperature Measuring System thermometer. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment.

[Signature] #7750
Coordinator's Signature


12/12/2025
Date


Alcotest 7110 Calibration Record


Equipment


Location:	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Calibration File No.:	03309	Calib. No.:	00057
Certification File No.:	03285	Cert. No.:	00046
Linearity File No.:	03286	Lin. No.:	00045
Solution File No.:	03306	Soln. No.:	00416
Sequential File No.:	03309	File Date:	12/12/2025
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	24220	Expires:	06/18/2026
		Bottle No.:	0975

Coordinator

Last Name:	HALL	First Name:	JOSHUA	MI:	B
Signature:				Badge No.:	7750
				Date:	12/12/2025

*Black Key Temperature Probe Serial.....# DDEE P2-081 

*Digital NIST Temperature Measuring System Serial.....# 240 418 745 



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests


Equipment

Alcotest 7110 MKIII-C
Location: STRATFORD POLICE DEPT.
Serial No.: ARNK-0039
Calibration File No.: 03309 Calib. Date: 12/12/2025 Calib. No.: 00057
Certification File No.: 03310 Cert. Date: 12/12/2025 Cert. No.: 00047
Linearity File No.: 03286 Lin. Date: 07/18/2025 Lin. No.: 00045
Solution File No.: 03306 Soln. Date: 11/26/2025 Soln. No.: 00416
Sequential File No.: 03310 File Date: 12/12/2025
Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 06/18/2026
Solution Control Lot: 24220 Bottle No.: 0975

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:45S		
Control 1 EC	0.099%	08:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		
Control 2 EC	0.098%	08:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 3 EC	0.098%	08:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		

All tests within acceptable tolerance.

Coordinator

Last Name: HALL First Name: JOSHUA MI: B
Signature:  Badge No.: 7750
Date: 12/12/2025

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment

Location: Alcotest 7110 MKIII-C
STRATFORD POLICE DEPT.

Serial No.: ARNK-0039

Calibration File No.: 03309 Calib. Date: 12/12/2025
Certification File No.: 03310 Cert. Date: 12/12/2025
Linearity File No.: 03311 Lin. Date: 12/12/2025
Solution File No.: 03306 Soln. Date: 11/26/2025
Sequential File No.: 03311 File Date: 12/12/2025

Calib. No.: 00057
Cert. No.: 00047
Lin. No.: 00046
Soln. No.: 00416

Calibrating Unit: WET
Control Solution %: 0.040%
Solution Control Lot: 24230

Model No.: CU-34

Serial No.: DDCB-0002
Expires: 06/25/2026
Bottle No.: 1168

Calibrating Unit: WET
Control Solution %: 0.080%
Solution Control Lot: 24240

Model No.: CU-34

Serial No.: DDCB-0001
Expires: 07/02/2026
Bottle No.: 0213

Calibrating Unit: WET
Control Solution %: 0.160%
Solution Control Lot: 24250

Model No.: CU-34

Serial No.: DDSC S3-0013
Expires: 07/17/2026
Bottle No.: 0781

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:09S		
Control 1 EC	0.041%	09:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.041%	09:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:11S		
Control 2 EC	0.041%	09:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.039%	09:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14S		
Control 3 EC	0.081%	09:15S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.080%	09:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16S		
Control 4 EC	0.080%	09:17S	33.9°C	*** TEST PASSED ***
Control 4 IR	0.080%	09:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18S		
Control 5 EC	0.159%	09:19S	33.9°C	*** TEST PASSED ***
Control 5 IR	0.159%	09:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21S		
Control 6 EC	0.157%	09:22S	33.9°C	*** TEST PASSED ***
Control 6 IR	0.159%	09:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:24S		

All tests within acceptable tolerance.

Coordinator

Last Name: HALL

First Name: JOSHUA

MI: B

Signature: 

Badge No.: 7750

Date: 12/12/2025

DEPARTMENT OF
Traffic and Public Safety
This is to certify that
Joshua B. Hall
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**
A METHOD TO DETERMINE INTOXICATION
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS **21st** DAY OF **December**
TWO THOUSAND AND SEVENTEEN

[Signature]
ACTING SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 8/22/19	NJSP Galloway	B. Hall
2. 1/12/21	ACCC	B. Hall
3. 10-3-23	ACCC	B. Hall
4. 5/28/25	BL ESTC	B. Hall
5.		
6.		
7.		
8.		
9.		

S.P. 293B (Rev. 11/17)

DEPARTMENT OF
Traffic and Public Safety
This is to certify that
Joshua B. Hall
Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**
A METHOD TO DETERMINE INTOXICATION
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS **29th** DAY OF **March**
TWO THOUSAND AND TWENTY TWO

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 293B (Rev. 02/22)



NEW JERSEY STATE POLICE
Alcohol Drug Testing Unit

Tpr. II Joshua B. Hall #7750

Hamilton Technology Complex
Suite #400
1200 Negron Drive
Hamilton, NJ 08691

Tel#:(609)584-5051 ext 5608
Email:Joshua.Hall@njsp.gov
Cell# (609) 947-3179



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY

Governor

TAHESHA L. WAY

Lt. Governor

MATTHEW J. PLATKIN

Attorney General

COLONEL PATRICK J. CALLAHAN

Superintendent

CERTIFICATION OF ANALYSIS **0.040 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0469 to 0.0499 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/23/2024

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 24230

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0479 to 0.0484 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 25, 2026.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 30 day of July, 2024.

Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 8/13/2024

NOTARY PUBLIC NEW JERSEY
Commission # 50110522
My Commission Expires 8/13/2024



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS **0.080 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0939 to 0.0997 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/24/2024

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 24240

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0957 to 0.0973 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 2, 2026.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 31 day of July, 2024.
Notary [Signature]

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 8/13/2024



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS **0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/22/2024

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 24220

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1201 to 0.1223 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 18, 2026.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 29 day of July, 2024.
[Signature]
Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 8/13/2024



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS 0.160 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1878 to 0.1994 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.


ANALYSIS DATE: 08/07/2024


BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 24250

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1924 to 0.1949 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 17, 2026.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Michael Kennedy
OFS Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 9 day of August, 2024.

Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ☒ Model: ALCOTEST CU34
☐ Model: MARK IIA
☐ X-Cal 2000 (Alcosim)
☐ Other: _____

Serial Number:

DDCB - 0002

Certification Date:

5-16-25

Technician:

DH / QR

Re-Certification Due Date:

5-16-26

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ☒ Model: ALCOTEST CU34
☐ Model: MARK IIA
☐ X-Cal 2000 (Alcosim)
☐ Other: _____

Serial Number:

DDCB - 0001

Certification Date:

5-16-25

Technician:

DH / QR

Re-Certification Due Date:

5-16-26

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ☒ Model: ALCOTEST CU34
☐ Model: MARK IIA
☐ X-Cal 2000 (Alcosim)
☐ Other: _____

Serial Number:

DDSCS3-0013

Certification Date:

5-16-25

Technician:

DH / QR

Re-Certification Due Date:

5-16-26

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDEEP2-081

Certification Date:

5-16-25

Next Certification Due:

5-16-26

Probe Value:

104

Draeger, Inc.

DH / QR

Calibrating Unit

New Standard Solution Report


Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03309 **Calib. Date:** 12/12/2025 **Calib. No.:** 00057
Certification File No.: 03310 **Cert. Date:** 12/12/2025 **Cert. No.:** 00047
Linearity File No.: 03311 **Lin. Date:** 12/12/2025 **Lin. No.:** 00046
Solution File No.: 03312 **Soln. Date:** 12/12/2025 **Soln. No.:** 00417
Sequential File No.: 03312 **File Date:** 12/12/2025


Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0163
Control Solution %: 0.100% **Expires:** 03/04/2027
Solution Control Lot: 25080 **Bottle No.:** 0032

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:32S		
Control 1 EC	0.100%	10:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33S		
Control 2 EC	0.099%	10:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S		
Control 3 EC	0.099%	10:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUL P2-030 

Changed By:
Last Name: HALL First Name: JOSHUA MI: B
Signature:  #7750 Badge No.: 7750
Date: 12/12/2025

Temperature Probe Serial Number:



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS **0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 03/31/2025

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 25080


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1208 to 0.1220 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 04, 2027.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 2 day of April, 2025.
[Signature]
Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable

