

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 01131 Calib. Date: 11/11/2024 Calib. No.: 00046
Certification File No.: 01132 Cert. Date: 11/11/2024 Cert. No.: 00041
Linearity File No.: 01133 Lin. Date: 11/11/2024 Lin. No.: 00043
Solution File No.: 01140 Soln. Date: 03/14/2025 Soln. No.: 00292
Sequential File No.: 01140 File Date: 03/14/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 06/05/2025
Solution Control Lot: 23200 Bottle No.: 1022

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:42S		
Control 1 EC	0.104%	10:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43S		
Control 2 EC	0.103%	10:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	10:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44S		
Control 3 EC	0.102%	10:45S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	10:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sgt. D. Davis #2753

Badge No.: 2753

Date: 03/14/2025

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 01131 Calib. Date: 11/11/2024 Calib. No.: 00046
Certification File No.: 01132 Cert. Date: 11/11/2024 Cert. No.: 00041
Linearity File No.: 01133 Lin. Date: 11/11/2024 Lin. No.: 00043
Solution File No.: 01139 Soln. Date: 02/14/2025 Soln. No.: 00291
Sequential File No.: 01139 File Date: 02/14/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 06/05/2025
Solution Control Lot: 23200 Bottle No.: 1171

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:52S		
Control 1 EC	0.103%	21:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:53S		
Control 2 EC	0.101%	21:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	21:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:54S		
Control 3 EC	0.102%	21:55S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	21:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sgt. D. Davis #2753

Badge No.: 2753

Date: 02/14/2025

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01131	Calib. Date: 11/11/2024
Certification File No.:	01132	Calib. No.: 00046
Linearity File No.:	01133	Cert. Date: 11/11/2024
Solution File No.:	01138	Cert. No.: 00041
Sequential File No.:	01138	Lin. Date: 11/11/2024
		Lin. No.: 00043
		Soln. Date: 01/13/2025
		Soln. No.: 00290
		File Date: 01/13/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 0973

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:22S		
Control 1 EC	0.103%	11:22S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23S		
Control 2 EC	0.101%	11:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24S		
Control 3 EC	0.101%	11:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORROTTO	First Name: CHRISTOPHER	MI: M
Signature: <u></u>	Badge No.: 2768	
	Date: 01/13/2025	