

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03238 Calib. Date: 11/18/2024 Calib. No.: 00054
Certification File No.: 03239 Cert. Date: 11/18/2024 Cert. No.: 00044
Linearity File No.: 03240 Lin. Date: 11/18/2024 Lin. No.: 00043
Solution File No.: 03254 Soln. Date: 01/10/2025 Soln. No.: 00403
Sequential File No.: 03254 File Date: 01/10/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 09/12/2025
Solution Control Lot: 23330 Bottle No.: 0560

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:24S		
Control 1 EC	0.101%	19:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:26S		
Control 2 EC	0.100%	19:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:27S		
Control 3 EC	0.100%	19:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3204

Date: 01/10/2025