

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03309	Calib. Date: 12/12/2025
Certification File No.:	03310	Calib. No.: 00057
Linearity File No.:	03311	Cert. Date: 12/12/2025
Solution File No.:	03321	Cert. No.: 00047
Sequential File No.:	03321	Lin. Date: 12/12/2025
		Lin. No.: 00046
		Soln. Date: 01/30/2026
		Soln. No.: 00419
		File Date: 01/30/2026
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	25080	Expires: 03/04/2027
		Bottle No.: 0250

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:45S		
Control 1 EC	0.099%	09:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47S		
Control 2 EC	0.098%	09:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48S		
Control 3 EC	0.096%	09:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3203

Date: 01/30/2026

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03309 **Calib. Date:** 12/12/2025 **Calib. No.:** 00057
Certification File No.: 03310 **Cert. Date:** 12/12/2025 **Cert. No.:** 00047
Linearity File No.: 03311 **Lin. Date:** 12/12/2025 **Lin. No.:** 00046
Solution File No.: 03316 **Soln. Date:** 01/12/2026 **Soln. No.:** 00418
Sequential File No.: 03316 **File Date:** 01/12/2026

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0163
Control Solution %: 0.100% **Expires:** 03/04/2027
Solution Control Lot: 25080 **Bottle No.:** 0411

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:01S		
Control 1 EC	0.098%	03:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:03S		
Control 2 EC	0.098%	03:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:04S		
Control 3 EC	0.098%	03:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3203

Date: 01/12/2026