

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0038
Location:	EASTAMPTON TOWNSHIP	
Calibration File No.:	00808	Calib. Date: 11/26/2024
Certification File No.:	00809	Cert. Date: 11/26/2024
Linearity File No.:	00810	Lin. Date: 11/26/2024
Solution File No.:	00817	Soln. Date: 01/28/2025
Sequential File No.:	00817	File Date: 01/28/2025

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUK S3-0130
Control Solution %:	0.100%		Expires: 07/23/2026
Solution Control Lot:	24260		Bottle No.: 0181

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:34S		
Control 1 EC	0.100%	17:35S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:35S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:36S		
Control 2 EC	0.099%	17:36S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:36S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:37S		
Control 3 EC	0.100%	17:38S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:38S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIPPIN

First Name: MATTHEW

MI: F

Signature: _____

Badge No.: 3411

Date: 01/28/2025