

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
Location: WESTAMPTON POLICE  
Calibration File No.: 01131 Calib. Date: 11/11/2024 Calib. No.: 00046  
Certification File No.: 01132 Cert. Date: 11/11/2024 Cert. No.: 00041  
Linearity File No.: 01133 Lin. Date: 11/11/2024 Lin. No.: 00043  
Solution File No.: 01136 Soln. Date: 12/13/2024 Soln. No.: 00289  
Sequential File No.: 01136 File Date: 12/13/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117  
Control Solution %: 0.100% Expires: 06/05/2025  
Solution Control Lot: 23200 Bottle No.: 1160

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:16S		
Control 1 EC	0.102%	06:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	06:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:17S		
Control 2 EC	0.100%	06:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	06:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:19S		
Control 3 EC	0.101%	06:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	06:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sgt. D. Davis

Badge No.: 2753

Date: 12/13/2024

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
Location: WESTAMPTON POLICE  
Calibration File No.: 01115 Calib. Date: 06/13/2024 Calib. No.: 00045  
Certification File No.: 01116 Cert. Date: 06/13/2024 Cert. No.: 00040  
Linearity File No.: 01117 Lin. Date: 06/13/2024 Lin. No.: 00042  
Solution File No.: 01127 Soln. Date: 10/16/2024 Soln. No.: 00287  
Sequential File No.: 01127 File Date: 10/16/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116  
Control Solution %: 0.100% Expires: 06/05/2025  
Solution Control Lot: 23200 Bottle No.: 1025

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:16D		
Control 1 EC	0.102%	00:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	00:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:17D		
Control 2 EC	0.101%	00:18D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18D		
Control 3 EC	0.101%	00:19D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sgt. D. Davis #2753

Badge No.: 2753

Date: 10/16/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01115	Calib. Date: 06/13/2024
Certification File No.:	01116	Calib. No.: 00045
Linearity File No.:	01117	Cert. Date: 06/13/2024
Solution File No.:	01126	Cert. No.: 00042
Sequential File No.:	01126	Lin. Date: 06/13/2024
		Lin. No.: 00042
		Soln. Date: 09/15/2024
		Soln. No.: 00286
		File Date: 09/15/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 1020

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:02D		
Control 1 EC	0.100%	21:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 2 EC	0.100%	21:04D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		
Control 3 EC	0.100%	21:05D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Det. D. Davis #2753

Badge No.: 2753

Date: 09/15/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01115	Calib. Date: 06/13/2024
Certification File No.:	01116	Calib. No.: 00045
Linearity File No.:	01117	Cert. Date: 06/13/2024
Solution File No.:	01123	Cert. No.: 00040
Sequential File No.:	01123	Lin. Date: 06/13/2024
		Lin. No.: 00042
		Soln. Date: 08/08/2024
		Soln. No.: 00285
		File Date: 08/08/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 1180

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:42D		
Control 1 EC	0.098%	15:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	15:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44D		
Control 2 EC	0.096%	15:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	15:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45D		
Control 3 EC	0.097%	15:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sgt. D Davis #2753

Badge No.: 2753

Date: 08/08/2024

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01115	Calib. Date: 06/13/2024
Certification File No.:	01116	Calib. No.: 00045
Linearity File No.:	01117	Cert. Date: 06/13/2024
Solution File No.:	01121	Cert. No.: 00040
Sequential File No.:	01121	Lin. Date: 06/13/2024
		Lin. No.: 00042
		Soln. Date: 07/10/2024
		Soln. No.: 00284
		File Date: 07/10/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 1023

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:00D		
Control 1 EC	0.099%	22:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:01D		
Control 2 EC	0.098%	22:02D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03D		
Control 3 EC	0.098%	22:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:04D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature:  #2753

Badge No.: 2753

Date: 07/10/2024

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
Location: WESTAMPTON POLICE  
Calibration File No.: 01097 Calib. Date: 12/27/2023 Calib. No.: 00044  
Certification File No.: 01098 Cert. Date: 12/27/2023 Cert. No.: 00039  
Linearity File No.: 01099 Lin. Date: 12/27/2023 Lin. No.: 00041  
Solution File No.: 01113 Soln. Date: 06/08/2024 Soln. No.: 00282  
Sequential File No.: 01113 File Date: 06/08/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117  
Control Solution %: 0.100% Expires: 06/05/2025  
Solution Control Lot: 23200 Bottle No.: 1460

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:50D		
Control 1 EC	0.099%	11:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		
Control 2 EC	0.097%	11:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53D		
Control 3 EC	0.097%	11:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sci. Dennis Davis #2753

Badge No.: 2753

Date: 06/08/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01097	Calib. Date: 12/27/2023
Certification File No.:	01098	Cert. No.: 00044
Linearity File No.:	01099	Cert. Date: 12/27/2023
Solution File No.:	01110	Lin. No.: 00039
Sequential File No.:	01110	Lin. Date: 12/27/2023
		Soln. No.: 00041
		Soln. Date: 05/08/2024
		File Date: 05/08/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 1185

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:15D		
Control 1 EC	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16D		
Control 2 EC	0.099%	02:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:18D		
Control 3 EC	0.099%	02:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: \_\_\_\_\_



Badge No.: 2753

Date: 05/08/2024

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
Location: WESTAMPTON POLICE  
Calibration File No.: 01097 Calib. Date: 12/27/2023 Calib. No.: 00044  
Certification File No.: 01098 Cert. Date: 12/27/2023 Cert. No.: 00039  
Linearity File No.: 01099 Lin. Date: 12/27/2023 Lin. No.: 00041  
Solution File No.: 01107 Soln. Date: 04/04/2024 Soln. No.: 00280  
Sequential File No.: 01107 File Date: 04/04/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117  
Control Solution %: 0.100% Expires: 06/05/2025  
Solution Control Lot: 23200 Bottle No.: 1481

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:17S		
Control 1 EC	0.099%	18:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:18S		
Control 2 EC	0.098%	18:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:19S		
Control 3 EC	0.098%	18:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Scpt. D. Davis #2753

Badge No.: 2753

Date: 04/04/2024



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01097	Calib. Date: 12/27/2023
Certification File No.:	01098	Calib. No.: 00044
Linearity File No.:	01099	Cert. Date: 12/27/2023
Solution File No.:	01105	Cert. No.: 00039
Sequential File No.:	01105	Lin. Date: 12/27/2023
		Lin. No.: 00041
		Soln. Date: 03/02/2024
		Soln. No.: 00279
		File Date: 03/02/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 1397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:49S		
Control 1 EC	0.100%	05:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:51S		
Control 2 EC	0.099%	05:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:52S		
Control 3 EC	0.099%	05:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: 

Badge No.: 2753

Date: 03/02/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	01097	Calib. Date: 12/27/2023
Certification File No.:	01098	Calib. No.: 00044
Linearity File No.:	01099	Cert. Date: 12/27/2023
Solution File No.:	01103	Cert. No.: 00039
Sequential File No.:	01103	Lin. Date: 12/27/2023
		Lin. No.: 00041
		Soln. Date: 01/25/2024
		Soln. No.: 00278
		File Date: 01/25/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	23200	Expires: 06/05/2025
		Bottle No.: 1008

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:25S		
Control 1 EC	0.099%	07:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		
Control 2 EC	0.097%	07:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28S		
Control 3 EC	0.098%	07:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVIS

First Name: DENNIS

MI: N

Signature: Sgt. D. Davis # 2753

Badge No.: 2753

Date: 01/25/2024