

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03238	Calib. Date: 11/18/2024
Certification File No.:	03239	Calib. No.: 00054
Linearity File No.:	03240	Cert. Date: 11/18/2024
Solution File No.:	03248	Cert. No.: 00044
Sequential File No.:	03248	Lin. Date: 11/18/2024
		Lin. No.: 00043
		Soln. Date: 12/13/2024
		Soln. No.: 00402
		File Date: 12/13/2024

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0133
Control Solution %:	0.100%			Expires:	09/12/2025
Solution Control Lot:	23330			Bottle No.:	0425

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:30S		
Control 1 EC	0.099%	02:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:31S		
Control 2 EC	0.098%	02:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:33S		
Control 3 EC	0.098%	02:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JESS		First Name: VINCENT	MI: N
Signature: <u>Sgt [Signature] #3203</u>		Badge No.: 3203	Date: 12/13/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03202	Calib. Date: 06/26/2024
Certification File No.:	03203	Calib. No.: 00053
Linearity File No.:	03204	Cert. Date: 06/26/2024
Solution File No.:	03234	Cert. No.: 00043
Sequential File No.:	03234	Lin. Date: 06/26/2024
		Lin. No.: 00042
		Soln. Date: 10/28/2024
		Soln. No.: 00400
		File Date: 10/28/2024

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 09/12/2025
Solution Control Lot: 23330		Bottle No.: 0420

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:26S		
Control 1 EC	0.097%	05:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28S		
Control 2 EC	0.096%	05:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:29S		
Control 3 EC	0.098%	05:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3204

Date: 10/28/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03202	Calib. Date: 06/26/2024
Certification File No.:	03203	Cert. No.: 00053
Linearity File No.:	03204	Cert. Date: 06/26/2024
Solution File No.:	03229	Lin. No.: 00042
Sequential File No.:	03229	Soln. No.: 00399
		Soln. Date: 09/28/2024
		File Date: 09/28/2024

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 0561

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:16D		
Control 1 EC	0.100%	04:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		
Control 2 EC	0.099%	04:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19D		
Control 3 EC	0.099%	04:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3204

Date: 09/28/2024

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.		
Calibration File No.: 03202	Calib. Date: 06/26/2024	Calib. No.: 00053
Certification File No.: 03203	Cert. Date: 06/26/2024	Cert. No.: 00043
Linearity File No.: 03204	Lin. Date: 06/26/2024	Lin. No.: 00042
Solution File No.: 03222	Soln. Date: 08/31/2024	Soln. No.: 00398
Sequential File No.: 03222	File Date: 08/31/2024	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 09/12/2025
Solution Control Lot: 23330		Bottle No.: 0312

Function	Result %BAC	Time HH:MM	Temperature Simufator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:23D		
Control 1 EC	0.101%	20:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:24D		
Control 2 EC	0.101%	20:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:26D		
Control 3 EC	0.100%	20:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3204

Date: 08/31/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03202	Calib. Date: 06/26/2024
Certification File No.:	03203	Cert. No.: 00053
Linearity File No.:	03204	Cert. Date: 06/26/2024
Solution File No.:	03221	Lin. No.: 00043
Sequential File No.:	03221	Lin. Date: 06/26/2024
		Soln. No.: 00397
		Soln. Date: 08/31/2024
		File Date: 08/31/2024

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 09/12/2025
Solution Control Lot: 23330		Bottle No.: 0481

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:56D		
Control 1 EC	0.092%	18:57D	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.091%	18:57D	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:09D		
Control 2 EC	0.095%	19:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	19:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:11D		
Control 3 EC	0.095%	19:12D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	19:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:13D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3204

Date: 08/31/2024

# Calibrating Unit

## New Standard Solution Report

### Equipment

Alcotest 7110 MKIII-C  
Location: STRATFORD POLICE DEPT.  
Serial No.: ARNK-0039  
Calibration File No.: 03202 Calib. Date: 06/26/2024 Calib. No.: 00053  
Certification File No.: 03203 Cert. Date: 06/26/2024 Cert. No.: 00043  
Linearity File No.: 03204 Lin. Date: 06/26/2024 Lin. No.: 00042  
Solution File No.: 03215 Soln. Date: 07/24/2024 Soln. No.: 00396  
Sequential File No.: 03215 File Date: 07/24/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163  
Control Solution %: 0.100% Expires: 09/12/2025  
Solution Control Lot: 23330 Bottle No.: 0531

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:40D		
Control 1 EC	0.100%	19:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:42D		
Control 2 EC	0.099%	19:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:43D		
Control 3 EC	0.099%	19:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3204

Date: 07/24/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03171	Calib. Date: 01/12/2024
Certification File No.:	03172	Calib. No.: 00052
Linearity File No.:	03173	Cert. Date: 01/12/2024
Solution File No.:	03201	Cert. No.: 00042
Sequential File No.:	03201	Lin. Date: 01/12/2024
		Lin. No.: 00041
		Soln. Date: 06/23/2024
		Soln. No.: 00394
		File Date: 06/23/2024

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUK S3-0133
Control Solution %: 0.100%		Expires: 09/12/2025
Solution Control Lot: 23330		Bottle No.: 0198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:55D		
Control 1 EC	0.100%	11:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 2 EC	0.098%	11:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		
Control 3 EC	0.098%	11:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3204

Date: 06/23/2024

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.		
Calibration File No.: 03171	Calib. Date: 01/12/2024	Calib. No.: 00052
Certification File No.: 03172	Cert. Date: 01/12/2024	Cert. No.: 00042
Linearity File No.: 03173	Lin. Date: 01/12/2024	Lin. No.: 00041
Solution File No.: 03196	Soln. Date: 05/24/2024	Soln. No.: 00393
Sequential File No.: 03196	File Date: 05/24/2024	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUK S3-0133
Control Solution %: 0.100%		Expires: 09/12/2025
Solution Control Lot: 23330		Bottle No.: 0608

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:28D		
Control 1 EC	0.104%	15:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29D		
Control 2 EC	0.101%	15:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30D		
Control 3 EC	0.101%	15:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3204

Date: 05/24/2024



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	03171	Calib. Date:	01/12/2024
Certification File No.:	03172	Cert. No.:	00052
Linearity File No.:	03173	Cert. Date:	01/12/2024
Solution File No.:	03193	Lin. No.:	00041
Sequential File No.:	03193	Soln. Date:	04/29/2024
		Soln. No.:	00392
		File Date:	04/29/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	23330	Expires:	09/12/2025
		Bottle No.:	0520

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:02D		
Control 1 EC	0.101%	08:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03D		
Control 2 EC	0.099%	08:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05D		
Control 3 EC	0.099%	08:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JESS

First Name: VINCENT

MI: N

Signature: \_\_\_\_\_



Badge No.: 3203

Date: 04/29/2024



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03171	Calib. Date: 01/12/2024
Certification File No.:	03172	Calib. No.: 00052
Linearity File No.:	03173	Cert. Date: 01/12/2024
Solution File No.:	03184	Cert. No.: 00042
Sequential File No.:	03184	Lin. Date: 01/12/2024
		Lin. No.: 00041
		Soln. Date: 03/20/2024
		Soln. No.: 00390
		File Date: 03/20/2024

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUK S3-0133
Control Solution %:	0.100%		Expires: 09/12/2025
Solution Control Lot:	23330		Bottle No.: 0638

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:03S		
Control 1 EC	0.099%	22:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:05S		
Control 2 EC	0.098%	22:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:07S		
Control 3 EC	0.099%	22:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3204

Date: 03/20/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03171	Calib. Date: 01/12/2024
Certification File No.:	03172	Calib. No.: 00052
Linearity File No.:	03173	Cert. Date: 01/12/2024
Solution File No.:	03181	Cert. No.: 00042
Sequential File No.:	03181	Lin. Date: 01/12/2024
		Lin. No.: 00041
		Soln. Date: 02/25/2024
		Soln. No.: 00389
		File Date: 02/25/2024

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 0261

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:58S		
Control 1 EC	0.099%	00:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:59S		
Control 2 EC	0.097%	01:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:01S		
Control 3 EC	0.098%	01:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3204

Date: 02/25/2024

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03171	Calib. Date: 01/12/2024
Certification File No.:	03172	Calib. No.: 00052
Linearity File No.:	03173	Cert. Date: 01/12/2024
Solution File No.:	03179	Cert. No.: 00042
Sequential File No.:	03179	Lin. Date: 01/12/2024
		Lin. No.: 00041
		Soln. Date: 01/29/2024
		Soln. No.: 00388
		File Date: 01/29/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 0318

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:25S		
Control 1 EC	0.098%	02:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27S		
Control 2 EC	0.097%	02:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28S		
Control 3 EC	0.098%	02:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3204

Date: 01/29/2024

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 03137 Calib. Date: 08/01/2023 Calib. No.: 00051  
Certification File No.: 03138 Cert. Date: 08/01/2023 Cert. No.: 00041  
Linearity File No.: 03139 Lin. Date: 08/01/2023 Lin. No.: 00040  
Solution File No.: 03170 Soln. Date: 01/07/2024 Soln. No.: 00386  
Sequential File No.: 03170 File Date: 01/07/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 0956

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:37S		
Control 1 EC	0.102%	23:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:39S		
Control 2 EC	0.101%	23:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:40S		
Control 3 EC	0.100%	23:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3204

Date: 01/07/2024