

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
Location: MERCHANTVILLE POLICE  
Calibration File No.: 01192 Calib. Date: 12/27/2024 Calib. No.: 00052  
Certification File No.: 01193 Cert. Date: 12/27/2024 Cert. No.: 00045  
Linearity File No.: 01194 Lin. Date: 12/27/2024 Lin. No.: 00045  
Solution File No.: 01198 Soln. Date: 01/25/2025 Soln. No.: 00350  
Sequential File No.: 01198 File Date: 01/25/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012  
Control Solution %: 0.100% Expires: 01/09/2026  
Solution Control Lot: 24020 Bottle No.: 0678

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:01S		
Control 1 EC	0.099%	20:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03S		
Control 2 EC	0.099%	20:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:04S		
Control 3 EC	0.099%	20:05S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: REGALBUTO

First Name: AUSTIN

MI: -

Signature: 

#78

Badge No.: 78

Date: 01/25/2025