

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	01229	Calib. Date:	10/01/2025
Certification File No.:	01230	Cert. Date:	10/01/2025
Linearity File No.:	01231	Lin. Date:	10/01/2025
Solution File No.:	01241	Soln. Date:	01/11/2026
Sequential File No.:	01241	File Date:	01/11/2026
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	24370	Expires:	10/08/2026
		Bottle No.:	0156

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:11S		
Control 1 EC	0.099%	08:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		
Control 2 EC	0.098%	08:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		
Control 3 EC	0.098%	08:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REGALBUTO

First Name: AUSTIN

MI: -

Signature: _____

Badge No.: 78

Date: 01/11/2026