

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	01131	Calib. Date:	11/11/2024
Certification File No.:	01132	Cert. No.:	00046
Linearity File No.:	01133	Cert. Date:	11/11/2024
Solution File No.:	01138	Lin. No.:	00041
Sequential File No.:	01138	Lin. Date:	11/11/2024
		Soln. No.:	00043
		Soln. Date:	01/13/2025
		File Date:	01/13/2025
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0117
Solution Control Lot:	23200	Expires:	06/05/2025
		Bottle No.:	0973

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:22S		
Control 1 EC	0.103%	11:22S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23S		
Control 2 EC	0.101%	11:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24S		
Control 3 EC	0.101%	11:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORROTTO	First Name: CHRISTOPHER	MI: M
Signature: <u></u>	Badge No.: 2768	
	Date: 01/13/2025	