

ALCOHOL INFLUENCE REPORT FORM, ALCOTEST 9510
Longport Borough

Sequential File No.: 50
Department Case No.: 25-005159

Subject

Last Name: TEST -
D.O.B.: 06/10/2000 Age: 25
Driver License Number: TEST

First Name: TEST
Gender: MALE

MI: T
Ht: 6 ft. 9 in.
Issuing State: NA

Arresting Officer

Last Name: TEST -
Badge No.: 000
Municipal Code/Arrest Location: 0115

First Name: TEST

Arrest Date: 06/17/2025 Arrest Time: 12:00

Equipment

Location: Longport Borough
Inst. Model No.: ALCOTEST 9510
Firmware: 8326739 1.5
Sequential File No.: 50
Linearity File No.: 36

Serial No.: ARMJ-0201
Config.: 8326737 3.10 WinCE: 8326738 2.9
File Date: 06/17/2025
Lin. Date: 03/24/2025 Lin. No.: 2

Active Dry Gas Cyl.: #1 (Upper) Lot #: 302-402921458 Exp. Date: 12/18/2026 Conc.: 0.100

Breath Test Information

Date of Test: 06/17/2025


Function	Result %BAC	Time hh:mm:ss	Volume (L)	Duration Sec. (s)	Barom. Pres.(hPa)	Status
Ambient Air Blank	0.000	13:52:39				
Control Test 1		13:53:13			1018	*TEST PASSED*
EC Result	0.103					
IR Result	0.101					
Ambient Air Blank	0.000	13:53:53				
Breath Sample 1		13:54:52	3.3	17.0		*TEST PASSED*
EC Result	0.000					
IR Result	0.000					
Ambient Air Blank	0.000	13:55:24				
Ambient Air Blank	0.000	13:57:55				
Breath Sample 2		13:58:35	2.6	11.2		*TEST PASSED*
EC Result	0.000					
IR Result	0.000					
Ambient Air Blank	0.000	13:58:57				
Control Test 2		13:59:26			1018	*TEST PASSED*
EC Result	0.103					
IR Result	0.102					
Ambient Air Blank	0.000	14:00:06				

REPORTED BREATH TEST RESULT: 0.000 %BAC

Breath Test Operator

Last Name: Wright -
Agency: 0115

First Name: Quinton
Badge No.: 333



Signature:

Date: 06/17/2025

ALCOHOL INFLUENCE REPORT FORM, ALCOTEST 9510
Longport Borough

Sequential File No.: 49
Department Case No.: 25-005159

Subject

Last Name: TEST -
D.O.B.: 10/10/2015 Age: 9
Driver License Number: TEST

First Name: TEST
Gender: MALE

MI: T
Ht: 5 ft. 5 in.
Issuing State: TE

Arresting Officer

Last Name: TEST -
Badge No.: 333
Municipal Code/Arrest Location: 0115

First Name: TEST
Arrest Date: 06/17/2025 Arrest Time: 12:00

Equipment

Location: Longport Borough
Inst. Model No.: ALCOTEST 9510
Firmware: 8326739 1.5
Sequential File No.: 49
Linearity File No.: 36

Serial No.: ARMJ-0201
Config.: 8326737 3.10 WinCE: 8326738 2.9
File Date: 06/17/2025
Lin. Date: 03/24/2025 Lin. No.: 2

Active Dry Gas Cyl.: #1 (Upper) Lot #: 302-402921458 Exp. Date: 12/18/2026 Conc.: 0.100

Breath Test Information

Date of Test: 06/17/2025

Function	Result %BAC	Time hh:mm:ss	Volume (L)	Duration Sec. (s)	Barom. Pres.(hPa)	Status
Ambient Air Blank	0.000	13:43:07				
Control Test 1		13:43:42			1018	*TEST PASSED*
EC Result	0.104					
IR Result	0.101					
Ambient Air Blank	0.000	13:44:22				
Breath Sample 1		13:45:16	2.3	11.6		*TEST PASSED*
EC Result	0.000					
IR Result	0.000					
Ambient Air Blank	0.000	13:45:39				
Ambient Air Blank	-----	--:--:--				
Breath Sample 2		--:--:--	-----	-----		
EC Result	-----					
IR Result	-----					
Ambient Air Blank	-----	--:--:--				
Control Test 2		13:46:42			1018	*TEST PASSED*
EC Result	0.103					
IR Result	0.101					
Ambient Air Blank	0.000	13:47:22				

TEST RESULT: BLOWING NOT ALLOWED

Breath Test Operator

Last Name: Wright -
Agency: 0115

First Name: Quinton
Badge No.: 333



Signature:

Date: 06/17/2025