

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: STRATFORD POLICE DEPT.
Serial No.: ARNK-0039
Calibration File No.: 03284 Calib. Date: 07/18/2025 Calib. No.: 00056
Certification File No.: 03285 Cert. Date: 07/18/2025 Cert. No.: 00046
Linearity File No.: 03286 Lin. Date: 07/18/2025 Lin. No.: 00045
Solution File No.: 03292 Soln. Date: 08/11/2025 Soln. No.: 00412
Sequential File No.: 03292 File Date: 08/11/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 03/04/2027
Solution Control Lot: 25080 Bottle No.: 0149

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:21D		
Control 1 EC	0.097%	02:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:23D		
Control 2 EC	0.097%	02:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:24D		
Control 3 EC	0.097%	02:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 08/11/2025

Calibrating Unit

New Standard Solution Report

Equipment

Location: Alcotest 7110 MKIII-C

Serial No.: ARNK-0039

Calibration File No.: 03267

Calib. Date: 03/14/2025

Calib. No.: 00055

Certification File No.: 03268

Cert. Date: 03/14/2025

Cert. No.: 00045

Linearity File No.: 03269

Lin. Date: 03/14/2025

Lin. No.: 00044

Solution File No.: 03281

Soln. Date: 07/03/2025

Soln. No.: 00410

Sequential File No.: 03281

File Date: 07/03/2025

Calibrating Unit: WET

Model No.: CU-34

Serial No.: DDUK S3-0133

Control Solution %: 0.100%

Expires: 09/12/2025

Solution Control Lot: 23330

Bottle No.: 0507

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:42D		
Control 1 EC	0.099%	19:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:44D		
Control 2 EC	0.097%	19:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:45D		
Control 3 EC	0.098%	19:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3204

Date: 07/03/2025

Calibrating Unit

New Standard Solution Report

Equipment

Location:

Alcotest 7110 MKIII-C
STRATFORD POLICE DEPT.

Serial No.: ARNK-0039

Calibration File No.: 03267

Calib. Date: 03/14/2025

Calib. No.: 00055

Certification File No.: 03268

Cert. Date: 03/14/2025

Cert. No.: 00045

Linearity File No.: 03269

Lin. Date: 03/14/2025

Lin. No.: 00044

Solution File No.: 03279

Soln. Date: 06/02/2025

Soln. No.: 00409

Sequential File No.: 03279

File Date: 06/02/2025

Calibrating Unit: WET

Model No.: CU-34

Serial No.: DDUK S3-0133

Control Solution %: 0.100%

Expires: 09/12/2025

Solution Control Lot: 23330

Bottle No.: 0609

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:26D		
Control 1 EC	0.100%	22:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27D		
Control 2 EC	0.099%	22:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:29D		
Control 3 EC	0.099%	22:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3204

Date: 06/02/2025

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03267 **Calib. Date:** 03/14/2025 **Calib. No.:** 00055
Certification File No.: 03268 **Cert. Date:** 03/14/2025 **Cert. No.:** 00045
Linearity File No.: 03269 **Lin. Date:** 03/14/2025 **Lin. No.:** 00044
Solution File No.: 03276 **Soln. Date:** 05/09/2025 **Soln. No.:** 00408
Sequential File No.: 03276 **File Date:** 05/09/2025

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUK S3-0133
Control Solution %: 0.100% **Expires:** 09/12/2025
Solution Control Lot: 23330 **Bottle No.:** 0508

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:51D		
Control 1 EC	0.099%	06:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	06:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:52D		
Control 2 EC	0.098%	06:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		
Control 3 EC	0.097%	06:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:55D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3204

Date: 05/09/2025

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03267 Calib. Date: 03/14/2025 Calib. No.: 00055
Certification File No.: 03268 Cert. Date: 03/14/2025 Cert. No.: 00045
Linearity File No.: 03269 Lin. Date: 03/14/2025 Lin. No.: 00044
Solution File No.: 03273 Soln. Date: 04/08/2025 Soln. No.: 00407
Sequential File No.: 03273 File Date: 04/08/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 09/12/2025
Solution Control Lot: 23330 Bottle No.: 0253

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:18D		
Control 1 EC	0.100%	06:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:19D		
Control 2 EC	0.099%	06:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		
Control 3 EC	0.097%	06:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3204

Date: 04/08/2025

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: STRATFORD POLICE DEPT.
Serial No.: ARNK-0039
Calibration File No.: 03238 Calib. Date: 11/18/2024 Calib. No.: 00054
Certification File No.: 03239 Cert. Date: 11/18/2024 Cert. No.: 00044
Linearity File No.: 03240 Lin. Date: 11/18/2024 Lin. No.: 00043
Solution File No.: 03266 Soln. Date: 03/14/2025 Soln. No.: 00405
Sequential File No.: 03266 File Date: 03/14/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 09/12/2025
Solution Control Lot: 23330 Bottle No.: 0530

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:18S		
Control 1 EC	0.101%	01:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:20S		
Control 2 EC	0.101%	01:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	01:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22S		
Control 3 EC	0.101%	01:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	01:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:23S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3204

Date: 03/14/2025

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: STRATFORD POLICE DEPT.
Serial No.: ARNK-0039
Calibration File No.: 03238
Certification File No.: 03239
Linearity File No.: 03240
Solution File No.: 03262
Sequential File No.: 03262
Calib. Date: 11/18/2024
Cert. Date: 11/18/2024
Lin. Date: 11/18/2024
Soln. Date: 02/10/2025
File Date: 02/10/2025
Calib. No.: 00054
Cert. No.: 00044
Lin. No.: 00043
Soln. No.: 00404

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 23330
Model No.: CU-34
Serial No.: DDUK S3-0133
Expires: 09/12/2025
Bottle No.: 0374

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:42S		
Control 1 EC	0.100%	02:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44S		
Control 2 EC	0.098%	02:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45S		
Control 3 EC	0.099%	02:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 02/10/2025

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: STRATFORD POLICE DEPT.
Serial No.: ARNK-0039
Calibration File No.: 03238
Calib. Date: 11/18/2024
Calib. No.: 00054
Certification File No.: 03239
Cert. Date: 11/18/2024
Cert. No.: 00044
Linearity File No.: 03240
Lin. Date: 11/18/2024
Lin. No.: 00043
Solution File No.: 03254
Soln. Date: 01/10/2025
Soln. No.: 00403
Sequential File No.: 03254
File Date: 01/10/2025

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 23330
Model No.: CU-34
Serial No.: DDUK S3-0133
Expires: 09/12/2025
Bottle No.: 0560

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:24S		
Control 1 EC	0.101%	19:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:26S		
Control 2 EC	0.100%	19:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:27S		
Control 3 EC	0.100%	19:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 01/10/2025