

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Longport Police Department 2305 Atlantic Ave Longport NJ 08403	2. Contact: Dean Tibbitt 3. Phone Number: 609-822-2141	4. Date: 08/06/2024 5. County: Atlantic
6. Alcotest Instrument Serial Number:	7. Simulator Component Serial Number: <p style="text-align: center;">DDXC S3-0154</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">DDWAP2-177</p>

9. Reason for Service:
Simulator (DDXC S3-0154) will not heat past 25 degrees Celsius.

10. Comments:
I arrived at the Longport Police Department to perform a Simulator and Temperature Probe change. Upon preparing the Simulator and Temperature Probe for a Solution Change, I noticed that the Simulator was not heating up to the required temperature. I allowed plenty of time for the simulator to heat but it only reached 25 Degree Celsius. It should be noted that this was the first time the simulator had been used since its return from Draeger Inc.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p style="text-align: center;"><u>Tpr. II William Widener #7817</u> Name & Badge Number (Print)</p>	<p><u><i>Tpr. Wm. F. Widener #7817</i></u> Signature</p>	<p><u>08/06/2024</u> Date</p>
-------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------	-----------------------------------