

ALCOTEST CHECKLIST

Municipality: BURLINGTON TWP. Alcotest Ser. #: ARTL - 0024
County: BURLINGTON Date of Calibration: 03-02-2015

- 1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
- 2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDWL 53 - 0167
- 3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDXA Pa - 011
- 4. **Digital Temperature Measuring System Report of Calibration.**
Ser. #: 130 602 271
- 5.
 - A. Alcotest 7110 Calibration Record
 - B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
 - C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
 - D. **Alcotest Card** of operator/coordinator who performed tests.
 - E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDUN Pa - 231

- 6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:

<input checked="" type="checkbox"/> A.	0.04% Solution.	<u>13</u>	<u>I</u>	<u>120</u>
<input checked="" type="checkbox"/> B.	0.08% Solution.	<u>13</u>	<u>I</u>	<u>121</u>
<input checked="" type="checkbox"/> C.	0.10% Solution.	<u>13</u>	<u>I</u>	<u>122</u>
<input checked="" type="checkbox"/> D.	0.16% Solution.	<u>14</u>	<u>H</u>	<u>131</u>

- 7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
 - A. 0.04% used in Calibration/Linearity Testing. DDMK 53-0003
 - B. 0.08% used in Calibration/Linearity Testing. DDAE -0022
 - C. 0.10% used in Calibration/Linearity Testing.
[Same as CU34 unit on instrument.]
 - D. 0.16% used in Calibration/Linearity Testing. DDCN -0052
- 8.
 - A. New Standard Solution Report following Calibration.
 - B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
 - C. Certificate of Analysis 0.10% solution for same.
Lot #: 14 H 132
 - D. Alcotest card of operator/coordinator who completed change.

Drägersafety

Alcotest[®] 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 48 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

SERIAL NUMBER:

03/08/200

ARTK-0024

Dräger Safety Diagnostics, Inc.



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDULS3-0167

Certification Date

8-19-14

Technician

BC

Re-Certification Due Date

8-19-15

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDXA P2-011

Certification date:

8-19-14

Next Certification due:

8-19-15

Probe Value

103

Draeger Safety Diagnostics, Inc.
Technical Service Department

BC



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4000-5427014

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, LLC, Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA 19067
Instrument Identification:

Model: 61220-601 S/N: 130602271 Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-178	A45240		
Thermistor Module	A17118	2/13/14	1000332071
Temperature Probe	128	2/20/14	6-B48Z9-30-1
Temperature Calibration Bath TC-231	A79341		
Thermistor Module	A17118	2/13/14	1000332071
Temperature Probe	3039	2/20/14	6-B48Z9-1-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Module	A27129	11/08/13	1000327281
Temperature Probe	5202	11/30/14	15-B15PW-1-1
Temperature Calibration Bath TC-275	B16388		
Digital Thermometer	B16015	8/12/14	1000341987
PRT Temperature Probe	02022	8/14/15	B3812004

Certificate Information:

Technician: 88 Procedure: CAL-08 Cal Date: 10/08/13 Cal Due: 10/08/15
Test Conditions: 24.0°C 37.0 %RH 1017 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	eU	TUR
°C		N.A.		-0.001	0.000	Y	-0.051	0.049	0.013	3.8:1
°C		N.A.		25.001	25.000	Y	24.951	25.051	0.023	2.2:1
°C		N.A.		50.003	50.001	Y	49.953	50.053	0.014	3.6:1
°C		N.A.		100.001	99.994	Y	99.951	100.051	0.018	2.8:1

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the items described. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal = Standard's Reading; As Found = Instrument's Reading; In Tol = In Tolerance; Min/Max = Acceptance Range; eU = Expanded Measurement Uncertainty; TUR = Test Uncertainty Ratio; Accuracy = (Max-Min)/2; Min = As Left Nominal (Rounded) - Tolerance; Max = As Left Nominal (Rounded) + Tolerance; Date = M/D/YYYY

Rodriguez
Rodriguez, Quality Manager

Aaron Adams
Aaron Adams, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and condensation.

Recalibration:

For factory calibration and recalibration traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4466 Rex Road Friendswood, TX 77546 USA
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2008 Calibration Laboratory Accredited by (AZLA) American Association for Laboratory Accreditation, Certificate No. 1750.01
Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-012005-2008-AC-HOU-114
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA)

Alcotest 7110 Calibration Record

Equipment

Alcotest 7110 MKIII-C
Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 01008 Calib. Date: 03/02/2015 Calib. No.: 00024
Certification File No.: 00972 Cert. Date: 09/10/2014 Cert. No.: 00018
Linearity File No.: 00973 Lin. Date: 09/10/2014 Lin. No.: 00018
Solution File No.: 01007 Soln. Date: 03/02/2015 Soln. No.: 00157
Sequential File No.: 01008 File Date: 03/02/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0167
Control Solution %: 0.100% Expires: 09/17/2015
Solution Control Lot: 13I122 Bottle No.: 1224

Coordinator

Last Name: GONCALVES First Name: MICHELLE MI: L
Signature: Sgt Michelle Gons #6040 Badge No.: 6040
Date: 03/02/2015

*Black Key Temperature Probe Serial.....# DDUNP2-231 (MA)
*Digital NIST Temperature Measuring System Serial.....# 130602271 (MA)

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 01008 **Calib. Date:** 03/02/2015 **Calib. No.:** 00024
Certification File No.: 01009 **Cert. Date:** 03/02/2015 **Cert. No.:** 00019
Linearity File No.: 00973 **Lin. Date:** 09/10/2014 **Lin. No.:** 00018
Solution File No.: 01007 **Soln. Date:** 03/02/2015 **Soln. No.:** 00157
Sequential File No.: 01009 **File Date:** 03/02/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 09/17/2015
Solution Control Lot: 13I122 **Bottle No.:** 1224

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:41S		
Control 1 EC	0.100%	12:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42S		
Control 2 EC	0.100%	12:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:44S		
Control 3 EC	0.099%	12:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:45S		

All tests within acceptable tolerance.

Coordinator

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature:

Sgt. Michelle Delo #6040

Badge No.: 6040

Date: 03/02/2015

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 01008 **Calib. Date:** 03/02/2015 **Calib. No.:** 00024
Certification File No.: 01009 **Cert. Date:** 03/02/2015 **Cert. No.:** 00019
Linearity File No.: 01010 **Lin. Date:** 03/02/2015 **Lin. No.:** 00019
Solution File No.: 01007 **Soln. Date:** 03/02/2015 **Soln. No.:** 00157
Sequential File No.: 01010 **File Date:** 03/02/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDMK S3-0003
Control Solution %: 0.040% **Expires:** 09/13/2015
Solution Control Lot: 13I120 **Bottle No.:** 1131

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAE-0022
Control Solution %: 0.080% **Expires:** 09/16/2015
Solution Control Lot: 13I121 **Bottle No.:** 0131

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDCN-0052
Control Solution %: 0.160% **Expires:** 08/05/2016
Solution Control Lot: 14H131 **Bottle No.:** 1055

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:59S		
Control 1 EC	0.043%	13:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.041%	13:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01S		
Control 2 EC	0.041%	13:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	13:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04S		
Control 3 EC	0.082%	13:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.080%	13:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06S		
Control 4 EC	0.082%	13:06S	34.0°C	*** TEST PASSED ***
Control 4 IR	0.078%	13:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08S		
Control 5 EC	0.164%	13:09S	33.9°C	*** TEST PASSED ***
Control 5 IR	0.158%	13:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10S		
Control 6 EC	0.160%	13:11S	33.9°C	*** TEST PASSED ***
Control 6 IR	0.158%	13:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		

All tests within acceptable tolerance.

Coordinator

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: 

Badge No.: 6040

Date: 03/02/2015

DEPARTMENT OF
Traffic and Public Safety
 State Police

Michelle Gonzalez
 New Jersey State Police

It is declared and certified that the following person has successfully completed the course and is eligible to receive a license to drive a motor vehicle in New Jersey. The license shall be issued on or before the date of the course completion.

7110 MOKU-C

December

Twelve

[Signature]
 Commissioner
 NEW JERSEY STATE POLICE

[Signature]
 Attorney General
 NEW JERSEY STATE POLICE

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 12-3-13	GCPA	Wm Lora
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

N.J. 2008 (Rev. 02/10)

DEPARTMENT OF
Traffic and Public Safety
 State Police

Michelle Gonzalez
 New Jersey State Police

It is declared and certified that the following person has successfully completed the course and is eligible to receive a license to drive a motor vehicle in New Jersey. The license shall be issued on or before the date of the course completion.

7110 MOKU-C

March

Twelve

[Signature]
 Commissioner
 NEW JERSEY STATE POLICE

[Signature]
 Attorney General
 NEW JERSEY STATE POLICE

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

N.J. 2008 (Rev. 02/10)



CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lieutenant Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF CRIMINAL JUSTICE
PO BOX 085
TRENTON, NJ 08625-0085
TELEPHONE: (609) 984-6300

JEFFREY S. CHIESA
Attorney General

STEPHEN J. TAYLOR
Director

March 23, 2012

Col. Joseph R. Fuentes, Superintendent
Division of State Police
Division Headquarters
P.O. Box 7068
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification -
Detective I Michelle L. Goncalves #6040

Dear Col. Fuentes:


Pursuant to the provisions of *N.J.A.C. 13:51-2.1 (b) and (c)*, as adopted and promulgated under the provisions of *N.J.S.A. 39:4-50.3, 39:3-10.25 and 12:7-56*, I hereby approve Detective I Michelle L. Goncalves #6040, as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,

Jeffrey S. Chiesa
Attorney General

- c. Detective I Michelle L. Goncalves, Alcohol/Drug Test Unit, Division of State Police
DSFC Patrick Guilfooy, Unit Head, Alcohol/Drug Test Unit, Division of State Police ✓





CERTIFICATE

This is to certify that


Michelle Goncalves

has successfully completed the two day Draeger Safety Diagnostics, Inc. Alcohol Coordinator Training Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: March 14, 2012



Instructor: Hansueli Ryser

Dräger



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JOHN J. HOFFMAN
 Acting Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.040 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0469 to 0.0499 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 10/25/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 131120

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0484 to 0.0489 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 13, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 28th day of October, 2013.

Notary



"An Internationally Accredited Agency"

New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JOHN J. HOFFMAN
 Acting Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.080 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0939 to 0.0997 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 10/31/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 131121

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0959 to 0.0971 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 16, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 1st day of November, 2013.

Mary E. McLaughlin
 Notary

MARY ELIZABETH MCLAUGHLIN
 NOTARY PUBLIC OF NEW JERSEY
 MY COMMISSION EXPIRES 12-24-2013



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 887-2000

CHRIS CHRISTIE
 Governor

JOHN J. HOFFMAN
 Acting Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dreger Safety, Inc.

ANALYSIS DATE: 10/23/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 131122

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1212 to 0.1222 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 17, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alarvia, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 20th day of October, 2013.

Anda L. DePorti
 Notary



"An Internationally Accredited Agency"
 New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRANTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

KIM GUADAGNO
 Lt. Governor

JOHN J. HOFFMAN
 Acting Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1878 to 0.1994 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.


ANALYSIS DATE: 8/14/2014

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 14H131


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1925 to 0.1944 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 5, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


 Ali M. Klauie, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 14th day of August, 2014.


 Notary
 MARY ELIZABETH MCLAUGHLIN

ID # 2062190
 NOTARY PUBLIC
 STATE OF NEW JERSEY
 My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"

New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDAE-0022

Certification Date

8-28-14

Technician

BC

Re-Certification Due Date

8-28-15

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUNP2-231

Certification date:

9-19-14

Next Certification due:

9-19-15

Probe Value

107

Dräger Safety Diagnostics, Inc.
Technical Service Department

BC



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDMKS3-0003

Certification Date

9-19-14

Technician

BC

Re-Certification Due Date

9-19-15



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDCN-0052

Certification Date

9-22-14

Technician

BC

Re-Certification Due Date

9-22-15

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	Calib. No.:	00024
Calibration File No.:	01008	Cert. Date:	03/02/2015
Certification File No.:	01009	Lin. Date:	03/02/2015
Linearity File No.:	01010	Soln. Date:	03/02/2015
Solution File No.:	01011	File Date:	03/02/2015
Sequential File No.:	01011		

Calibrating Unit:	WET	Model No.:	CÜ-34	Serial No.:	DDUL S3-0167
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0894

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:18S		
Control 1 EC	0.100%	14:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:20S		
Control 2 EC	0.100%	14:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21S		
Control 3 EC	0.100%	14:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXAP2 - 011 MG

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: Sgt Michelle Goncalves #6040

Badge No.: 6040

Date: 03/02/2015



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST BRUNTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

KIM GUADAGNO
 Lt. Governor

JOHN J. HOFFMAN
 Acting Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 9/11/2014

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 14H132

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1211 to 0.1220 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 22, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 15th day of September, 2014.

Linda L. Desantis
 Notary

LINDA L DESANTIS
 ID # 22288889
 NOTARY PUBLIC
 STATE OF NEW JERSEY
 My Commission Expires August 17, 2019



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable

