

ALCOTEST CHECKLIST

Municipality: MERCHANTVILLE Alcotest Ser.#: ARUM-0073

County: CAMDEN Date of Calibration: 11-03-2010

- 1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
- 2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDWA 53-0011
- 3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDUJ P2-168
- 4. **Digital Temperature Measuring System Report of Calibration.**
Ser. #: 101733532
- 5.
 - A. Alcotest 7110 Calibration Record
 - B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
 - C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
 - D. **Alcotest Card** of operator/coordinator who performed tests.
 - E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDMB P1-0002
- 6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:

<input checked="" type="checkbox"/> A.	0.04% Solution.	<u>10</u>	<u>A</u>	<u>073</u>
<input checked="" type="checkbox"/> B.	0.08% Solution.	<u>10</u>	<u>A</u>	<u>074</u>
<input checked="" type="checkbox"/> C.	0.10% Solution.	<u>10</u>	<u>F</u>	<u>080</u>
<input checked="" type="checkbox"/> D.	0.16% Solution.	<u>10</u>	<u>A</u>	<u>075</u>
- 7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
 - A. 0.04% used in Calibration/Linearity Testing. DDWF 53-0215
 - B. 0.08% used in Calibration/Linearity Testing. DDXD 53-0193
 - C. 0.10% used in Calibration/Linearity Testing DDWA 53-0011
[Same as instrument's CU34 unit].
 - D. 0.16% used in Calibration/Linearity Testing. DDWF 53-0241
- 8.
 - A. New Standard Solution Report following Calibration.
 - B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
 - C. Certificate of Analysis 0.10% solution for same.
Lot #: 10 C 077
 - D. Alcotest card of operator/coordinator who completed change.

Dräger safety

Alcotest[®] 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standards for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48254, 49 FR 48984 and 50 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:


03/29/2006

SERIAL NUMBER:

ARUM-0073

Dräger Safety Diagnostics, Inc.
Durango, CO

Ante Dely



Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDWAS3-0011

Certification Date

5-7-10

Technician

mm

Re-Certification Due Date

5-7-11

Dräger safety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUJ P2-168

Certification date:

5/7/10

Next Certification due:

5/7/11

Probe Value

1.02

Dräger Safety Diagnostics, Inc.
Technical Service Department

mm



Calibration
Certificate No. 1750.01

Calibration complies with ISO 9001
ISO/IEC 17025 AND ANSI/NCSL Z540-1



Cert. No.: 4000-2966287

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, P.O. Box 2158, Secaucus, NJ 07094 U.S.A.

Instrument Identification:

Model: 61220-601

S/N: 101733532

Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-179	A45240		
Thermistor Module	A17118	11/19/10	A9B21010
Temperature Probe	128	12/10/10	A9B23079
Temperature Calibration Bath TC-231	A79341		
Temperature Probe	3039	12/10/10	A9B23080-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Module	A27129	7/09/10	1000264338
Temperature Probe	5202	3/11/11	90310050
Temperature Calibration Bath TC-256	B01375		
Temperature Probe	157	7/27/10	A9708011-4

Certificate Information:

Technician: 104

Procedure: CAL-06

Cal Date: 6/08/10

Cal Due: 6/08/12

Test Conditions: 24.0°C 42.0%RH 1015 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C		N.A.		0.002	-0.002	Y	-0.048	0.052	0.013	3.8:1
°C		N.A.		24.999	25.000	Y	24.949	25.049	0.013	3.8:1
°C		N.A.		60.001	60.002	Y	59.951	60.051	0.018	2.8:1
°C		N.A.		100.001	100.001	Y	99.951	100.051	0.013	3.8:1

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test, and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2, to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=(Max-Min)/2; Min = Nominal(Rounded) - Tolerance; Max = Nominal(Rounded) + Tolerance; Date=MM/DD/YY

Nicol Rodriguez
Nicol Rodriguez, Quality Manager

Wallace Berry
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock and contamination.

Recalibration:

For factory calibration and recertification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MORGANTHAU POLICE
Calibration File No.: 00527 **Calib. Date:** 11/03/2010 **Calib. No.:** 00013
Certification File No.: 00528 **Cert. Date:** 11/03/2010 **Cert. No.:** 00011
Linearity File No.: 00500 **Lin. Date:** 05/05/2010 **Lin. No.:** 00010
Solution File No.: 00526 **Soln. Date:** 11/01/2010 **Soln. No.:** 00097
Sequential File No.: 00528 **File Date:** 11/03/2010

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWA S3-0011
Control Solution %: 0.100% **Expires:** 06/14/2012
Solution Control Lot: 10F080 **Bottle No.:** 0558

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:23S	34.0°C	
Control 1 EC	0.099%	09:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:24S		
Control 2 EC	0.097%	09:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26S		
Control 3 EC	0.098%	09:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27S		

All tests within acceptable tolerance.

Coordinator

Last Name: SULLIVAN First Name: FRANCIS MI: MI
Signature: *[Handwritten Signature]* #5103 Badge No.: 5103
Date: 11/03/2010

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00527	Calib. Date: 11/03/2010
Certification File No.:	00528	Cert. No.: 00013
Linearity File No.:	00529	Cert. Date: 11/03/2010
Solution File No.:	00526	Lin. No.: 00011
Sequential File No.:	00529	Lin. Date: 11/03/2010
		Soln. No.: 00097
		File Date: 11/03/2010
Calibrating Unit:	WET	Model No.: CU 34
Control Solution %:	0.040%	Serial No.: DDWF S3-0215
Solution Control Lot:	10A073	Expires: 01/12/2012
		Bottle No.: 1090
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.080%	Serial No.: DDXD S3-0193
Solution Control Lot:	10A074	Expires: 01/15/2012
		Bottle No.: 1232
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.160%	Serial No.: DDWF S3-0241
Solution Control Lot:	10A075	Expires: 01/21/2012
		Bottle No.: 0013

Function	Result	Time	Temperature	Comment(s)
	%BAC	MM:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:50S	10.500 ₃₃	
Control 1 EC	0.040%	09:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	09:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52S		
Control 2 EC	0.040%	09:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	09:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54S		
Control 3 EC	0.081%	09:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.081%	09:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		
Control 4 EC	0.080%	09:57S	34.0°C	*** TEST PASSED ***
Control 4 IR	0.078%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		
Control 5 EC	0.161%	09:59S	33.9°C	*** TEST PASSED ***
Control 5 IR	0.160%	09:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01S		
Control 6 EC	0.160%	10:02S	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	10:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03S	11.03 ₀₃	

All tests within acceptable tolerance.

Coordinator

Last Name: SULLIVAN

First Name: FRANCIS

MI: -

Signature: *Francis J. Sullivan #5103*

Badge No.: 5103

Date: 11/03/2010

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Francis Sullivan
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CERTIFICATE COURSE IN ACCORDANCE WITH CHAPTER 140 OF
 THE LAWS OF NEW JERSEY IN THE OPERATION OF THE **ALSTON 7110 MKIII-C**
 A MOTOR TO OPERATE MOTORCYCLE
 LICENSES BY NAME AT LOCATION, NEW JERSEY

DATE March

TWO THOUSAND AND **SIX**

[Signature]
 CLERK
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. <i>05/16/08</i>	<i>Seyreville PO</i>	<i>[Signature]</i>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____



JON S. CORZINE
Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
PO BOX 080
TRENTON, NJ 08625-0080

ANNE MILGRAM
Attorney General

November 5, 2008

Col. Joseph R. Fuentes, Superintendent
Division of State Police
Division Headquarters
P.O. Box 7068
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification - Trooper I Francis Sullivan # 5103

Dear Col. Fuentes:

Pursuant to the provisions of *N.J.A.C.* 13:51-2.1 (b) and (c), as adopted and promulgated under the provisions of *N.J.S.A.* 39:4-50.3, 39:3-10.25 and 12:7-56, I hereby approve Trooper I Francis Sullivan # 5103 as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,

Anne Milgram
Attorney General

AM:HA:tlh

c: Trooper I Francis Sullivan # 5103, Alcohol/Drug Test Unit, Division of State Police
Lt. Paul Spirit, Unit Head, Alcohol/Drug Test Unit, Division of State Police





CERTIFICATE

This is to certify that


Francis Sullivan
Trooper I, #5103

has successfully completed the two day Draeger Safety Diagnostics, Inc. Alcohol Coordinator Training Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: September 18, 2008


Instructor: Hansueli Fyser

Dräger



Francis Sullivan #5103
Hamilton Technology Complex
Suite 400, 1200 Negrón Drive
Hamilton, NJ 08691
lpp5103@gw.njsp.org

Current Rank & Date of Promotion & Time of Service:

Trooper I 7/16/2002 16 years of Service

Career Experience/Assignment:

April 26, 2008 to Present:

Special Investigation Section
Alcohol/Drug Testing Unit

May 29 2004 to April 25 2008:

Special Operations Section
Motor Coach/ Compliance Review Unit

December 29 2001 to May 28 2004:

Special Operations Section
Diesel Emissions

November 3 2001 to December 28 2001:

Field Operations Section
Tuckerton Station
Troop A

April 1 1995 to November 2 2001:

Field Operations Section
Bass River Station
Troop E

October 15 1994 to March 31 1996:

Field Operations Section
Port Norris Station
Troop A

October 16 1993 to October 14 1994:

Field Operations Section
Tuckerton Station
Troop A

January 16, 1993 to October 15 1993:

Field Operations Section
Woodbine Station
Troop A

Specialized Training:

Breath Test Coordinator/Instructor, Certification

November 2008

Draeger Safety Diagnostics, Inc	May 2009
Alcotest 7110-MKIII-C Memory Data Download Training	
Instructor Training in DWI Detection and Standardized	October 2008
Field Sobriety Testing	
Draeger Safety Diagnostics, Inc., Alcohol Coordinator	September 2008
Training Course on the New Jersey specific Alcotest	
7110 MKIII-C	
The Robert F. Borkenstein Course on Alcohol and	May 2008
Highway Safety	
Inspection and Investigation of Commercial Vehicle	September 2007
Crashes	
Accident Investigation 1	January 2007
Accident Investigation 2	March 2007
Instructor Training Course	May 2006
Laser Operator Training Course	March 2006
Alcotest Certification	March 2006
Post Crash Vehicle Equipment Inspection	November 2004
Compliance Review Training	September 2003
Suspension and Steering	August 2003
Commercial Motor Vehicle Criminal Interdiction	April 2003
General Hazardous Material Inspection	December 2002
Domestic Cannabis Eradication & Suppression Seminar	July 2002
Trucks and Terrorism	May 2002
Motorcoach	May 2002
NASI-North American Standard Inspection Part B	February 2002
NASI-Weights & Dimensions	February 2002
NASI-North American Standard Inspection Part A	January 2002
PR-24 Training Course	July 1999
Narcotics Trafficking Trends Seminar	June 1998
Firearms Instructor Course	May 1997
Driving While Intoxicated/Standardized Field Sobriety	July 1996

Educational/Certification Summary:

Associates of Science	
Comm. College of the Air Force	August 1987

Commendation & Awards:

Certificate of Commendation	June 2002
-----------------------------	-----------

DEPARTMENT OF
Traffic and Public Safety

This is to certify that

Frank J. Sullivan

New Jersey State Police

IS QUALIFIED AND COMPETENT TO OPERATE AN ANALYZER PURSUANT TO CHAPTER 162 OF THE LAWS OF 1964 IN THE PERFORMANCE OF HIS DUTY AS A MEMBER OF THE NEW JERSEY STATE POLICE



29th DAY OF Apr.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS

ONE THOUSAND NINE HUNDRED AND 94

Carl Williams
 SUPERVISOR
 NEW JERSEY STATE POLICE

John F. Pate
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1	9-6-96	ACTC	J. P. [unclear]
2	12-15-97	ACTC	[unclear]
3	12-8-99	ACTC	[unclear]
4	10-9-01	OCFA	[unclear]
5	7-17-03	ACTC	[unclear]
6	6-10-04	ACTC	[unclear]
7	11-16-07	OCFA	[unclear]
8.			
9.			

Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe:

DDMBP7-0000

Certification date:

2-3-10

Next Certification due:

2-3-11

Probe Value

101

Draeger Safety Diagnostics, Inc.
Technical Service Department

mm



CHRIS CHRISTIE
Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
Post Office Box 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

PAULA T. DOW
Acting Attorney General
COLONEL JOSEPH R. FUE
Superintendent

CERTIFICATION OF ANALYSIS
0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution:

MANUFACTURER: Drager Safety, Inc.


ANALYSIS DATE: 2/2/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A073


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0479 to 0.0481 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 12, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Kenneth W. Klawalek, M.S.
Assistant Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 1st day of February, 2010.


Notary

Linda L. Decaris
Notary Public, New Jersey
Commission Expires 8-17-14





CHRIS CHRISTIE
Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08626-0068
(609) 882-2000

PAULA T. DOW
Acting Attorney General
COLONEL JOSEPH R. FUEN
Superintendent

**CERTIFICATION OF ANALYSIS
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.


MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 2/3/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A074


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0954 to 0.0958 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Kenneth W. Kawalek, M.S.
Assistant Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 19th day of February, 2010.


Notary

Linda L. Deonno
Notary Public, New Jersey
My Commission Expires 8-17-14





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

KIM GUADAGNOLI
 Lt. Governor

PAULA T. DOW
 Attorney General

COLUMBIA JOSEPH R. FUENTES
 Superintendent

**CERTIFICATION OF ANALYSIS
 0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.

ANALYSIS DATE: 7/8/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10F080

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1197 to 0.1207 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 14, 2012.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Ajit R. Tungarc

 Ajit R. Tungarc
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 8th day of July, 2010.

Dorinda K. DeFantis

 Notary



New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





CHRIS CHRISTIE
Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

PAULA T. DOW
Acting Attorney General

COLONEL JOSEPH R. FUE
Superintendent

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.


MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 2/4/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A075


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1913 to 0.1919 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 21, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Kenneth W. Kawalek, M.S.
Assistant Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 19th day of February, 2010.


Notary

Linda L. Desantis
Notary Public, New Jersey
My Commission Expires 8-17-14





Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

Model: ALCOTEST[®] CU34

Model: MARK IIA

Other: _____

Serial Number:

DDWFS3-0215

Certification Date

Technician

Re-Certification Due Date

2/3/10

JM

2/3/11



Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(P.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Serial Number:

Model: MARK IIA

Other: _____

DDXDS 3-0193

Certification Date

Technician

Re-Certification Due Date

2/3/10

AM

2/3/11



- Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: _____

Serial Number:

DDWF 53-0241

Certification Date

Technician

Re-Certification Due Date

2/3/10

DM

2/3/11

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCRIANTVILLE POLICE	Calib. No.:	00013
Calibration File No.:	00527	Cert. Date:	11/03/2010
Certification File No.:	00528	Lin. Date:	11/03/2010
Linearity File No.:	00529	Soln. Date:	11/03/2010
Solution File No.:	00530	File Date:	11/03/2010
Sequential File No.:	00530		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	10C077	Expires:	03/10/2012
		Bottle No.:	0574

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:11S	34.0°C	
Control 1 EC	0.101%	11:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13S		
Control 2 EC	0.099%	11:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		
Control 3 EC	0.099%	11:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: 00UJPP-16803

Changed By:

Last Name: SULLIVAN	First Name: FRANCIS	MI: -
Signature: <u><i>Tr. I J Sullivan #5103</i></u>	Badge No.: 5103	Date: 11/03/2010



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE Box 7068
 WEST TRENTON NJ 08628 0068
 (609) 822-2000

CHRIS CHRISTIE
 Governor

PAULA T. DOW
 Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENZALDA
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.


MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 3/24/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10C877


Representative samples of the above referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1198 to 0.1202 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 10, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


 Kenneth W. Kawalek, M.S.
 Assistant Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 24th day of April, 2010.


 Notary



New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable

