



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

JOHN S. COZZINE
Governor

ANNE MILGRAM
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc. **ANALYSIS DATE:** 06/28/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07B046

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1221 to 0.1228 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 17:27-3.4, of approved breath test instruments (N.J.A.C. 17:27-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 24, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tongare
Ajit R. Tongare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 13th day of July, 2007.

Linda L. DeSantis
Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



New Jersey is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

JON S. CORZINE
Governor

ZULMA V. FARBER
Attorney General

COLONEL JOSEPH R. PUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. **ANALYSIS DATE:** 4/19/06

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 06C030

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1196 to 0.1201 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 16, 2008.

As Forensic Laboratory Director of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Thomas A. Brettell
Thomas A. Brettell, Ph.D.
Forensic Laboratory Director
Division of State Police

Sworn to and subscribed before me this 23rd day of May, 2006.

Linda L. DeSantia
Notary

Linda L. DeSantia
My Commission
Expires Aug. 17, 2008





Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDULS3-0163

Certification Date

Technician

Re-Certification Due Date

02/14/2007

CRD

02/14/2008

Dräger safety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDUTP2-030

02/14/2007

02/14/2008

Probe Value

98

Dräger Safety Diagnostics, Inc.
Technical Service Department

CRD

Draeger safety

Alcotest[®] 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48654, 49 FR 48664 and 58 FR 46705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

SERIAL NUMBER:

11 APR 05

AE0K50039

Draeger Safety Diagnostics Inc.
Durango, CO

[Signature]

Alcotest 7110 Calibration Record

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00308 Calib. Date: 04/25/2007 Calib. No.: 00006
Certification File No.: 00146 Cert. Date: 04/25/2006 Cert. No.: 00003
Linearity File No.: 00147 Lin. Date: 04/25/2006 Lin. No.: 00002
Solution File No.: 00297 Soln. Date: 04/02/2007 Soln. No.: 00033
Sequential File No.: 00308 File Date: 04/25/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 11/29/2007
Solution Control Lot: 05K021 Bottle No.: 0406

Coordinator

Last Name: CROSS First Name: WILLIAM MI: M
Signature: TPR. I WILLIAM M. CROSS #5342 Badge No.: 5342
Date: 04/25/2007

*Black Key Temperature Probe Serial# DDWAPJ-001 WCR

*Ertco-Hart Digital Temperature Measuring System Serial# A23725 WCR

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00308	Calib. Date:	04/25/2007
Certification File No.:	09309	Cert. Date:	04/25/2007
Linearity File No.:	00147	Lin. Date:	04/25/2006
Solution File No.:	00297	Soln. Date:	04/02/2007
Sequential File No.:	00309	File Date:	04/25/2007
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	05K021	Expires:	11/29/2007
		Bottle No.:	0406

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:46D		
Control 1 EC	0.100%	12:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47D		
Control 2 EC	0.099%	12:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48D		
Control 3 EC	0.099%	12:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50D		

All tests within acceptable tolerance.

Coordinator

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: _____

TPR. W. M. Cross

Badge No.: 5343

Date: 04/25/2007

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00308	Calib. Date:	04/25/2007
Certification File No.:	00309	Cert. No.:	00004
Linearity File No.:	00310	Lin. Date:	04/25/2007
Solution File No.:	00297	Soln. No.:	00033
Sequential File No.:	00310	File Date:	04/25/2007
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.040%	Serial No.:	DDMK S3-0003
Solution Control Lot:	06B026	Expires:	02/25/2008
		Bottle No.:	0247
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.080%	Serial No.:	DDRK S3-0005
Solution Control Lot:	06B024	Expires:	02/24/2008
		Bottle No.:	0267
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.160%	Serial No.:	DDRF S3-0009
Solution Control Lot:	06B025	Expires:	02/25/2008
		Bottle No.:	0324

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:08D		
Control 1 EC	0.040%	13:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.039%	13:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10D		
Control 2 EC	0.040%	13:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	13:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12D		
Control 3 EC	0.080%	13:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.078%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		
Control 4 EC	0.079%	13:15D	34.0°C	*** TEST PASSED ***
Control 4 IR	0.079%	13:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:16D		
Control 5 EC	0.158%	13:17D	34.0°C	*** TEST PASSED ***
Control 5 IR	0.157%	13:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18D		
Control 6 EC	0.157%	13:19D	34.0°C	*** TEST PASSED ***
Control 6 IR	0.155%	13:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20D		

All tests within acceptable tolerance.

Coordinator

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: TPR I WILL M. Cross #5340

Badge No.: 5342

Date: 04/25/2007



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

RICHARD J. CODEY
Acting Governor

PETER C. HARVEY
Attorney General

COLONEL JOSEPH R. PUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 12/12/05

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 05K021

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1208 to 0.1210 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is November 29, 2007.

As Forensic Laboratory Director of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Thomas A. Brettell
Thomas A. Brettell, Ph.D.
Forensic Laboratory Director
Division of State Police

Sworn to and subscribed before me this 20th day of December, 2005.

Linda L. DeSantis
Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



New Jersey is An Equal Opportunity Employer • Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
Post Office Box 7068
West Trenton NJ 08628-0068
(609) 882-2000

ION S. CORLEONE
Governor

ZILMA V. FAEBER
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

**CERTIFICATION OF ANALYSIS
0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. **ANALYSIS DATE:** 3/17/06

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 06B626

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0477 to 0.0479 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 23, 2008.

As Forensic Laboratory Director of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Thomas A. Bretell
Thomas A. Bretell, Ph.D.
Forensic Laboratory Director
Division of State Police

Sworn to and subscribed before me this 10th day of April, 2006.

Linda L. DeSantis
Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
Post Office Box 7068
West Trenton NJ 08628-0068
(609) 882-2000

JON S. CORZINE
Governor

ZULIMA V. FARBER
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 3/17/06

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 06B024

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0956 to 0.0959 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 24, 2008.

As Forensic Laboratory Director of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


Handwritten signature of Thomas A. Bretteil

Thomas A. Bretteil, Ph.D.
Forensic Laboratory Director
Division of State Police

Sworn to and subscribed before me this 10th day of April, 2006.

Handwritten signature of Notary and the word Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
Post Office Box 7068
West Trenton NJ 08628-0068
(609) 882-2000

ZULMA V. FARBER
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

ION S. CORZINE
Governor

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 3/17/06

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 06B025

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1916 to 0.1929 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 25, 2008.

As Forensic Laboratory Director of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Thomas A. Brenell, Ph.D.
Forensic Laboratory Director
Division of State Police

Sworn to and subscribed before me this 10 day of April, 2006.

Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00308 Calib. Date: 04/25/2007 Calib. No.: 00006
Certification File No.: 00309 Cert. Date: 04/25/2007 Cert. No.: 00004
Linearity File No.: 00310 Lin. Date: 04/25/2007 Lin. No.: 00003
Solution File No.: 00311 Soln. Date: 04/25/2007 Soln. No.: 00034
Sequential File No.: 00311 File Date: 04/25/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 03/16/2008
Solution Control Lot: 06C030 Bottle No.: 0282

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:27D		
Control 1 EC	0.099%	14:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	14:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28D		
Control 2 EC	0.099%	14:29D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:29D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30D		
Control 3 EC	0.098%	14:30D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:30D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUMP2 - 316 wne

Changed By:

Last Name: CROSS


First Name: WILLIAM

MI: M

Signature: TPR. I. WILLIAM CROSS #5342

Badge No.: 5342

Date: 04/25/2007



Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST[®] CU34
 Model: MARK IIA
 Other: _____

Serial Number:

DDMK53-0003

Certification Date


Technician

Re-Certification Due Date

FEB 19 2007



FEB 19 2008



Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST[®] CU34
 Model: MARK IIA
 Other: _____

Serial Number:

DDPK 53-0005

Certification Date

Technician

Re-Certification Due Date

FEB 19 2007

FEB 19 2008



Draeger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
 Model: MARK IIA
 Other: _____

Serial Number:

DDRF 33-8009

Certification Date

FEB 19 2007

Technician

Re-Certification Due Date

FEB 19 2008



Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWAP2 - 001

Certification date:

02/19/2007

Next Certification due:

02/19/2008

Probe Value

101

Draeger Safety Diagnostics, Inc.
Technical Service Department

CRD

Dräger safety

Ertco-Hart Digital Temperature Measuring System

REPORT OF CALIBRATION

This is to certify that the Ertco-Hart Digital Temperature Measuring System has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). Draeger Safety Diagnostics, Inc. (DSDI) recommends accuracy verification of the Ertco-Hart Digital Temperature Measuring System within 12 months of the certification date below, or sooner, according to your state specification.

DSDI equipment used for temperature verification Serial Number: AH41 303176

Digital Units Serial Number: A23725

Probe Serial Number: 583984

Certification Date: 02/19/2007

Next Certification Due: 02/19/2008

At 34.00 °C digital unit displays 34.00 °C

Draeger Safety Diagnostics, Inc. Technician: CRD

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

William M. Cross
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL ANALYSIS PURSUANT TO CHAPTER 147 OF THE LAWS OF 1986 IN THE OPERATING OF THE **Alcotest 7110 MKIII-C** BREATHALYZER
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS **1st** DAY OF **February**
 ONE THOUSAND **05**

[Signature]
 SUPERVISOR
 NEW JERSEY STATE POLICE



ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1		
2		
3		
4		
5		
6		
7		
8		
9		

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

William M. Cross

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL ANALYSIS PURSUANT TO CHAPTER 147 OF THE LAWS OF 1986 IN THE OPERATING OF THE **Breathalyzer**
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS **31st** DAY OF **Jan.**
 ONE THOUSAND **NINE HUNDRED AND** **97**

[Signature]
 SUPERVISOR
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 2-11-99	GCDA	<i>[Signature]</i>
2. 10-10-01	BCPA	<i>[Signature]</i>
3. 4-17-03	BCDA	<i>[Signature]</i>
4. 6-22-05	ACTC	<i>[Signature]</i>
5		
6		
7		
8		
9		



CERTIFICATE

This is to certify that


William M. Cross
Trooper II #5342

has successfully completed the two day Draeger Safety Diagnostics, Inc. Operator Training and Preventive Maintenance Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: October 28, 2004



Instructor: Hansueli Ryser

Draegersafety

William M. Cross #5342
Department of Law and Public Safety
New Jersey State Police
P.O. Box 7068-River Road
West Trenton, NJ 08628
609-882-2000
lpp5342@gw.njsp.org

Enlistment
Date

December 23, 1995

Career
Experience

10 years, 9 months

May 2003 to present

Field Operations Section
Alcohol/Drug Test Unit

December 2001 to May 2003

Field Operations Section
Tactical Patrol Unit #1
Troop D

March 1999 to December 2001

Field Operations Section
Moorestown Barracks
Troop D

November 1997 to March 1999

Field Operations Section
Bellmawr Barracks
Troop A

October 1997 to November 1997

Field Operations Section
Red Lion Barracks
Troop A

January 1997 to October 1997

Field Operations Section
Tuckerton Barracks
Troop A

June 1996 to January 1997

Field Operations Section
Bridgeton Barracks
Troop A

December 1995 to June 1996

Field operations Section
Red Lion Barracks
Troop A

July 1995 to December 1995


New Jersey State Police Academy

Education

- May 1992 Bachelor of Arts degree, Political Science,
Kutztown University, Kutztown, PA
- June 1988 Graduated Haddon Township High School,
Haddon Township, NJ

Technical/ Professional Training

- February 2005 Alcotest® 7110 MKIII-C Certified
- October 2004 Draeger Safety Diagnostics, Inc. Operator Training
and Preventative Maintenance Course on the
New Jersey specific Alcotest® 7110 MKIII-C
- January 2004 Instructor Training in DWI Detection and
Standardized Field Sobriety Testing
- January 2004 Certified in Standardized Field Sobriety
Tests including Horizontal Gaze Nystagmus
- September 2003 Certified Breath Test Coordinator/Instructor
- August 2003 National Draeger, Inc. Breathalyzer® 900/900A
Maintenance Course
- July 2003 Methods of Instruction Course
- June 2003 Driving While Intoxicated/Standardized Field
Sobriety Testing Course
- May 2003 The Robert F. Borkestein Course on Alcohol,
Drugs, and Highway Safety
- May 2003 Gang Awareness Course
- March 2000 Spanish for Law Enforcement Course
- March 1999 Traffic Law Enforcement Course
- June 1998 Accident Investigation I Course
- February 1997 Tracktest Fifth Wheel Certified
- January 1997 5-Day Breathalyzer® 900/900A Operator Certified
- February 1996 K-55 Radar Operator Certified



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: _____

Serial Number:

DDULSS-0163

Certification Date

02/14/2008

Technician

CRD

Re-Certification Due Date

02/14/2008

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Ronald M. Morello
 Stratford Boro

IS EXAMINED AND COMPETENT TO CONDUCT CHEMICAL (DRUG) ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1964 IN THE OPERATION OF THE **ALCOTEST 7110 MK III C**
 A METHOD TO DETERMINE INTOXICATION.
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 1st DAY OF February

[Signature]
 SUPERVISOR
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1-1-607	GCRA	C. P. Miller
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. FORM (Rev. 12-64)

DEPARTMENT OF
Fire and Public Safety

This is to certify that
Stephen J. McBride
Stratford Boro.

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL URINALYSIS PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**
 A METHOD TO OBTAIN DRUG INTOXICATION
 GIVEN UNDER MY HANDS AT TRENTON, NEW JERSEY THIS 14th DAY OF February

TWO THOUSAND AND Five

[Signature]
 SUPERVISOR
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course	
	PLACE	INSTRUCTOR
1. 2-2-07	GCPA	WM Ross
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

*Advise subject:

"I want you to take a deep breath and blow into the mouthpiece with one long continuous breath."

Continue to blow until I tell you to stop.

Do you understand these instructions?

**NO
PORTABLE
RADIOS ON
DURING
BREATH
TESTS**

NO PORTABLE
RADIOS OR
CELLULAR
TELEPHONES IN
THIS AREA OF THE
PATROL ROOM
DURING BREATH
TESTING
KEEP PORTABLE
RADIOS OFF
DURING TESTING

**Advise subject:*

I want you to take a deep breath and blow into the mouthpiece with one long continuous breath.

Continue to blow until I tell you to stop.

Do you understand these instructions?