

# ALCOTEST CHECKLIST

Municipality: BUENA BORO Alcotest Ser.#: ARXB-0074  
County: ATLANTIC Date of Calibration: 08-15-2011

- 1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
- 2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.  
Ser. #: DDX J 53-0210
- 3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.  
Ser. #: DDWJ P2-204
- 4. **Digital Temperature Measuring System Report of Calibration.**  
Ser. #: 101 733534
- 5.
  - A. Alcotest 7110 Calibration Record
  - B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
  - C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
  - D. **Alcotest Card** of operator/coordinator who performed tests.
  - E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].  
Ser. #: DDWA P2-016
- 6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:
  - A. 0.04% Solution. 10 A 073
  - B. 0.08% Solution. 10 A 074
  - C. 0.10% Solution. 10 I 082
  - D. 0.16% Solution. 10 A 075
- 7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
  - A. 0.04% used in Calibration/Linearity Testing DDWE 53-0196
  - B. 0.08% used in Calibration/Linearity Testing DDWE 53-0203
  - C. 0.10% used in Calibration/Linearity Testing DDXJ 53-0210  
[Same as instrument's CU34 unit].
  - D. 0.16% used in Calibration/Linearity Testing DDWE 53-0205
- 8.
  - A. New Standard Solution Report following Calibration.
  - B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
  - C. Certificate of Analysis 0.10% solution for same.  
Lot #: 09 L 072
  - D. Alcotest card of operator/coordinator who completed change.

Dräger safety

Alcotest<sup>®</sup> 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 40705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:


SERIAL NUMBER:

8-19-2010

ARXB-0074

Dräger Safety Diagnostics, Inc.

\_\_\_\_\_



**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Draeger Safety Diagnostics, Inc.**

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDXJ53-0210

Certification Date

8-5-11

Technician

MM

Re-Certification Due Date

8-5-12

# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

### CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWJP2-204

Certification date:

8-5-11

Next Certification due:

8-5-12

Probe Value

104

**Draeger Safety Diagnostics, Inc.**  
Technical Service Department

MM



Calibration  
Certificate No. 1750.01

Calibration complies with ISO 9001  
ISO/IEC 17025 AND ANSI/NCCL Z540-1



Cert. No.: 4000-2966289

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, P.O. Box 2158, Secaucus, NJ 07094 U.S.A.  
Instrument Identification:

Model: 61220-601      S/N: 101733534      Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-179	A45240		
Thermistor Module	A17118	11/19/10	A9821010
Temperature Probe	128	12/10/10	A9823079
Temperature Calibration Bath TC-231	A79341		
Temperature Probe	3039	12/10/10	A9823080-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Module	A27129	7/09/10	1000264338
Temperature Probe	5202	3/11/11	80310050
Temperature Calibration Bath TC-256	B01375		
Temperature Probe	157	7/27/10	A9708011-4

Certificate Information:

Technician: 104      Procedure: CAL-06      Cal Date: 6/08/10      Cal Due: 6/08/12  
Test Conditions: 24.0°C      42.0 %RH      1015 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C		N.A.		0.002	-0.001	Y	-0.048	0.052	0.013	3.8:1
°C		N.A.		24.999	25.000	Y	24.949	25.049	0.013	3.8:1
°C		N.A.		60.001	59.998	Y	59.951	60.051	0.018	2.8:1
°C		N.A.		100.001	100.002	Y	99.951	100.051	0.013	3.8:1

This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy= $\pm(\text{Max}-\text{Min})/2$ ; Min = Nominal(Rounded) - Tolerance; Max = Nominal(Rounded) + Tolerance; Date=MM/DD/YY

*Nicol Rodriguez*  
Nicol Rodriguez, Quality Manager

*Wallace Berry*  
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 : USA  
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.  
Control Company is ISO 9001:2008 Quality Certified by (DNY) Det Norske Veritas, Certificate No. CERT-01805-2006-AD-HOU-ANAB.  
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

# Alcotest 7110 Calibration Record

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074		
Location:	BUENA BOROUGH POLICE				
Calibration File No.:	00140	Calib. Date:	08/15/2011	Calib. No.:	00012
Certification File No.:	00106	Cert. Date:	02/15/2011	Cert. No.:	00006
Linearity File No.:	00107	Lin. Date:	02/15/2011	Lin. No.:	00006
Solution File No.:	00134	Soln. Date:	08/02/2011	Soln. No.:	00043
Sequential File No.:	00140	File Date:	08/15/2011		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXJ S3-0210
Control Solution %:	0.100%			Expires:	09/09/2012
Solution Control Lot:	10I082			Bottle No.:	0325

## Coordinator

Last Name: DELLANOCE

First Name: JOSEPH

MI: S


Signature: Joseph Dellanoce # 6027

Badge No.: 6027

Date: 08/15/2011

\*Black Key Temperature Probe Serial.....# DDWAP2-016

\*Digital NIST Temperature Measuring System Serial.....# 101733534



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

# Alcotest 7110 Calibration Certificate

## Part I - Control Tests

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00140	Calib. Date: 08/15/2011
Certification File No.:	00141	Calib. No.: 00012
Linearity File No.:	00107	Cert. Date: 08/15/2011
Solution File No.:	00134	Cert. No.: 00007
Sequential File No.:	00141	Lin. Date: 02/15/2011
		Lin. No.: 00006
		Soln. Date: 08/02/2011
		Soln. No.: 00043
		File Date: 08/15/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXJ S3-0210
Solution Control Lot:	10I082	Expires: 09/09/2012
		Bottle No.: 0325

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:04D		
Control 1 EC	0.099%	15:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05D		
Control 2 EC	0.099%	15:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06D		
Control 3 EC	0.098%	15:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08D		

All tests within acceptable tolerance.

**Coordinator**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*Joseph Dellanoce*  
# 6027

Badge No: 6027

Date: 08/15/2011

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

# Alcotest 7110 Calibration Certificate

## Part II - Linearity Tests

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE		
Calibration File No.:	00140	Calib. Date:	08/15/2011
Certification File No.:	00141	Cert. Date:	08/15/2011
Linearity File No.:	00142	Lin. Date:	08/15/2011
Solution File No.:	00134	Soln. Date:	08/02/2011
Sequential File No.:	00142	File Date:	08/15/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.040%		Serial No.:
Solution Control Lot:	10A073		Expires:
			Bottle No.:
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.080%		Serial No.:
Solution Control Lot:	10A074		Expires:
			Bottle No.:
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.160%		Serial No.:
Solution Control Lot:	10A075		Expires:
			Bottle No.:

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:27D		
Control 1 EC	0.041%	15:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.039%	15:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29D		
Control 2 EC	0.040%	15:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.039%	15:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31D		
Control 3 EC	0.080%	15:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.079%	15:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33D		
Control 4 EC	0.080%	15:34D	34.0°C	*** TEST PASSED ***
Control 4 IR	0.079%	15:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35D		
Control 5 EC	0.161%	15:36D	34.0°C	*** TEST PASSED ***
Control 5 IR	0.159%	15:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:38D		
Control 6 EC	0.160%	15:38D	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	15:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:40D		

All tests within acceptable tolerance.

**Coordinator**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

Tp I [Signature] #6027

Badge No.: 6027

Date: 08/15/2011

DEPARTMENT OF  
**Motor and Public Safety**  
*This is to certify that*

Joseph S. Dellanoce  
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT EXAMINATIONS IN ACCORDANCE WITH CHAPTER 12 OF  
 THE LAWS OF THE STATE IN THE OPERATION OF THE **Alcohol 7110 MKII-C**  
 & LICENSE TO DETROCKING REGISTRATION  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29th DAY OF March

TWO THOUSAND AND Five

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	2/15/07	CCPA	C. Potter
2.	12/7/09	OLPA	Adam Stanke
3.	7/20/11	STOCKTON	Adam Stanke
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 12/04)



DEPARTMENT OF  
**Motor and Public Safety**  
This is to certify that

Joseph S. Dellanoce  
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL SCREENING ANALYSIS PURSUANT TO CHAPTER 10 OF  
 THE LAWS OF 1964 IN THE OPERATION OF THE Breath Test Coordinator/Instructor  
 A METHOD TO DETERMINE INTOXICATION  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 14th DAY OF January  
 TWO THOUSAND AND EIGHT

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	12/7/09	OCPS	Adam Stankovic
2.	7/20/11	STOCKTON	Adam Stankovic
3.			
4.			
5.			
6.			
7.			
8.			
9.			

DEPARTMENT OF  
*Peace and Public Safety*  
 This is to certify that  
Joseph B. Bellanocce  
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT MEDICAL BREATH ANALYSIS (PURSUANT TO CHAPTER 114 OF THE LAWS OF 1946 IN THE OPERATION OF THE BREATHALYZER  
 A METHOD TO DETERMINE INTOXICATION)

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 13th DAY OF Dec.  
 TWO THOUSAND AND 01

*Car. Filby*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*John P. ...*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES: 8/30 - 9/3/99

Refresher Course

	DATE	PLACE	INSTRUCTOR
1.	<u>12/12/01</u>	<u>A.G.P.C.</u>	<u>Potter</u>
2.	<u>7-30-03</u>	<u>CCPA</u>	<u>C. Potter</u>
3.	<u>10/3-05</u>	<u>ACTC</u>	<u>A. ...</u>
4.			
5.			
6.			
7.			
8.			
9.			

SP-283B (Rev. 4/00)



JON S. CORZINE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
PO Box 080  
TRENTON, NJ 08625-0080

ANNE MILGRAM  
Attorney General

January 14, 2008

Col. Joseph R. Fuentes, Superintendent  
Division of State Police  
Division Headquarters  
P.O. Box 7068  
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification -  
Trooper Joseph S. Dellanoce #6027

Dear Colonel Fuentes:

Pursuant to the provisions of *N.J.A.C. 13:51-2.1 (b) and (c)*, as adopted and promulgated under the provisions of *N.J.S.A. 39:4-50.3, 39:3-10.25 and 12:7-56*, I hereby approve Trooper Joseph S. Dellanoce #6027, as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,


  
Anne Milgram  
Attorney General

jk

- c. Trooper Joseph S. Dellanoce #6027, Alcohol/Drug Test Unit, Division of State Police  
Lt. Mark Kolodzieski, Unit Head, Alcohol/Drug Test Unit, Division of State Police



829474



# CERTIFICATE

This is to certify that


**Joseph S. Dellanoce #6027**

has successfully completed the two day Draeger Safety  
Diagnostics, Inc. Operator Training and Preventive  
Maintenance Course on the New Jersey specific  
Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and  
Maintenance Technician**

Completion of this course qualifies this individual to train  
and certify Operators in the proper use and operation as  
well as perform Preventive Maintenance on the New Jersey  
specific Alcotest® 7110 MKIII-C.

Date: 4/12/2007

  
\_\_\_\_\_  
Instructor: Hansueli Ryser

**Draeger**



ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWAP2-016

Certification date:

February 7, 2011

Next Certification due:

February 7, 2012

Probe Value

107

Draeger Safety Diagnostics, Inc.  
Technical Service Department



CHRIS CHRISTIE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

PAULA I. DOW  
Acting Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution.


MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/2/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A073


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0479 to 0.0481 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 12, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
Kenneth W. Kawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

  
Notary

Linda L. Dwan  
Notary Public, New Jersey  
Commission Expires 8-17-14



New Jersey Is An Equal Opportunity Employer





CHRIS CHRISTIE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

PAULA T. DOW  
Acting Attorney General  
COLONEL JOSEPH R. FUENTES  
Superintendent

CERTIFICATION OF ANALYSIS  
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/3/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A074

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0954 to 0.0958 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Kenneth W. Kawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

Notary

Linda L. DiStasio  
Notary Public, New Jersey  
My Commission Expires 8-17-14





State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

PAULA T. DOW  
 Attorney General

KIM GUADAGNO  
 Lt. Governor

COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.


ANALYSIS DATE: 9/23/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 101082


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1190 to 0.1194 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 9, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
 Kenneth W. Kawalek, M.S.  
 Assistant Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 28<sup>th</sup> day of September, 2010.

  
 Notary

Linda L. Deantis  
 Notary Public, New Jersey  
 My Commission Expires 8-17-14



New Jersey Is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable







CHRIS CHRISTIE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

PAULA T. DOW  
Acting Attorney General  
COLONEL JOSEPH R. FUENTES  
Superintendent

CERTIFICATION OF ANALYSIS  
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.


MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 2/4/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A075


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1913 to 0.1919 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 21, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
Kenneth W. Kawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

  
Notary

Linda L. Desantis  
Notary Public, New Jersey  
My Commission Expires 8-17-14





Dräger

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

DDWE 53-0196

Certification Date

Technician

Re-Certification Due Date

2/28/11

JM

2/28/12



Dräger

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34  
 Model: MARK IIA  
 Other: \_\_\_\_\_

Serial Number:

DDW53-0203

Certification Date

Technician

Re-Certification Due Date

February 7, 2011

DM

February 7, 2012



Dräger

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDWES3-0205

Certification Date

Technician

Re-Certification Due Date

February 7, 2011

DM

February 7, 2012

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE	<b>Calib. No.:</b>	00012
<b>Calibration File No.:</b>	00140	<b>Calib. Date:</b>	08/15/2011
<b>Certification File No.:</b>	00141	<b>Cert. No.:</b>	00007
<b>Linearity File No.:</b>	00142	<b>Lin. Date:</b>	08/15/2011
<b>Solution File No.:</b>	00143	<b>Soln. No.:</b>	00044
<b>Sequential File No.:</b>	00143	<b>File Date:</b>	08/15/2011
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDXJ S3-0210
<b>Solution Control Lot:</b>	09L072	<b>Expires:</b>	12/31/2011
		<b>Bottle No.:</b>	0931

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:53D		
Control 1 EC	0.101%	16:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55D		
Control 2 EC	0.100%	16:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:56D		
Control 3 EC	0.100%	16:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWJP2-204<sub>20</sub>

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*J. Dellanoce* # 6027

Badge No.: 6027

Date: 08/15/2011



State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

PAULA T. DOW  
 Acting Attorney General  
 COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 1/15/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 09L072

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1203 to 0.1210 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is December 31, 2011.

As the Director of Office of Forensic Sciences, Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Howard J. Baum, Ph.D.  
 Director, Office of Forensic Sciences  
 Division of State Police

Sworn to and subscribed before me this 21 day of January, 2010.

Notary

JOHN R. LEAVER  
 NOTARY PUBLIC STATE OF NEW JERSEY  
 MY COMMISSION EXPIRES 12/14/2012



New Jersey Is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable

