

ALCOTEST CHECKLIST

Municipality: BUENA BORO Alcotest Ser. #: ARXB-0074

County: ATLANTIC Date of Calibration: 04-24-2009

1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R..
2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDXJ S3-0210
3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDWJ P2-204
4. **Digital Temperature Measuring System Report of Calibration.**
5.
 - A. Alcotest 7110 Calibration Record
 - B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
 - C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
 - D. **Alcotest Card** of operator/coordinator who performed tests.
 - E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDMB P1-0002
6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:
 - A. 0.04% Solution. 08 F 054
 - B. 0.08% Solution. 08 F 055
 - C. 0.10% Solution. 08 G 056
 - D. 0.16% Solution. 08 G 057
7. **Certificate of Accuracy Alcotest CU34 Simulators** from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
 - A. 0.04% used in Calibration/Linearity Testing. DDWF-S3-0215
 - B. 0.08% used in Calibration/Linearity Testing. DDXD-S3-0193
 - C. 0.10% used in Calibration/Linearity Testing DDXJ-S3-0210
[Same as instrument's CU34 unit].
 - D. 0.16% used in Calibration/Linearity Testing. DDWF-S3-0241
8.
 - A. New Standard Solution Report - following Calibration.
 - B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
 - C. Certificate of Analysis 0.10% solution for same.
Lot #: 08 B050
 - D. Alcotest card of operator/coordinator who completed change.

Drugsafety

Alcotest[®] 7110 MKII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 88854, 49 FR 48884 and 58 FR 48705. The manufacturer recommends accuracy verification of the instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

08/04/2006

SERIAL NUMBER:

ARXB-0074

Dräger Safety Diagnostics, Inc.
Durango, CO

Art Dly



Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: _____

Serial Number:

1201553-0210

Certification Date

Technician

Re-Certification Due Date

10-08-2008

JS

10-08-2009

Dräger safety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

22WTP2-204

Certification date:

10-08-2008

Next Certification due:

10-08-2009

Probe Value

104

Dräger Safety Diagnostics, Inc.
Technical Service Department

JS



Calibration
Certificate No. 1750.01

**Calibration complies with ISO 9001
ISO/IEC 17025 AND ANSI/NCCL Z540-1**



Cert. No.: 4000-1967485

Traceable® Certificate of Calibration for Digital Thermometer

Instrument Identification:

Model: 61220-601 S/N: 80497135 Manufacturer : Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC191	A79341		
Thermistor Module	A27129	8/22/09	1000248949
Temperature Probe	3039	10/26/08	A7710039-4
Temperature Probe	149	3/06/09	A82225037-3
Thermistor Module	A17118	8/30/08	A7831032
Temperature Calibration Bath TC218	A73332		

Certificate Information:

Technician: 68 Procedure: CAL-06 Cal Date: 8/28/08 Cal Due: 8/28/10
 Test Conditions: 22.5°C 49.0 %RH 1013 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±uc	TUR
°C		N.A.		0.001	0.000	Y	-0.049	0.051	0.013	3.8:1
°C		N.A.		25.001	24.998	Y	24.951	25.051	0.013	3.8:1
°C		N.A.		60.001	59.999	Y	59.951	60.051	0.018	2.8:1
°C		N.A.		100.001	99.998	Y	99.951	100.051	0.013	3.8:1

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±uc=Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Date=MM/DD/YY

Wallace Berry
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA
 Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
 Control Company is ISO 9001:2000 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-AQ-HOU.
 International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

Alcotest 7110 Calibration Record

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE	Calib. No.:	00003
Calibration File No.:	00022	Cert. Date:	11/05/2008
Certification File No.:	00005	Lin. Date:	11/05/2008
Linearity File No.:	00006	Soln. Date:	03/30/2009
Solution File No.:	00021	File Date:	04/24/2009
Sequential File No.:	00022		
 		Serial No.:	DDXJ S3-0210
Calibrating Unit:	WET	Expires:	07/02/2010
Control Solution %:	0.100%	Bottle No.:	0477
Solution Control Lot:	08G056		


Coordinator

Last Name: SULLIVAN First Name: FRANCIS MI: -

Signature: Tp1 F.J. Sullivan # 5103 Badge No.: 5103 Date: 04/24/2009

*Black Key Temperature Probe Serial.....# DDMBPI-0002 25

*Digital NIST Temperature Measuring System Serial.....# 80497135 25



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator in the State of New Jersey, in my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks for approved instruments employing infrared analysis and electrochemical analysis. The instrument used in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00022	Calib. Date: 04/24/2009
Certification File No.:	00023	Calib. No.: 00003
Linearity File No.:	00006	Cert. Date: 04/24/2009
Solution File No.:	00021	Cert. No.: 00002
Sequential File No.:	00023	Lin. Date: 11/05/2008
		Lin. No.: 00001
		Soln. Date: 03/30/2009
		Soln. No.: 00009
		File Date: 04/24/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXJ S3-0210
Solution Control Lot:	08G056	Expires: 07/02/2010
		Bottle No.: 0477

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.100%	09:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 2 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 3 EC	0.099%	09:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		

All tests within acceptable tolerance.

Coordinator

Last Name: SULLIVAN

First Name: FRANCIS


MI: -

Signature: *Francis Sullivan*

Badge No.: 5103

Date: 04/24/2009

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator in my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I performed calibration checks on approved instruments employing infrared analysis and electrochemical analysis as well as included in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.



Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Serial No.: ARXB-0074

Calib. No.: 00003

Cert. No.: 00002

Lin. No.: 00002

Soln. No.: 00009

Calib. Date: 04/24/2009

Cert. Date: 04/24/2009

Lin. Date: 04/24/2009

Soln. Date: 03/30/2009

File Date: 04/24/2009

Equipment

Alcotest 7110 MKII-C

BUENA BOROUGH POLICE

Location:

Calibration File No.: 00022

Certification File No.: 00023

Linearity File No.: 00024

Solution File No.: 00021

Sequential File No.: 00024

Calibrating Unit:

WET

Control Solution %:

0.040%

Solution Control Lot:

08F054

Calibrating Unit:

WET

Control Solution %:

0.080%

Solution Control Lot:

08F055

Calibrating Unit:

WET

Control Solution %:

0.160%

Solution Control Lot:

08G057

Serial No.: DDWF S3-0241

Expires: 07/07/2010

Bottle No.: 0442

Serial No.: DDXD S3-0193

Expires: 06/28/2010

Bottle No.: 0477

Serial No.: DDWF S3-0215

Expires: 06/26/2010

Bottle No.: 1256

Function
Result
Time
Temperature
Simulator (°C)
Comment(s)
or Error(s)

Function	Result	Time	Temperature	Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:56D	34.0°C		*** TEST PASSED ***
Control 1 EC	0.042%	09:57D	34.0°C		*** TEST PASSED ***
Control 1 IR	0.039%	09:57D	34.0°C		*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D	34.0°C		*** TEST PASSED ***
Control 2 EC	0.041%	09:59D	34.0°C		*** TEST PASSED ***
Control 2 IR	0.040%	09:59D	34.0°C		*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D	34.0°C		*** TEST PASSED ***
Control 3 EC	0.082%	10:01D	34.0°C		*** TEST PASSED ***
Control 3 IR	0.080%	10:01D	34.0°C		*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D	34.0°C		*** TEST PASSED ***
Control 4 EC	0.081%	10:03D	34.0°C		*** TEST PASSED ***
Control 4 IR	0.080%	10:03D	34.0°C		*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D	34.0°C		*** TEST PASSED ***
Control 5 EC	0.163%	10:05D	34.0°C		*** TEST PASSED ***
Control 5 IR	0.162%	10:05D	34.0°C		*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D	34.0°C		*** TEST PASSED ***
Control 6 EC	0.162%	10:07D	34.0°C		*** TEST PASSED ***
Control 6 IR	0.161%	10:07D	34.0°C		*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D	34.0°C		*** TEST PASSED ***

All tests within acceptable tolerance.

Coordinator



Last Name: SULLIVAN

First Name: FRANCIS

Signature: *Francis Sullivan #5103*

Badge No.: 5103
Date: 04/24/2009

MI: -



CERTIFICATE

This is to certify that


Francis Sullivan
Trooper I, #5103

has successfully completed the two day Draeger Safety Diagnostics, Inc. Alcohol Coordinator Training Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: September 18, 2008



Instructor: Hansuelli Flyser

Dräger



State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
PO Box 080
TRENTON, NJ 08625-0080

JON S. CORZINE
Governor

ANNE MILGRAM
Attorney General

November 5, 2008

Col. Joseph R. Fuentes, Superintendent
Division of State Police
Division Headquarters
P.O. Box 7068
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification - Trooper I Francis Sullivan # 5103

Dear Col. Fuentes:

Pursuant to the provisions of *N.J.A.C.* 13:51-2.1 (b) and (c), as adopted and promulgated under the provisions of *N.J.S.A.* 39:4-50.3, 39:3-10.25 and 12:7-56, I hereby approve Trooper I Francis Sullivan # 5103 as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,


Anne Milgram
Attorney General

AM:HA:tlh

c: Trooper I Francis Sullivan # 5103, Alcohol/Drug Test Unit, Division of State Police
Lt. Paul Spirit, Unit Head, Alcohol/Drug Test Unit, Division of State Police

G:\Prosecutor Services\COPROSLETTERS\Sullivan Breath Test Ltr 11.5.08.wpd



DEPARTMENT OF
Water and Public Safety
 This is to certify that

OF THE STATE
Frank Sullivan
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN ACCORDANCE WITH CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE COURSE NUMBER 7110 MKIII-C
 A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 16th DAY OF March
 TWO THOUSAND AND SIX

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	05/26/08	Sayreville PD	<i>[Signature]</i>
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2908 (Rev. 0108)

DEPARTMENT OF
Law and Public Safety

This is to certify that

Frank J. Sullivan
 New Jersey State Police



IS QUALIFIED AND COMPETENT TO PERFORM THE ANALYSIS PURSUANT TO CHAPTER
 143 OF THE LAWS OF 1986 IN THE MATTER OF Deathly IN ACCORDANCE WITH
 DETERMINED INVESTIGATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29th DAY OF Apr.
 ONE THOUSAND NINE HUNDRED AND 94

C. A. Williams
 SUPERVISOR
 NEW JERSEY STATE POLICE

Robert T. Pitt
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course	INSTRUCTOR
1.	9-6-96	ACTC	J. J. [Signature]
2.	12-15-97	ACTC	J. J. [Signature]
3.	12-8-99	ACTC	[Signature]
4.	10-9-01	OCPP	[Signature]
5.	7-17-03	ACTC	[Signature]
6.	6-10-04	ACTC	[Signature]
7.	11-16-07	OCPP	[Signature]
8.			
9.			

Francis Sullivan #5103
Hamilton Technology Complex
Suite 400, 1200 Negron Drive
Hamilton, NJ 08691
fp5103@gw.njsp.org

Current Rank & Date of Promotion & Time of Service:

Trooper I 7/16/2002 16 years of Service

Career Experience/Assignment:

April 26, 2008 to Present:

Special Investigation Section
Alcohol/Drug Testing Unit

May 29 2004 to April 25 2008:

Special Operations Section
Motor Coach/ Compliance Review Unit

December 29 2001 to May 28 2004:

Special Operations Section
Diesel Emissions

November 3 2001 to December 28 2001:

Field Operations Section
Tuckerton Station
Troop A

April 1 1995 to November 2 2001:

Field Operations Section
Bass River Station
Troop E

October 15 1994 to March 31 1995:

Field Operations Section
Port Norris Station
Troop A

October 16 1993 to October 14 1994:

Field Operations Section
Tuckerton Station
Troop A

January 16, 1993 to October 15 1993:

Field Operations Section
Woodbine Station
Troop A

Specialized Training:

Breath Test Coordinator/Instructor, Certification

November 2008

Instructor Training in DWI Detection and Standardized Field Sobriety Testing	October 2008
Draeger Safety Diagnostics, Inc., Alcohol Coordinator Training Course on the New Jersey specific Alcotest 7110 MKIII-C	September 2008
The Robert F. Borkenstein Course on Alcohol and Highway Safety	May 2008
Inspection and Investigation of Commercial Vehicle Crashes	September 2007
Accident Investigation 1	January 2007
Accident Investigation 2	March 2007
Instructor Training Course	May 2006
Laser Operator Training Course	March 2006
Alcotest Certification	March 2006
Post Crash Vehicle Equipment Inspection	November 2004
Compliance Review Training	September 2003
Suspension and Steering	August 2003
Commercial Motor Vehicle Criminal Interdiction	April 2003
General Hazardous Material Inspection	December 2002
Domestic Cannabis Eradication & Suppression Seminar	July 2002
Trucks and Terrorism	May 2002
Motorcoach	May 2002
NASI-North American Standard Inspection Part B	February 2002
NASI-Weights & Dimensions	February 2002
NASI-North American Standard Inspection Part A	January 2002
PR-24 Training Course	July 1999
Narcotics Trafficking Trends Seminar	June 1998
Firearms Instructor Course	May 1997
Driving While Intoxicated/Standardized Field Sobriety	July 1996

Educational/Certification Summary:

Associates of Science Comm. College of the Air Force	August 1987
---	-------------

Commendation & Awards:

Certificate of Commendation	June 2002
-----------------------------	-----------

Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDMBPI-0002

02/03/2009

02/03/2010

Probe Value

112

Dräger Safety Diagnostics, Inc.
Technical Service Department





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

JON S. CORZINE
 Governor

ANNE MILGRAM
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.040 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution:

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 07/16/2008

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 08F054

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0487 to 0.0493 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 26, 2010.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare

 Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 19th day of August, 2008.

John R. Leaver
 Notary
 JOHN R. LEAVER
 NOTARY PUBLIC STATE OF NEW JERSEY
 MY COMMISSION EXPIRES 12/14/2012



JON S. CORZINE
Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
Post Office Box 7068
WEST TRENTON NJ 08625-0068
(609) 882-2000

ANNE MILGRAM
Attorney General
COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.080 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 07/25/2008

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 08R055

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0977 to 0.0984 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 28, 2010.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 19th day of August, 2008.

Notary

JOHN R. LEAVER
NOTARY PUBLIC STATE OF NEW JERSEY



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

JON S. CORZINE
 Governor

ANNE MILGRAM
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 07/24/2008

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 08G056

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1227 to 0.1236 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13-51-3.4, of approved breath test instruments (N.J.A.C. 13-51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 2, 2010.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare

 Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 13th day of August, 2008.

Linda L. DeSantis

 Notary

Linda L. DeSantis
 My Commission
 Expires August 25, 2010



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7058
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

JON S. CORZINE
 Governor

ANNE MILGRAM
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.160 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 07/28/2008

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 08G057

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1956 to 0.1959 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 7, 2010.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 19th day of August, 2008.

Notary

JOHN R. LEAVER
 NOTARY PUBLIC STATE OF NEW JERSEY
 LV COMMISSION EXPIRES 10/1/2010



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: _____

Serial Number:

DDWF53-0215

Certification Date

Technician

Re-Certification Due Date

03/19/2009

JS

03/19/2010



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDXDS3-0193

Certification Date

02-03-09

Technician

[Signature]

Re-Certification Due Date

02-03-10



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: _____

Serial Number:

DDWFS3-0241

Certification Date

02-03-09

Technician

[Signature]

Re-Certification Due Date

02-03-10

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00022	Calib. Date: 04/24/2009
Certification File No.:	00023	Calib. No.: 00003
Linearity File No.:	00024	Cert. Date: 04/24/2009
Solution File No.:	00025	Cert. No.: 00002
Sequential File No.:	00025	Lin. Date: 04/24/2009
		Lin. No.: 00002
		Soln. Date: 04/24/2009
		Soln. No.: 00010
		File Date: 04/24/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXJ S3-0210
Solution Control Lot:	08B050	Expires: 02/13/2010
		Bottle No.: 0483

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:16D		
Control 1 EC	0.102%	11:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		
Control 2 EC	0.100%	11:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19D		
Control 3 EC	0.100%	11:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DOWJ P2-20433

Changed By:

Last Name: SULLIVAN

First Name: FRANCIS

MI: -

Signature: *Tpr T. J. Sullivan #5103*

Badge No.: 5103

Date: 04/24/2009



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 WEST TUNTON NJ 08628-0068
 (609) 882-2000

JOHN S. CORZINE
 Governor

ANNE MILGRAM
 Attorney General

COLONEL JOSEPH R. FLORES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc. ANALYSIS DATE: 03/14/2008

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 08B050

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1229 to 0.1234 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 13, 2010.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
 Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 31 day of March, 2008.

Linda L. DeBartis

NOTARY
 Linda L. DeBartis
 My Commission
 Expires Aug. 17, 2009



New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable

