



STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF CONSUMER AFFAIRS
 OFFICE OF WEIGHTS AND MEASURES
 1261 ROUTES 1 & 9 SOUTH
 AVENEL, NEW JERSEY 07001



Date: December 5, 2011

TOWNSHIP OF BURLINGTON
 POLICE DEPARTMENT
 851 OLD YORK ROAD
 BURLINGTON, NJ 08016

Subject: TEST REPORT - NJ120331 - LASER SPEED GUN DISTANCE CALIBRATION COURSE

TEST OF ACCURACY ONLY

The course(s) which you submitted to the New Jersey State Metrology Laboratory were/was compared with the State Standard (NIST Test Number 821/263710-00) on this date.

<u>DESIGNATION</u>	<u>ACTUAL</u>	<u>ID MARKS</u>
CENTER OF BASE POST TO:		BTPD LC
CENTER OF FIRST TARGET (1)	50 FEET (+/- <3/16")	
CENTER OF SECOND TARGET (2)	75 FEET (+/- <3/16")	

This certificate shall not be reproduced except in full without the written permission of the State Superintendent of Weights and Measures.



Robert Campanelli
 Acting State Superintendent