

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00265 Calib. Date: 09/09/2009 Calib. No.: 00012
Certification File No.: 00266 Cert. Date: 09/09/2009 Cert. No.: 00007
Linearity File No.: 00267 Lin. Date: 09/09/2009 Lin. No.: 00007
Solution File No.: 00285 Soln. Date: 12/22/2009 Soln. No.: 00065
Sequential File No.: 00285 File Date: 12/22/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0472

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 09:49S | | |
| Control 1 EC | 0.097% | 09:49S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 09:49S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:50S | | |
| Control 2 EC | 0.098% | 09:51S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 09:51S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:52S | | |
| Control 3 EC | 0.097% | 09:52S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 09:52S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:53S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 12/22/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00265 **Calib. Date:** 09/09/2009 **Calib. No.:** 00012
Certification File No.: 00266 **Cert. Date:** 09/09/2009 **Cert. No.:** 00007
Linearity File No.: 00267 **Lin. Date:** 09/09/2009 **Lin. No.:** 00007
Solution File No.: 00282 **Soln. Date:** 11/24/2009 **Soln. No.:** 00064
Sequential File No.: 00282 **File Date:** 11/24/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0189
Control Solution %: 0.100% **Expires:** 06/09/2010
Solution Control Lot: 08F053 **Bottle No.:** 0451

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:55S | | |
| Control 1 EC | 0.098% | 10:55S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 10:55S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:56S | | |
| Control 2 EC | 0.097% | 10:57S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 10:57S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:57S | | |
| Control 3 EC | 0.097% | 10:58S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 10:58S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:59S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____

 #142

Badge No.: 16142

Date: 11/24/2009

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | | |
| Calibration File No.: | 00265 | Calib. Date: | 09/09/2009 |
| Certification File No.: | 00266 | Cert. Date: | 09/09/2009 |
| Linearity File No.: | 00267 | Lin. Date: | 09/09/2009 |
| Solution File No.: | 00280 | Soln. Date: | 10/27/2009 |
| Sequential File No.: | 00280 | File Date: | 10/27/2009 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWE S3-0189 |
| Solution Control Lot: | 08F053 | Expires: | 06/09/2010 |
| | | Bottle No.: | 0452 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:44S | | |
| Control 1 EC | 0.098% | 07:45S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 07:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:45S | | |
| Control 2 EC | 0.099% | 07:46S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 07:46S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:47S | | |
| Control 3 EC | 0.099% | 07:47S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 07:47S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:48S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 10/27/2009

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|------------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | |
| Calibration File No.: | 00265 | Calib. Date: 09/09/2009 |
| Certification File No.: | 00266 | Calib. No.: 00012 |
| Linearity File No.: | 00267 | Cert. Date: 09/09/2009 |
| Solution File No.: | 00276 | Cert. No.: 00007 |
| Sequential File No.: | 00276 | Lin. Date: 09/09/2009 |
| | | Lin. No.: 00007 |
| | | Soln. Date: 10/06/2009 |
| | | Soln. No.: 00062 |
| | | File Date: 10/06/2009 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDWE S3-0189 |
| Control Solution %: | 0.100% | | | Expires: | 06/09/2010 |
| Solution Control Lot: | 08F053 | | | Bottle No.: | 0473 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 10:43D | | |
| Control 1 EC | 0.097% | 10:44D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 10:44D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:44D | | |
| Control 2 EC | 0.098% | 10:45D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 10:45D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:46D | | |
| Control 3 EC | 0.097% | 10:46D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 10:46D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:47D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOMON

First Name: KINAMO

MI: A

Signature: _____

Badge No.: 16164

Date: 10/06/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00241 Calib. Date: 03/12/2009 Calib. No.: 00011
Certification File No.: 00242 Cert. Date: 03/12/2009 Cert. No.: 00006
Linearity File No.: 00243 Lin. Date: 03/12/2009 Lin. No.: 00006
Solution File No.: 00262 Soln. Date: 08/23/2009 Soln. No.: 00060
Sequential File No.: 00262 File Date: 08/23/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0471

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:44D | | |
| Control 1 EC | 0.100% | 10:44D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 10:44D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:45D | | |
| Control 2 EC | 0.100% | 10:45D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 10:45D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:46D | | |
| Control 3 EC | 0.099% | 10:47D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:47D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:47D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 08/23/2009

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | |
| Calibration File No.: | 00241 | Calib. Date: 03/12/2009 |
| Certification File No.: | 00242 | Calib. No.: 00011 |
| Linearity File No.: | 00243 | Cert. Date: 03/12/2009 |
| Solution File No.: | 00260 | Cert. No.: 00006 |
| Sequential File No.: | 00260 | Lin. Date: 03/12/2009 |
| | | Lin. No.: 00006 |
| | | Soln. Date: 07/26/2009 |
| | | Soln. No.: 00059 |
| | | File Date: 07/26/2009 |

| | | |
|-----------------------|--------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWE S3-0189 |
| Solution Control Lot: | 08F053 | Expires: 06/09/2010 |
| | | Bottle No.: 0465 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:57D | | |
| Control 1 EC | 0.099% | 10:58D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 10:58D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:59D | | |
| Control 2 EC | 0.099% | 10:59D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 10:59D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:00D | | |
| Control 3 EC | 0.099% | 11:01D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 11:01D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:01D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 07/26/2009

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | |
| Calibration File No.: | 00241 | Calib. Date: 03/12/2009 |
| Certification File No.: | 00242 | Cert. No.: 00011 |
| Linearity File No.: | 00243 | Cert. Date: 03/12/2009 |
| Solution File No.: | 00256 | Lin. No.: 00006 |
| Sequential File No.: | 00256 | Soln. Date: 06/28/2009 |
| | | Soln. No.: 00058 |
| | | File Date: 06/28/2009 |

| | | |
|-----------------------|--------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWE S3-0189 |
| Solution Control Lot: | 08F053 | Expires: 06/09/2010 |
| | | Bottle No.: 0469 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 09:46D | | |
| Control 1 EC | 0.099% | 09:46D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 09:46D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:47D | | |
| Control 2 EC | 0.099% | 09:48D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 09:48D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:48D | | |
| Control 3 EC | 0.099% | 09:49D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 09:49D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:50D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 06/28/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00241 Calib. Date: 03/12/2009 Calib. No.: 00011
Certification File No.: 00242 Cert. Date: 03/12/2009 Cert. No.: 00006
Linearity File No.: 00243 Lin. Date: 03/12/2009 Lin. No.: 00006
Solution File No.: 00250 Soln. Date: 06/05/2009 Soln. No.: 00057
Sequential File No.: 00250 File Date: 06/05/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0459

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 18:02D | | |
| Control 1 EC | 0.098% | 18:03D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 18:03D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:04D | | |
| Control 2 EC | 0.099% | 18:04D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 18:04D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:05D | | |
| Control 3 EC | 0.099% | 18:06D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 18:06D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:06D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 06/05/2009

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | |
| Calibration File No.: | 00241 | Calib. Date: 03/12/2009 |
| Certification File No.: | 00242 | Calib. No.: 00011 |
| Linearity File No.: | 00243 | Cert. Date: 03/12/2009 |
| Solution File No.: | 00249 | Cert. No.: 00006 |
| Sequential File No.: | 00249 | Lin. Date: 03/12/2009 |
| | | Lin. No.: 00006 |
| | | Soln. Date: 05/08/2009 |
| | | Soln. No.: 00056 |
| | | File Date: 05/08/2009 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWE S3-0189 |
| Solution Control Lot: | 08F053 | Expires: 06/09/2010 |
| | | Bottle No.: 0458 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 10:46D | | |
| Control 1 EC | 0.100% | 10:47D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 10:47D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:48D | | |
| Control 2 EC | 0.100% | 10:48D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 10:48D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:49D | | |
| Control 3 EC | 0.100% | 10:50D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 10:50D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:50D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 05/08/2009

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARWC-0060 |
| Location: | WILLINGBORO TWP POLICE | | |
| Calibration File No.: | 00241 | Calib. Date: | 03/12/2009 |
| Certification File No.: | 00242 | Cert. Date: | 03/12/2009 |
| Linearity File No.: | 00243 | Lin. Date: | 03/12/2009 |
| Solution File No.: | 00247 | Soln. Date: | 04/10/2009 |
| Sequential File No.: | 00247 | File Date: | 04/10/2009 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDWE S3-0189 |
| Control Solution %: | 0.100% | | | Expires: | 06/09/2010 |
| Solution Control Lot: | 08F053 | | | Bottle No.: | 0467 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:11D | | |
| Control 1 EC | 0.101% | 10:12D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 10:12D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:12D | | |
| Control 2 EC | 0.101% | 10:13D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 10:13D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:14D | | |
| Control 3 EC | 0.101% | 10:14D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:14D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:15D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 04/10/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00200 Calib. Date: 08/18/2008 Calib. No.: 00008
Certification File No.: 00201 Cert. Date: 08/18/2008 Cert. No.: 00005
Linearity File No.: 00202 Lin. Date: 08/18/2008 Lin. No.: 00005
Solution File No.: 00234 Soln. Date: 02/07/2009 Soln. No.: 00052
Sequential File No.: 00234 File Date: 02/07/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0457

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:49S | | |
| Control 1 EC | 0.096% | 08:49S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 08:49S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:50S | | |
| Control 2 EC | 0.097% | 08:51S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.096% | 08:51S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:52S | | |
| Control 3 EC | 0.096% | 08:52S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 08:52S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:53S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 02/07/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00200 **Calib. Date:** 08/18/2008 **Calib. No.:** 00008
Certification File No.: 00201 **Cert. Date:** 08/18/2008 **Cert. No.:** 00005
Linearity File No.: 00202 **Lin. Date:** 08/18/2008 **Lin. No.:** 00005
Solution File No.: 00231 **Soln. Date:** 01/30/2009 **Soln. No.:** 00051
Sequential File No.: 00231 **File Date:** 01/30/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0188
Control Solution %: 0.100% **Expires:** 06/09/2010
Solution Control Lot: 08F053 **Bottle No.:** 0463

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 09:29S | | |
| Control 1 EC | 0.096% | 09:30S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 09:30S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:30S | | |
| Control 2 EC | 0.097% | 09:31S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.096% | 09:31S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:32S | | |
| Control 3 EC | 0.097% | 09:33S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.096% | 09:33S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:33S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 01/30/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00200 Calib. Date: 08/18/2008 Calib. No.: 00008
Certification File No.: 00201 Cert. Date: 08/18/2008 Cert. No.: 00005
Linearity File No.: 00202 Lin. Date: 08/18/2008 Lin. No.: 00005
Solution File No.: 00225 Soln. Date: 01/02/2009 Soln. No.: 00050
Sequential File No.: 00225 File Date: 01/02/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0466

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:32S | | |
| Control 1 EC | 0.096% | 10:33S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.096% | 10:33S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:34S | | |
| Control 2 EC | 0.096% | 10:34S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.096% | 10:34S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:35S | | |
| Control 3 EC | 0.096% | 10:36S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.096% | 10:36S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:36S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

M1: A

Signature: _____

Sgt. Raymond Norfo "102"

Badge No.: 102

Date: 01/02/2009