

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
 Location: WILLINGBORO TWP POLICE  
 Calibration File No.: 00454 Calib. Date: 09/09/2013 Calib. No.: 00024  
 Certification File No.: 00455 Cert. Date: 09/09/2013 Cert. No.: 00018  
 Linearity File No.: 00456 Lin. Date: 09/09/2013 Lin. No.: 00018  
 Solution File No.: 00463 Soln. Date: 12/02/2013 Soln. No.: 00118  
 Sequential File No.: 00463 File Date: 12/02/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
 Control Solution %: 0.100% Expires: 01/05/2014  
 Solution Control Lot: 12A097 Bottle No.: 1256

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:40S		
Control 1 EC	0.098%	14:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41S		
Control 2 EC	0.098%	14:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42S		
Control 3 EC	0.099%	14:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DD UN P2 - 389 (na)

**Changed By:**

Last Name: GONCALVES First Name: MICHELLE MI: L

Signature: #1 Sgt. Michel #6040 Badge No.: 6040  
Date: 12/02/2013

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00454	Calib. Date: 09/09/2013
Certification File No.:	00455	Cert. No.: 00024
Linearity File No.:	00456	Cert. Date: 09/09/2013
Solution File No.:	00460	Lin. No.: 00018
Sequential File No.:	00460	Soln. No.: 00117
		File Date: 10/12/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 01/05/2014
Solution Control Lot:	12A097		Bottle No.: 1260

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:54D		
Control 1 EC	0.099%	03:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:55D		
Control 2 EC	0.099%	03:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57D		
Control 3 EC	0.099%	03:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 10/12/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00431	Calib. Date:	03/26/2013
Certification File No.:	00432	Cert. Date:	03/26/2013
Linearity File No.:	00433	Lin. Date:	03/26/2013
Solution File No.:	00445	Soln. Date:	08/13/2013
Sequential File No.:	00445	File Date:	08/13/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	12A097	Expires:	01/05/2014
		Bottle No.:	1251

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:58D		
Control 1 EC	0.099%	02:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:00D		
Control 2 EC	0.099%	03:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:01D		
Control 3 EC	0.099%	03:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:03D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: 

Badge No.: 16187

Date: 08/13/2013





# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
<b>Location:</b>	WILLINGBORO TWP POLICE		
<b>Calibration File No.:</b>	00431	<b>Calib. Date:</b>	03/26/2013
<b>Certification File No.:</b>	00432	<b>Cert. Date:</b>	03/26/2013
<b>Linearity File No.:</b>	00433	<b>Lin. Date:</b>	03/26/2013
<b>Solution File No.:</b>	00441	<b>Soln. Date:</b>	05/22/2013
<b>Sequential File No.:</b>	00441	<b>File Date:</b>	05/22/2013
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-14
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDWE 51-0189
<b>Solution Control Lot:</b>	12A097	<b>Expires:</b>	01/05/2014
		<b>Bottle No.:</b>	1265

Function	Result	Time	Temperature	Comment(s) or Error(s)
	%BAC	HH:MM	Simulator (°C)	
Ambient Air Blank	0.000%	03:11D		
Control 1 EC	0.099%	03:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:12D		
Control 2 EC	0.099%	03:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:14D		
Control 3 EC	0.099%	03:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGEE

First Name: WILLIAM

MI: I

Signature:  186

Badge No. : 186

Date: 05/22/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
<b>Location:</b>	WILLINGBORO TWP POLICE		
<b>Calibration File No.:</b>	00431	<b>Calib. Date:</b>	03/26/2013
<b>Certification File No.:</b>	00432	<b>Cert. No.:</b>	00020
<b>Linearity File No.:</b>	00433	<b>Cert. Date:</b>	03/26/2013
<b>Solution File No.:</b>	00438	<b>Lin. No.:</b>	00015
<b>Sequential File No.:</b>	00438	<b>Soln. Date:</b>	04/23/2013
		<b>Soln. No.:</b>	00110
		<b>File Date:</b>	04/23/2013

<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34	<b>Serial No.:</b>	DDWE 53-0189
<b>Control Solution %:</b>	0.100%			<b>Expires:</b>	01/05/2014
<b>Solution Control Lot:</b>	12A097			<b>Bottle No.:</b>	1270

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:55D		
Control 1 EC	0.099%	21:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:56D		
Control 2 EC	0.099%	21:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58D		
Control 3 EC	0.099%	21:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 04/23/2013

# Calibrating Unit


## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
<b>Location</b>	WILLINGBORO TWP POLICE	
<b>Calibration File No.:</b>	00403	<b>Calib. Date:</b> 10/01/2012
<b>Certification File No.:</b>	00404	<b>Calib. No.:</b> 00019
<b>Linearity File No.:</b>	00405	<b>Cert. Date:</b> 10/01/2012
<b>Solution File No.:</b>	00430	<b>Lim. No.:</b> 00014
<b>Sequential File No.:</b>	00430	<b>Lim. Date:</b> 10/01/2012
		<b>Soln. No.:</b> 00108
		<b>Soln. Date:</b> 03/21/2013
		<b>File Date:</b> 03/21/2013
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-14
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDWE 57-0149
<b>Solution Control Lot:</b>	12A097	<b>Expires:</b> 01/05/2014
		<b>Bottle No.:</b> 1274

Function	Result	Time	Temperature	Comment(s) or Error(s)
	%BAC	HH:MM	Simulator (°C)	
Ambient Air Blank	0.000%	20:37S		
Control 1 EC	0.100%	20:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39S		
Control 2 EC	0.100%	20:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40S		
Control 3 EC	0.100%	20:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

**Changed By:**  
 Last Name: REYNOLDS      First Name: TIMOTHY      MI: G  
 Signature:       Badge No.: 16142  
 Date: 03/21/2013



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKII-C	Serial No.: ARWC-0060
<b>Location:</b>	WILLINGBORO TWP POLICE	
<b>Calibration File No.:</b>	00403	<b>Calib. Date:</b> 10/01/2012
<b>Certification File No.:</b>	00404	<b>Calib. No.:</b> 00019
<b>Linearity File No.:</b>	00405	<b>Cert. Date:</b> 10/01/2012
<b>Solution File No.:</b>	00427	<b>Cert. No.:</b> 00014
<b>Sequential File No.:</b>	00427	<b>Lin. Date:</b> 10/01/2012
		<b>Lin. No.:</b> 00014
		<b>Soln. Date:</b> 02/20/2013
		<b>Soln. No.:</b> 00107
		<b>File Date:</b> 02/20/2013

<b>Calibrating Unit:</b>	WET	Model No.: CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDWE 53-0189
<b>Solution Control Lot:</b>	12A097	<b>Expires:</b> 01/05/2014
		<b>Batch No.:</b> 1269

Function	Result %BAC	Time HH:MM	Temperature Simulant (°C)	Comment(s) or Errors!
Ambient Air Blank	0.000%	14:18S		
Control 1 EC	0.101%	14:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:20S		
Control 2 EC	0.101%	14:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21S		
Control 3 EC	0.100%	14:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISI) Chief Forensic Scientist.

Temperature Probe Serial Number DDUNP2-293 (New)

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature:

*Det I Michelles #6040*

Badge No.: 6040

Date: 02/20/2013



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC 0060
<b>Location:</b>	WILLINGBORO TWP POLICE	
<b>Calibration File No.:</b>	00403	<b>Calib. Date:</b> 10/01/2012
<b>Certification File No.:</b>	00404	<b>Calib. No.:</b> 00019
<b>Linearity File No.:</b>	00405	<b>Cert. Date:</b> 10/01/2012
<b>Solution File No.:</b>	00422	<b>Cert. No.:</b> 00014
<b>Sequential File No.:</b>	00422	<b>Lim. Date:</b> 10/01/2012
		<b>Lim. No.:</b> 00014
		<b>Soln. Date:</b> 01/01/2013
		<b>Soln. No.:</b> 00005
		<b>File Date:</b> 01/01/2013
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDWE S3-0188
<b>Solution Control Lot:</b>	12A097	<b>Expires:</b> 01/05/2014
		<b>Drule No.:</b> 1263

Function	Result	Time	Temperature	Comment (if
	%BAC	HH:MM	Simulator (°C)	or Error(s))
Ambient Air Blank	0.000%	06:14S		
Control 1 EC	0.100%	06:15S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:15S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:16S		
Control 2 EC	0.101%	06:16S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:16S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:17S		
Control 3 EC	0.101%	06:18S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:18S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 01/01/2013