

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00560	Calib. Date:	10/13/2016
Certification File No.:	00561	Cert. Date:	10/13/2016
Linearity File No.:	00562	Lin. Date:	10/13/2016
Solution File No.:	00568	Soln. Date:	12/09/2016
Sequential File No.:	00568	File Date:	12/09/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0782

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:03S		
Control 1 EC	0.103%	09:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04S		
Control 2 EC	0.101%	09:05S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06S		
Control 3 EC	0.102%	09:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 12/09/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00560	Calib. Date:	10/13/2016
Certification File No.:	00561	Cert. Date:	10/13/2016
Linearity File No.:	00562	Lin. Date:	10/13/2016
Solution File No.:	00565	Soln. Date:	11/10/2016
Sequential File No.:	00565	File Date:	11/10/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0781

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:53S		
Control 1 EC	0.081%	21:54S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.085%	21:54S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	21:55S		
Control 2 EC	0.090%	21:56S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.092%	21:56S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	21:57S		
Control 3 EC	0.095%	21:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:59S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 11/10/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00560	Calib. Date:	10/13/2016
Certification File No.:	00561	Cert. Date:	10/13/2016
Linearity File No.:	00562	Lin. Date:	10/13/2016
Solution File No.:	00566	Soln. Date:	11/11/2016
Sequential File No.:	00566	File Date:	11/11/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0784

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:35S		
Control 1 EC	0.103%	00:36S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37S		
Control 2 EC	0.100%	00:37S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:38S		
Control 3 EC	0.100%	00:39S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 11/11/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00543	Calib. Date: 05/10/2016
Certification File No.:	00544	Cert. No.: 00031
Linearity File No.:	00545	Cert. Date: 05/10/2016
Solution File No.:	00559	Lin. Date: 05/10/2016
Sequential File No.:	00559	Lin. No.: 00024
		Soln. Date: 10/03/2016
		Soln. No.: 00161
		File Date: 10/03/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0641

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:06D		
Control 1 EC	0.100%	20:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:07D		
Control 2 EC	0.098%	20:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 3 EC	0.098%	20:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 10/03/2016



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDWES3-0189

Certification Date

8-11-16

Technician

BC

Re-Certification Due Date

8-11-17

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY
ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUNP2-293

Certification date:

8-10-16

Next Certification due:

8-10-17

Probe Value

101

Draeger Safety Diagnostics, Inc.
Technical Service Department

BC

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00543 Calib. Date: 05/10/2016 Calib. No.: 00031
Certification File No.: 00544 Cert. Date: 05/10/2016 Cert. No.: 00024
Linearity File No.: 00545 Lin. Date: 05/10/2016 Lin. No.: 00024
Solution File No.: 00554 Soln. Date: 08/04/2016 Soln. No.: 00158
Sequential File No.: 00554 File Date: 08/04/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 0638

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:15D		
Control 1 EC	0.100%	20:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:16D		
Control 2 EC	0.099%	20:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17D		
Control 3 EC	0.099%	20:18D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 08/04/2016

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00543 Calib. Date: 05/10/2016 Calib. No.: 00031
Certification File No.: 00544 Cert. Date: 05/10/2016 Cert. No.: 00024
Linearity File No.: 00545 Lin. Date: 05/10/2016 Lin. No.: 00024
Solution File No.: 00552 Soln. Date: 07/07/2016 Soln. No.: 00157
Sequential File No.: 00552 File Date: 07/07/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 0639

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:36D		
Control 1 EC	0.100%	01:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:37D		
Control 2 EC	0.100%	01:38D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:38D		
Control 3 EC	0.099%	01:39D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:40D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: 

Badge No.: 16187

Date: 07/07/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00543	Calib. Date:	05/10/2016
Certification File No.:	00544	Cert. Date:	05/10/2016
Linearity File No.:	00545	Lin. Date:	05/10/2016
Solution File No.:	00549	Soln. Date:	06/03/2016
Sequential File No.:	00549	File Date:	06/03/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	06/13/2017
Solution Control Lot:	15F139			Bottle No.:	1233

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:05D		
Control 1 EC	0.095%	14:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	14:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:06D		
Control 2 EC	0.095%	14:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	14:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:08D		
Control 3 EC	0.094%	14:08D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.095%	14:08D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:09D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 06/03/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00526	Calib. Date:	01/04/2016
Certification File No.:	00527	Cert. Date:	01/04/2016
Linearity File No.:	00528	Lin. Date:	01/04/2016
Solution File No.:	00541	Soln. Date:	04/05/2016
Sequential File No.:	00541	File Date:	04/05/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	16A150	Expires:	01/05/2017
		Bottle No.:	0900

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:02D		
Control 1 EC	0.100%	20:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03D		
Control 2 EC	0.099%	20:04D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:05D		
Control 3 EC	0.098%	20:06D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 04/05/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00526	Calib. Date:	01/04/2016
Certification File No.:	00527	Cert. No.:	00023
Linearity File No.:	00528	Lin. Date:	01/04/2016
Solution File No.:	00540	Soln. No.:	00152
Sequential File No.:	00540	File Date:	03/08/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE 53-0188
Solution Control Lot:	14L134	Expires:	12/04/2016
		Bottle No.:	0754

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:49S		
Control 1 EC	0.100%	13:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50S		
Control 2 EC	0.100%	13:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:52S		
Control 3 EC	0.098%	13:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NUSP) Chief Forensic Scientist.

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature:

Sgt. Michelle Goncalves #6040

Badge No.: 6040

Date: 03/08/2016



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRINTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

JOHN J. HOFFMAN
Acting Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION; Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 11/18/2014

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 14L134

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1205 to 0.1218 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is December 04, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature]

AJF M. Alarulo, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 19th day of December, 2014.

[Handwritten signature of Linda L. DeSantis]
Notary

LINDA L DESANTIS
ID # 2228599
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires August 17, 2018



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00526	Calib. Date:	01/04/2016
Certification File No.:	00527	Cert. Date:	01/04/2016
Linearity File No.:	00528	Lin. Date:	01/04/2016
Solution File No.:	00539	Soln. Date:	03/07/2016
Sequential File No.:	00539	File Date:	03/07/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE 53-0188
Solution Control Lot:	15F139	Expires:	05/13/2017
		Bottle No.:	1229

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:07S		
Control 1 EC	0.096%	22:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:08S		
Control 2 EC	0.096%	22:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:10S		
Control 3 EC	0.096%	22:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECITT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 15187

Date: 03/07/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00526	Calib. Date:	01/04/2016
Certification File No.:	00527	Cert. No.:	00023
Linearity File No.:	00528	Lin. No.:	00023
Solution File No.:	00538	Soln. No.:	00150
Sequential File No.:	00538	File Date:	03/07/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	15F139	Expires:	06/13/2017
		Bottle No.:	1230

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:13S		
Control 1 EC	0.096%	20:13S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	20:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:14S		
Control 2 EC	0.096%	20:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	20:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15S		
Control 3 EC	0.094%	20:16S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.094%	20:16S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:17S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

ML: A

Signature: _____



Badge No.: 16187

Date: 03/07/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00526	Calib. Date:	01/04/2016
Certification File No.:	00527	Cert. Date:	01/04/2016
Linearity File No.:	00528	Lin. Date:	01/04/2016
Solution File No.:	00535	Soln. Date:	02/14/2016
Sequential File No.:	00535	File Date:	02/14/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDW16 S3-0188
Solution Control Lot:	15F139	Expires:	06/12/2017
		Drift No.:	1234

Function	Result	Time	Temperature	Comment(s)
	%BAC	HR:MM	Simulator (°C)	or Temp (°C)
Ambient Air Blank	0.000%	03:43S		
Control 1 BC	0.099%	03:43S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:44S		
Control 2 BC	0.097%	03:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	03:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:46S		
Control 3 BC	0.096%	03:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	03:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: *J. Kn*

Dodge No. : 16187

Date: 02/14/2016

Calibrating Unit New Standard Solution Report

Equipment: Alcotest 7110 MKIII-C	Serial No.: ARWC-0050
Location: WILLINGBORO TWP POLICE	
Calibration File No.: 00514	Calib. Date: 07/17/2015
Certification File No.: 00515	Cert. No.: 00029
Linearity File No.: 00516	Cert. No.: 00022
Solution File No.: 00525	Lin. No.: 00022
Sequential File No.: 00525	Soln. No.: 00145
	File Date: 01/01/2016
Calibrating Unit: WET	Model No.: CTF-34
Control Solution %: 0.100%	Serial No.: DJDWE S3-0122
Solution Control Lot: 15F139	Expires: 05/13/2017
	Bottle No.: 1217

Function	Result	Time	Temperature	Comment(s) or Error(s)
	%BAC	HH:MM	Simulator (°C)	
Ambient Air Blank	0.000%	19:53S		
Control 1 RC	0.099%	19:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54S		
Control 2 RC	0.096%	19:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	19:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56S		
Control 3 RC	0.096%	19:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	19:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 01/01/2016