

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00589	Calib. Date: 09/12/2017
Certification File No.:	00590	Calib. No.: 00034
Linearity File No.:	00591	Cert. Date: 09/12/2017
Solution File No.:	00601	Cert. No.: 00027
Sequential File No.:	00601	Lin. Date: 09/12/2017
		Lin. No.: 00027
		Soln. Date: 12/13/2017
		Soln. No.: 00178
		File Date: 12/13/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	05/17/2018
Solution Control Lot:	201605D			Bottle No.:	0298

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:59S		
Control 1 EC	0.102%	21:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	21:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:01S		
Control 2 EC	0.100%	21:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	21:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02S		
Control 3 EC	0.100%	21:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____

Badge No.: 16142

Date: 12/13/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00589	Calib. Date:	09/12/2017
Certification File No.:	00590	Cert. No.:	00034
Linearity File No.:	00591	Cert. Date:	09/12/2017
Solution File No.:	00596	Lin. Date:	09/12/2017
Sequential File No.:	00596	Lin. No.:	00027
		Soln. Date:	11/15/2017
		Soln. No.:	00177
		File Date:	11/15/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0300

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39S		
Control 1 EC	0.101%	11:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40S		
Control 2 EC	0.100%	11:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41S		
Control 3 EC	0.100%	11:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 11/15/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00589	Calib. Date:	09/12/2017
Certification File No.:	00590	Cert. Date:	09/12/2017
Linearity File No.:	00591	Lin. Date:	09/12/2017
Solution File No.:	00593	Soln. Date:	10/14/2017
Sequential File No.:	00593	File Date:	10/14/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0302

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:32D		
Control 1 EC	0.098%	09:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:34D		
Control 2 EC	0.097%	09:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:35D		
Control 3 EC	0.097%	09:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____

Sst / K 107

Badge No.: 16187

Date: 10/14/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00574	Calib. Date:	03/01/2017
Certification File No.:	00575	Cert. Date:	03/01/2017
Linearity File No.:	00576	Lin. Date:	03/01/2017
Solution File No.:	00586	Soln. Date:	07/21/2017
Sequential File No.:	00586	File Date:	07/21/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	1099

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:25D		
Control 1 EC	0.101%	09:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		
Control 2 EC	0.099%	09:27D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		
Control 3 EC	0.099%	09:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 07/21/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00574	Calib. Date: 03/01/2017
Certification File No.:	00575	Calib. No.: 00033
Linearity File No.:	00576	Cert. Date: 03/01/2017
Solution File No.:	00584	Cert. No.: 00026
Sequential File No.:	00584	Lin. Date: 03/01/2017
		Lin. No.: 00026
		Soln. Date: 06/22/2017
		Soln. No.: 00173
		File Date: 06/22/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 1103

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:51D		
Control 1 EC	0.102%	14:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		
Control 2 EC	0.099%	14:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		
Control 3 EC	0.099%	14:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 06/22/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00574	Calib. Date:	03/01/2017
Certification File No.:	00575	Cert. Date:	03/01/2017
Linearity File No.:	00576	Lin. Date:	03/01/2017
Solution File No.:	00582	Soln. Date:	05/24/2017
Sequential File No.:	00582	File Date:	05/24/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	1104

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:06D		
Control 1 EC	0.099%	19:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:07D		
Control 2 EC	0.100%	19:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:08D		
Control 3 EC	0.099%	19:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 05/24/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00574	Calib. Date:	03/01/2017
Certification File No.:	00575	Cert. Date:	03/01/2017
Linearity File No.:	00576	Lin. Date:	03/01/2017
Solution File No.:	00579	Soln. Date:	04/27/2017
Sequential File No.:	00579	File Date:	04/27/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	1102

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:42D		
Control 1 EC	0.100%	16:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:43D		
Control 2 EC	0.098%	16:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45D		
Control 3 EC	0.099%	16:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____

Badge No.: 16187

Date: 04/27/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00574	Calib. Date:	03/01/2017
Certification File No.:	00575	Cert. No.:	00033
Linearity File No.:	00576	Cert. Date:	03/01/2017
Solution File No.:	00578	Lin. No.:	00026
Sequential File No.:	00578	Soln. Date:	03/29/2017
		Soln. No.:	00169
		File Date:	03/29/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	1101

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:54S		
Control 1 EC	0.102%	07:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 2 EC	0.100%	07:56S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56S		
Control 3 EC	0.099%	07:57S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 03/29/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00560	Calib. Date: 10/13/2016
Certification File No.:	00561	Calib. No.: 00032
Linearity File No.:	00562	Cert. Date: 10/13/2016
Solution File No.:	00573	Cert. No.: 00025
Sequential File No.:	00573	Lin. Date: 10/13/2016
		Lin. No.: 00025
		Soln. Date: 02/06/2017
		Soln. No.: 00167
		File Date: 02/06/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 0780

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:11S		
Control 1 EC	0.102%	10:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S		
Control 2 EC	0.102%	10:14S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14S		
Control 3 EC	0.102%	10:15S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: Sgt. J. Knecht 187

Badge No.: 16187

Date: 02/06/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00560	Calib. Date: 10/13/2016
Certification File No.:	00561	Calib. No.: 00032
Linearity File No.:	00562	Cert. Date: 10/13/2016
Solution File No.:	00570	Cert. No.: 00025
Sequential File No.:	00570	Lin. Date: 10/13/2016
		Lin. No.: 00025
		Soln. Date: 01/05/2017
		Soln. No.: 00166
		File Date: 01/05/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 0783

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:17S		
Control 1 EC	0.101%	09:18S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19S		
Control 2 EC	0.101%	09:19S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20S		
Control 3 EC	0.101%	09:21S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____

Sgt. Jennifer Knecht

Badge No.: 16187

Date: 01/05/2017