

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00620	Calib. Date: 07/12/2018
Certification File No.:	00621	Cert. No.: 00036
Linearity File No.:	00622	Cert. Date: 07/12/2018
Solution File No.:	00631	Lin. No.: 00029
Sequential File No.:	00631	Lin. Date: 07/12/2018
		Soln. No.: 00194
		Soln. Date: 12/10/2018
		File Date: 12/10/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1121

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:10S		
Control 1 EC	0.103%	21:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	21:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:11S		
Control 2 EC	0.103%	21:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	21:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:13S		
Control 3 EC	0.101%	21:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 12/10/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00620	Calib. Date:	07/12/2018
Certification File No.:	00621	Cert. Date:	07/12/2018
Linearity File No.:	00622	Lin. Date:	07/12/2018
Solution File No.:	00629	Soln. Date:	11/08/2018
Sequential File No.:	00629	File Date:	11/08/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	0961

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:03S		
Control 1 EC	0.103%	01:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.104%	01:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:05S		
Control 2 EC	0.101%	01:05S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	01:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:06S		
Control 3 EC	0.102%	01:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	01:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_

Sgt Knecht JP

Badge No.: 16187

Date: 11/08/2018

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE	Calib. No.:	00036
Calibration File No.:	00620	Calib. Date:	07/12/2018
Certification File No.:	00621	Cert. Date:	07/12/2018
Linearity File No.:	00622	Lin. Date:	07/12/2018
Solution File No.:	00628	Soln. Date:	10/08/2018
Sequential File No.:	00628	File Date:	10/08/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1122

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:58D		
Control 1 EC	0.103%	19:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.104%	19:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59D		
Control 2 EC	0.101%	20:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00D		
Control 3 EC	0.101%	20:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_

Badge No.: 16142

Date: 10/08/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00620	Calib. Date: 07/12/2018
Certification File No.:	00621	Calib. No.: 00036
Linearity File No.:	00622	Cert. Date: 07/12/2018
Solution File No.:	00627	Cert. No.: 00029
Sequential File No.:	00627	Lin. Date: 07/12/2018
		Lin. No.: 00029
		Soln. Date: 09/08/2018
		Soln. No.: 00190
		File Date: 09/08/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1003

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:34D		
Control 1 EC	0.103%	20:35D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	20:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:36D		
Control 2 EC	0.100%	20:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:37D		
Control 3 EC	0.102%	20:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 09/08/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00620	Calib. Date: 07/12/2018
Certification File No.:	00621	Calib. No.: 00036
Linearity File No.:	00622	Cert. Date: 07/12/2018
Solution File No.:	00626	Cert. No.: 00029
Sequential File No.:	00626	Lin. Date: 07/12/2018
		Lin. No.: 00029
		Soln. Date: 08/11/2018
		Soln. No.: 00189
		File Date: 08/11/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1090

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:43D		
Control 1 EC	0.104%	01:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	01:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:45D		
Control 2 EC	0.101%	01:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	01:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:46D		
Control 3 EC	0.101%	01:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 08/11/2018

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00606	Calib. Date: 01/19/2018
Certification File No.:	00607	Calib. No.: 00035
Linearity File No.:	00608	Cert. Date: 01/19/2018
Solution File No.:	00618	Cert. No.: 00028
Sequential File No.:	00618	Lin. Date: 01/19/2018
		Lin. No.: 00028
		Soln. Date: 06/23/2018
		Soln. No.: 00187
		File Date: 06/23/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1106

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:19D		
Control 1 EC	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		
Control 2 EC	0.097%	17:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:22D		
Control 3 EC	0.097%	17:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 06/23/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00606	Calib. Date:	01/19/2018
Certification File No.:	00607	Cert. Date:	01/19/2018
Linearity File No.:	00608	Lin. Date:	01/19/2018
Solution File No.:	00617	Soln. Date:	05/21/2018
Sequential File No.:	00617	File Date:	05/21/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1250

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:40D		
Control 1 EC	0.097%	17:41D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	17:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:42D		
Control 2 EC	0.096%	17:42D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	17:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:43D		
Control 3 EC	0.095%	17:44D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 05/21/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00606	Calib. Date: 01/19/2018
Certification File No.:	00607	Calib. No.: 00035
Linearity File No.:	00608	Cert. Date: 01/19/2018
Solution File No.:	00614	Cert. No.: 00028
Sequential File No.:	00614	Lin. Date: 01/19/2018
		Lin. No.: 00028
		Soln. Date: 04/19/2018
		Soln. No.: 00185
		File Date: 04/19/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1097

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:39D		
Control 1 EC	0.098%	10:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40D		
Control 2 EC	0.096%	10:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42D		
Control 3 EC	0.096%	10:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: \_\_\_\_\_



Badge No.: 16187

Date: 04/19/2018



# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0060  
Location: WILLINGBORO TWP POLICE  
Calibration File No.: 00606 Calib. Date: 01/19/2018 Calib. No.: 00035  
Certification File No.: 00607 Cert. Date: 01/19/2018 Cert. No.: 00028  
Linearity File No.: 00608 Lin. Date: 01/19/2018 Lin. No.: 00028  
Solution File No.: 00613 Soln. Date: 03/20/2018 Soln. No.: 00184  
Sequential File No.: 00613 File Date: 03/20/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188  
Control Solution %: 0.100% Expires: 10/17/2019  
Solution Control Lot: 17360 Bottle No.: 1108

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:09S		
Control 1 EC	0.098%	02:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:11S		
Control 2 EC	0.096%	02:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:12S		
Control 3 EC	0.098%	02:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	02:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: J + J. Knecht 187

Badge No.: 16187

Date: 03/20/2018

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00606	Calib. Date: 01/19/2018
Certification File No.:	00607	Cert. No.: 00035
Linearity File No.:	00608	Cert. Date: 01/19/2018
Solution File No.:	00611	Lin. Date: 01/19/2018
Sequential File No.:	00611	Lin. No.: 00028
		Soln. Date: 02/17/2018
		Soln. No.: 00183
		File Date: 02/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1251

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:25S		
Control 1 EC	0.097%	21:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:26S		
Control 2 EC	0.098%	21:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27S		
Control 3 EC	0.096%	21:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 02/17/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00589	Calib. Date: 09/12/2017
Certification File No.:	00590	Calib. No.: 00034
Linearity File No.:	00591	Cert. Date: 09/12/2017
Solution File No.:	00605	Cert. No.: 00027
Sequential File No.:	00605	Lin. Date: 09/12/2017
		Lin. No.: 00027
		Soln. Date: 01/12/2018
		Soln. No.: 00181
		File Date: 01/12/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 0297

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:13S		
Control 1 EC	0.104%	06:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:15S		
Control 2 EC	0.101%	06:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	06:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:16S		
Control 3 EC	0.101%	06:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	06:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: \_\_\_\_\_



Badge No.: 16142

Date: 01/12/2018