

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	00989	Calib. Date:	08/24/2018
Certification File No.:	00990	Cert. No.:	00037
Linearity File No.:	00991	Cert. Date:	08/24/2018
Solution File No.:	01006	Lin. No.:	00029
Sequential File No.:	01006	Lin. Date:	08/24/2018
		Soln. Date:	12/14/2018
		Soln. No.:	00220
		File Date:	12/14/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0225
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	0757

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:29S		
Control 1 EC	0.099%	22:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:31S		
Control 2 EC	0.098%	22:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:32S		
Control 3 EC	0.098%	22:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: _____



Badge No.: 4047

Date: 12/14/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	00989	Calib. Date:	08/24/2018
Certification File No.:	00990	Cert. Date:	08/24/2018
Linearity File No.:	00991	Lin. Date:	08/24/2018
Solution File No.:	01004	Soln. Date:	11/28/2018
Sequential File No.:	01004	File Date:	11/28/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0225
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0979

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:17S		
Control 1 EC	0.096%	02:17S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	02:17S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:18S		
Control 2 EC	0.096%	02:19S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:20S		
Control 3 EC	0.096%	02:20S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CZEPIEJ

First Name: NICHOLAS

MI: B

Signature: _____



Badge No.: 4041

Date: 11/28/2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	00989	Calib. Date:	08/24/2018
Certification File No.:	00990	Cert. Date:	08/24/2018
Linearity File No.:	00991	Lin. Date:	08/24/2018
Solution File No.:	01000	Soln. Date:	11/09/2018
Sequential File No.:	01000	File Date:	11/09/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0225
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	1335

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:22S		
Control 1 EC	0.099%	06:22S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:22S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23S		
Control 2 EC	0.097%	06:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:24S		
Control 3 EC	0.098%	06:25S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CORRENTI

First Name: MAURO

MI: J

Signature: _____



Badge No.: 4025

Date: 11/09/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	
Calibration File No.:	00989	Calib. Date: 08/24/2018
Certification File No.:	00990	Calib. No.: 00037
Linearity File No.:	00991	Cert. Date: 08/24/2018
Solution File No.:	00997	Cert. No.: 00029
Sequential File No.:	00997	Lin. Date: 08/24/2018
		Lin. No.: 00029
		Soln. Date: 10/11/2018
		Soln. No.: 00217
		File Date: 10/11/2018

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUM S3-0225
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 1108

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:39D		
Control 1 EC	0.100%	23:40D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:40D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:41D		
Control 2 EC	0.098%	23:41D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:41D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:42D		
Control 3 EC	0.098%	23:43D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:43D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CZEPIEL

First Name: NICHOLAS

MI: B

Signature: _____



Badge No.: 4041

Date: 10/11/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	
Calibration File No.:	00989	Calib. Date: 08/24/2018
Certification File No.:	00990	Calib. No.: 00037
Linearity File No.:	00991	Cert. Date: 08/24/2018
Solution File No.:	00994	Cert. No.: 00029
Sequential File No.:	00994	Lin. Date: 08/24/2018
		Lin. No.: 00029
		Soln. Date: 09/13/2018
		Soln. No.: 00216
		File Date: 09/13/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUM S3-0225
Control Solution %:	0.100%		Expires: 01/31/2020
Solution Control Lot:	18050		Bottle No.: 1352

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:42D		
Control 1 EC	0.100%	00:42D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:42D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:43D		
Control 2 EC	0.099%	00:44D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:44D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:44D		
Control 3 EC	0.099%	00:45D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:45D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CORRENTI

First Name: MAURO

MI: J

Signature: _____



Badge No.: 4025

Date: 09/13/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	Calib. No.:	00036
Calibration File No.:	00972	Calib. Date:	04/18/2018
Certification File No.:	00973	Cert. Date:	04/18/2018
Linearity File No.:	00974	Lin. Date:	04/18/2018
Solution File No.:	00985	Soln. Date:	08/16/2018
Sequential File No.:	00985	File Date:	08/16/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0222
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0023

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:19D		
Control 1 EC	0.100%	01:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:21D		
Control 2 EC	0.099%	01:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22D		
Control 3 EC	0.099%	01:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SHAFER

First Name: JASON

MI: A

Signature: _____



Badge No.: 4045

Date: 08/16/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 00972 Calib. Date: 04/18/2018 Calib. No.: 00036
Certification File No.: 00973 Cert. Date: 04/18/2018 Cert. No.: 00028
Linearity File No.: 00974 Lin. Date: 04/18/2018 Lin. No.: 00028
Solution File No.: 00984 Soln. Date: 07/14/2018 Soln. No.: 00213
Sequential File No.: 00984 File Date: 07/14/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0222
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 0365

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:51D		
Control 1 EC	0.099%	23:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:53D		
Control 2 EC	0.098%	23:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:54D		
Control 3 EC	0.097%	23:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: YOUNG

First Name: BRIAN

MI: R

Signature: _____



Badge No.: 4034

Date: 07/14/2018

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 00972 Calib. Date: 04/18/2018 Calib. No.: 00036
Certification File No.: 00973 Cert. Date: 04/18/2018 Cert. No.: 00028
Linearity File No.: 00974 Lin. Date: 04/18/2018 Lin. No.: 00028
Solution File No.: 00979 Soln. Date: 06/27/2018 Soln. No.: 00212
Sequential File No.: 00979 File Date: 06/27/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0222
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 1342

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:28D		
Control 1 EC	0.102%	03:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:29D		
Control 2 EC	0.100%	03:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:30D		
Control 3 EC	0.099%	03:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: CZEPIEL First Name: NICHOLAS MI: B
Signature:  Badge No.: 4041
Date: 06/27/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	00951	Calib. Date:	11/08/2017
Certification File No.:	00952	Cert. Date:	11/08/2017
Linearity File No.:	00953	Lin. Date:	11/08/2017
Solution File No.:	00970	Soln. Date:	04/07/2018
Sequential File No.:	00970	File Date:	04/07/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0222
Solution Control Lot:	17110	Expires:	03/20/2019
		Bottle No.:	1494

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:42D		
Control 1 EC	0.100%	02:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44D		
Control 2 EC	0.098%	02:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		
Control 3 EC	0.098%	02:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SHAFER

First Name: JASON

MI: A

Signature: _____



Badge No.: 4045

Date: 04/07/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 00951 **Calib. Date:** 11/08/2017 **Calib. No.:** 00035
Certification File No.: 00952 **Cert. Date:** 11/08/2017 **Cert. No.:** 00027
Linearity File No.: 00953 **Lin. Date:** 11/08/2017 **Lin. No.:** 00027
Solution File No.: 00966 **Soln. Date:** 03/10/2018 **Soln. No.:** 00208
Sequential File No.: 00966 **File Date:** 03/10/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUM S3-0222
Control Solution %: 0.100% **Expires:** 03/20/2019
Solution Control Lot: 17110 **Bottle No.:** 1469

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:33S		
Control 1 EC	0.099%	00:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35S		
Control 2 EC	0.099%	00:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:36S		
Control 3 EC	0.100%	00:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CZEPIEL

First Name: NICHOLAS

MI: B

Signature: _____

Badgc No.: 4041

Date: 03/10/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	00951	Calib. Date:	11/08/2017
Certification File No.:	00952	Cert. Date:	11/08/2017
Linearity File No.:	00953	Lin. Date:	11/08/2017
Solution File No.:	00964	Soln. Date:	02/09/2018
Sequential File No.:	00964	File Date:	02/09/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0222
Solution Control Lot:	17110	Expires:	03/20/2019
		Bottle No.:	1466

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:57S		
Control 1 EC	0.100%	03:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58S		
Control 2 EC	0.099%	03:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:59S		
Control 3 EC	0.099%	04:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCULLY

First Name: DANIEL

MI: S

Signature: _____



Badge No.: 4065

Date: 02/09/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	00951	Calib. Date:	11/08/2017
Certification File No.:	00952	Cert. Date:	11/08/2017
Linearity File No.:	00953	Lin. Date:	11/08/2017
Solution File No.:	00963	Soln. Date:	02/09/2018
Sequential File No.:	00963	File Date:	02/09/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0222
Solution Control Lot:	17110	Expires:	03/20/2019
		Bottle No.:	1466

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:28S		
Control 1 EC	-.---%	01:28S	23.2°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	01:28S	23.2°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCULLY

First Name: DANIEL

MI: S

Signature: _____



Badge No.: 4065

Date: 02/09/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	
Calibration File No.:	00951	Calib. Date: 11/08/2017
Certification File No.:	00952	Calib. No.: 00035
Linearity File No.:	00953	Cert. Date: 11/08/2017
Solution File No.:	00962	Cert. No.: 00027
Sequential File No.:	00962	Lin. Date: 11/08/2017
		Lin. No.: 00027
		Soln. Date: 01/04/2018
		Soln. No.: 00205
		File Date: 01/04/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUM S3-0222
Control Solution %:	0.100%		Expires: 03/20/2019
Solution Control Lot:	17110		Bottle No.: 1463

Function	Rcsult %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:12S		
Control 1 EC	0.101%	21:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:14S		
Control 2 EC	0.100%	21:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15S		
Control 3 EC	0.100%	21:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CZEPIEL

First Name: NICHOLAS

MI: B

Signature: _____



Badge No.: 4041

Date: 01/04/2018