

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00670	Calib. Date:	11/26/2019
Certification File No.:	00671	Cert. Date:	11/26/2019
Linearity File No.:	00672	Lin. Date:	11/26/2019
Solution File No.:	00676	Soln. Date:	12/10/2019
Sequential File No.:	00676	File Date:	12/10/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0518

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:20S		
Control 1 EC	0.097%	04:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21S		
Control 2 EC	0.096%	04:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	04:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23S		
Control 3 EC	0.096%	04:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	04:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 12/10/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00651	Calib. Date: 06/07/2019
Certification File No.:	00652	Calib. No.: 00038
Linearity File No.:	00653	Cert. Date: 06/07/2019
Solution File No.:	00669	Cert. No.: 00031
Sequential File No.:	00669	Lin. Date: 06/07/2019
		Lin. No.: 00031
		Soln. Date: 11/20/2019
		Soln. No.: 00210
		File Date: 11/20/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0514

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:32S		
Control 1 EC	0.100%	22:33S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:33S		
Control 2 EC	0.097%	22:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35S		
Control 3 EC	0.098%	22:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 11/20/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00651	Calib. Date: 06/07/2019
Certification File No.:	00652	Calib. No.: 00038
Linearity File No.:	00653	Cert. Date: 06/07/2019
Solution File No.:	00667	Cert. No.: 00031
Sequential File No.:	00667	Lin. Date: 06/07/2019
		Lin. No.: 00031
		Soln. Date: 10/23/2019
		Soln. No.: 00209
		File Date: 10/23/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0475

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:32D		
Control 1 EC	0.100%	22:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:33D		
Control 2 EC	0.098%	22:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35D		
Control 3 EC	0.098%	22:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 10/23/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00651	Calib. Date: 06/07/2019
Certification File No.:	00652	Cert. No.: 00038
Linearity File No.:	00653	Cert. Date: 06/07/2019
Solution File No.:	00664	Lin. No.: 00031
Sequential File No.:	00664	Lin. Date: 06/07/2019
		Soln. No.: 00208
		Soln. Date: 09/21/2019
		File Date: 09/21/2019

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %: 0.100%		Expires: 06/10/2021
Solution Control Lot: 19150		Bottle No.: 0645

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:13D		
Control 1 EC	0.098%	22:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:14D		
Control 2 EC	0.097%	22:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:16D		
Control 3 EC	0.097%	22:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 09/21/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00651	Calib. Date: 06/07/2019
Certification File No.:	00652	Calib. No.: 00038
Linearity File No.:	00653	Cert. Date: 06/07/2019
Solution File No.:	00663	Cert. No.: 00031
Sequential File No.:	00663	Lin. Date: 06/07/2019
		Lin. No.: 00031
		Soln. Date: 09/21/2019
		Soln. No.: 00207
		File Date: 09/21/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0464

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:59D		
Control 1 EC	-.---%	20:00D	21.0°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	20:00D	21.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 09/21/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00651	Calib. Date: 06/07/2019
Certification File No.:	00652	Calib. No.: 00038
Linearity File No.:	00653	Cert. Date: 06/07/2019
Solution File No.:	00660	Cert. No.: 00031
Sequential File No.:	00660	Lin. Date: 06/07/2019
		Lin. No.: 00031
		Soln. Date: 08/22/2019
		Soln. No.: 00206
		File Date: 08/22/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1111

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:28D		
Control 1 EC	0.099%	11:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		
Control 2 EC	0.097%	11:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30D		
Control 3 EC	0.097%	11:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NAPOLITANO

First Name: DAVID

MI: M.

Signature:  7237

Badge No.: 7237

Date: 08/22/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00651	Calib. Date:	06/07/2019
Certification File No.:	00652	Cert. Date:	06/07/2019
Linearity File No.:	00653	Lin. Date:	06/07/2019
Solution File No.:	00659	Soln. Date:	08/15/2019
Sequential File No.:	00659	File Date:	08/15/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1091

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:34D		
Control 1 EC	0.000%	23:34D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.000%	23:34D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:35D		
Control 2 EC	-.---%	23:35D	33.9°C	CTRL GAS SUPPLY
Control 2 IR	-.---%	23:35D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:---		
Control 3 EC	-.---%	--:---	N/A	
Control 3 IR	-.---%	--:---	N/A	
Ambient Air Blank	-.---%	--:---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 08/15/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00651	Calib. No.:	00038
Certification File No.:	00652	Cert. Date:	06/07/2019
Linearity File No.:	00653	Lin. Date:	06/07/2019
Solution File No.:	00656	Soln. Date:	07/09/2019
Sequential File No.:	00656	File Date:	07/09/2019
Calibrating Unit: WET		Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1008

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:18D		
Control 1 EC	0.099%	12:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		
Control 2 EC	0.098%	12:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		
Control 3 EC	0.098%	12:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Temperature Probe Serial Number: DDUN P2-293 DW

Changed By:

Last Name: NAPOLITANO

First Name: DAVID

MI: M.

Signature:  7237

Badge No.: 7237

Date: 07/09/2019

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- X-Cal 2000 (Alcosim)
- Other: _____

Serial Number:

DDWES3-0189

Certification Date:

6-10-19

Technician:

BS

Re-Certification Due Date:

6-10-20

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications.
For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDUNP2-293

Certification Date:

6-10-19

Next Certification Due:

6-10-20

Probe Value:

102

Draeger, Inc.

BS



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

CHRISTOPHER S. PORRINO
Attorney General

COLONEL PATRICK J. CALLAHAN
Acting Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 11/01/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17360

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1197 to 0.1228 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is October 17, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature of Ali M. Alaoui]

Ali M. Alaoui, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 2nd day of November, 2017.

[Handwritten signature of Mary E. McLaughlin]
Notary

MARY ELIZABETH MCLAUGHLIN

ID # 2052190
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00651	Calib. Date: 06/07/2019
Certification File No.:	00652	Cert. No.: 00038
Linearity File No.:	00653	Cert. Date: 06/07/2019
Solution File No.:	00655	Lin. No.: 00031
Sequential File No.:	00655	Lin. Date: 06/07/2019
		Soln. No.: 00203
		Soln. Date: 07/04/2019
		File Date: 07/04/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1095

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:21D		
Control 1 EC	0.100%	07:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23D		
Control 2 EC	0.099%	07:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24D		
Control 3 EC	0.099%	07:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____

Badge No.: 16142

Date: 07/04/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00632	Calib. Date:	01/07/2019
Certification File No.:	00633	Cert. No.:	00037
Linearity File No.:	00634	Cert. Date:	01/07/2019
Solution File No.:	00650	Lin. No.:	00030
Sequential File No.:	00650	Lin. Date:	01/07/2019
		Soln. No.:	00201
		File Date:	05/31/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1252

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:19D		
Control 1 EC	0.098%	02:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:20D		
Control 2 EC	0.096%	02:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	02:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:22D		
Control 3 EC	0.097%	02:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 05/31/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00632	Calib. Date: 01/07/2019
Certification File No.:	00633	Calib. No.: 00037
Linearity File No.:	00634	Cert. Date: 01/07/2019
Solution File No.:	00646	Cert. No.: 00030
Sequential File No.:	00646	Lin. Date: 01/07/2019
		Lin. No.: 00030
		Soln. Date: 05/03/2019
		Soln. No.: 00200
		File Date: 05/03/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1007

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:38D		
Control 1 EC	0.099%	20:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	20:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39D		
Control 2 EC	0.096%	20:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40D		
Control 3 EC	0.097%	20:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REYNOLDS

First Name: TIMOTHY

MI: G

Signature: _____



Badge No.: 16142

Date: 05/03/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00632	Calib. Date: 01/07/2019
Certification File No.:	00633	Calib. No.: 00037
Linearity File No.:	00634	Cert. Date: 01/07/2019
Solution File No.:	00644	Cert. No.: 00030
Sequential File No.:	00644	Lin. Date: 01/07/2019
		Lin. No.: 00030
		Soln. Date: 04/01/2019
		Soln. No.: 00199
		File Date: 04/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1105

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:06S		
Control 1 EC	0.098%	16:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08S		
Control 2 EC	0.098%	16:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:09S		
Control 3 EC	0.098%	16:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 04/01/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00632	Calib. Date:	01/07/2019
Certification File No.:	00633	Cert. Date:	01/07/2019
Linearity File No.:	00634	Lin. Date:	01/07/2019
Solution File No.:	00642	Soln. Date:	02/28/2019
Sequential File No.:	00642	File Date:	02/28/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1254

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:16S		
Control 1 EC	0.100%	15:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18S		
Control 2 EC	0.098%	15:18S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19S		
Control 3 EC	0.098%	15:20S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDHHP2-094 BMG

Changed By:

Last Name: GAMBONE

First Name: BRIAN

MI: M

Signature: *T.P. I. B...* # 7029

Badge No.: 7029

Date: 02/28/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00632	Calib. Date: 01/07/2019
Certification File No.:	00633	Calib. No.: 00037
Linearity File No.:	00634	Cert. Date: 01/07/2019
Solution File No.:	00641	Cert. No.: 00030
Sequential File No.:	00641	Lin. Date: 01/07/2019
		Lin. No.: 00030
		Soln. Date: 02/24/2019
		Soln. No.: 00197
		File Date: 02/24/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1104

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:02S		
Control 1 EC	0.097%	21:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04S		
Control 2 EC	0.096%	21:04S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05S		
Control 3 EC	0.097%	21:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KNECHT

First Name: JENNIFER

MI: A

Signature: _____

Sgt Knecht 107

Badge No.: 16187

Date: 02/24/2019

