

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00705	Calib. Date:	11/13/2020
Certification File No.:	00706	Cert. Date:	11/13/2020
Linearity File No.:	00707	Lin. Date:	11/13/2020
Solution File No.:	00710	Soln. Date:	12/11/2020
Sequential File No.:	00710	File Date:	12/11/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0060

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:31S		
Control 1 EC	0.102%	06:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:33S		
Control 2 EC	0.100%	06:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:34S		
Control 3 EC	0.100%	06:35S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 12/11/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE	Calib. No.:	00040
Calibration File No.:	00689	Cert. Date:	05/21/2020
Certification File No.:	00690	Lin. Date:	05/21/2020
Linearity File No.:	00691	Soln. Date:	11/09/2020
Solution File No.:	00704	File Date:	11/09/2020
Sequential File No.:	00704		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0058

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:23S		
Control 1 EC	0.100%	04:24S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:25S		
Control 2 EC	0.098%	04:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:26S		
Control 3 EC	0.098%	04:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 11/09/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00689	Calib. Date:	05/21/2020
Certification File No.:	00690	Cert. Date:	05/21/2020
Linearity File No.:	00691	Lin. Date:	05/21/2020
Solution File No.:	00699	Soln. Date:	09/14/2020
Sequential File No.:	00699	File Date:	09/14/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0178

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:07D		
Control 1 EC	0.099%	00:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08D		
Control 2 EC	0.097%	00:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:10D		
Control 3 EC	0.098%	00:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 09/14/2020

Calibrating Unit New Standard Solution Report

Equipment

Location:	Alcotest 7110 MKIII-C WILLINGBORO TWP POLICE	Serial No.: ARWC-0060
Calibration File No.:	00689	Calib. Date: 05/21/2020
Certification File No.:	00690	Calib. No.: 00040
Linearity File No.:	00691	Cert. Date: 05/21/2020
Solution File No.:	00698	Lin. Date: 05/21/2020
Sequential File No.:	00698	Lin. No.: 00033
		Soln. No.: 00222
		Soln. Date: 08/16/2020
		File Date: 08/16/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0189
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0346

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:24D		
Control 1 EC	0.100%	04:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:26D		
Control 2 EC	0.099%	04:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:27D		
Control 3 EC	0.099%	04:28D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 08/16/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00689	Calib. Date:	05/21/2020
Certification File No.:	00690	Cert. No.:	00040
Linearity File No.:	00691	Cert. Date:	05/21/2020
Solution File No.:	00695	Lin. No.:	00033
Sequential File No.:	00695	Soln. Date:	07/17/2020
		Soln. No.:	00221
		File Date:	07/17/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0235

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:44D		
Control 1 EC	0.100%	21:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:46D		
Control 2 EC	0.099%	21:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:47D		
Control 3 EC	0.099%	21:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 07/17/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00689	Calib. Date:	05/21/2020
Certification File No.:	00690	Cert. Date:	05/21/2020
Linearity File No.:	00691	Lin. Date:	05/21/2020
Solution File No.:	00693	Soln. Date:	06/20/2020
Sequential File No.:	00693	File Date:	06/20/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0238

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:03D		
Control 1 EC	0.100%	02:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:04D		
Control 2 EC	0.100%	02:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:06D		
Control 3 EC	0.100%	02:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 06/20/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00670	Calib. Date:	11/26/2019
Certification File No.:	00671	Cert. Date:	11/26/2019
Linearity File No.:	00672	Lin. Date:	11/26/2019
Solution File No.:	00684	Soln. Date:	03/10/2020
Sequential File No.:	00684	File Date:	03/10/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0516

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:37S		
Control 1 EC	0.099%	04:38S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:38S		
Control 2 EC	0.096%	04:39S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:40S		
Control 3 EC	0.096%	04:41S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 03/10/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE	Calib. No.:	00039
Calibration File No.:	00670	Cert. Date:	11/26/2019
Certification File No.:	00671	Lin. Date:	11/26/2019
Linearity File No.:	00672	Soln. Date:	02/01/2020
Solution File No.:	00681	File Date:	02/01/2020
Sequential File No.:	00681		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0521

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:59S		
Control 1 EC	0.099%	01:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:01S		
Control 2 EC	0.097%	01:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:02S		
Control 3 EC	0.096%	01:03S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	01:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____

Sgt Jennifer Mitnick

Badge No.: 16187

Date: 02/01/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00670	Calib. Date: 11/26/2019
Certification File No.:	00671	Calib. No.: 00039
Linearity File No.:	00672	Cert. Date: 11/26/2019
Solution File No.:	00679	Cert. No.: 00032
Sequential File No.:	00679	Lin. Date: 11/26/2019
		Lin. No.: 00032
		Soln. Date: 01/08/2020
		Soln. No.: 00213
		File Date: 01/08/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0476

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:25S		
Control 1 EC	0.099%	21:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:26S		
Control 2 EC	0.097%	21:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:28S		
Control 3 EC	0.097%	21:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 01/08/2020