

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00722	Calib. No.: 00042
Certification File No.:	00723	Cert. No.: 00035
Linearity File No.:	00724	Lin. No.: 00035
Solution File No.:	00738	Soln. No.: 00239
Sequential File No.:	00738	
	Calib. Date: 04/26/2021	
	Cert. Date: 04/26/2021	
	Lin. Date: 04/26/2021	
	Soln. Date: 10/13/2021	
	File Date: 10/13/2021	
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0519

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:06D		
Control 1 EC	0.101%	23:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07D		
Control 2 EC	0.099%	23:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:09D		
Control 3 EC	0.099%	23:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 10/13/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00722	Calib. Date: 04/26/2021
Certification File No.:	00723	Calib. No.: 00042
Linearity File No.:	00724	Cert. Date: 04/26/2021
Solution File No.:	00737	Cert. No.: 00035
Sequential File No.:	00737	Lin. Date: 04/26/2021
		Lin. No.: 00035
		Soln. Date: 09/17/2021
		Soln. No.: 00238
		File Date: 09/17/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	21150	Expires: 05/04/2023
		Bottle No.: 0976

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:38D		
Control 1 EC	0.100%	00:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:39D		
Control 2 EC	0.098%	00:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:41D		
Control 3 EC	0.098%	00:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 09/17/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00722	Calib. Date: 04/26/2021
Certification File No.:	00723	Calib. No.: 00042
Linearity File No.:	00724	Cert. Date: 04/26/2021
Solution File No.:	00736	Cert. No.: 00035
Sequential File No.:	00736	Lin. Date: 04/26/2021
		Lin. No.: 00035
		Soln. Date: 08/20/2021
		Soln. No.: 00237
		File Date: 08/20/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0408

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:17D		
Control 1 EC	0.100%	21:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:19D		
Control 2 EC	0.098%	21:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:20D		
Control 3 EC	0.098%	21:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 08/20/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00722	Calib. Date:	04/26/2021
Certification File No.:	00723	Cert. No.:	00042
		Cert. Date:	04/26/2021
		Cert. No.:	00035
Linearity File No.:	00724	Lin. Date:	04/26/2021
Solution File No.:	00733	Lin. No.:	00035
Sequential File No.:	00733	Soln. Date:	07/23/2021
		Soln. No.:	00236
		File Date:	07/23/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	0980

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:42D		
Control 1 EC	0.097%	04:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:44D		
Control 2 EC	0.097%	04:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:45D		
Control 3 EC	0.097%	04:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 07/23/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00722	Calib. No.: 00042
Certification File No.:	00723	Cert. No.: 00035
Linearity File No.:	00724	Lin. No.: 00035
Solution File No.:	00728	Soln. No.: 00235
Sequential File No.:	00728	File Date: 06/25/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0179

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:20D		
Control 1 EC	0.102%	00:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:22D		
Control 2 EC	0.099%	00:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:23D		
Control 3 EC	0.099%	00:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 06/25/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00722	Calib. Date:	04/26/2021
Certification File No.:	00723	Cert. No.:	00042
Linearity File No.:	00724	Cert. Date:	04/26/2021
Solution File No.:	00727	Lin. Date:	04/26/2021
Sequential File No.:	00727	Lin. No.:	00035
		Soln. Date:	05/27/2021
		Soln. No.:	00234
		File Date:	05/27/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0120

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:36D		
Control 1 EC	0.099%	01:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:37D		
Control 2 EC	0.098%	01:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:39D		
Control 3 EC	0.098%	01:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 05/27/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00705	Calib. Date:	11/13/2020
Certification File No.:	00706	Cert. Date:	11/13/2020
Linearity File No.:	00707	Lin. Date:	11/13/2020
Solution File No.:	00721	Soln. Date:	04/20/2021
Sequential File No.:	00721	File Date:	04/20/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0459

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:56D		
Control 1 EC	0.101%	03:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	03:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58D		
Control 2 EC	0.100%	03:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:59D		
Control 3 EC	0.100%	04:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	04:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____

Badge No.: 16187

Date: 04/20/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE	Calib. No.:	00041
Calibration File No.:	00705	Cert. Date:	11/13/2020
Certification File No.:	00706	Lin. Date:	11/13/2020
Linearity File No.:	00707	Soln. Date:	03/18/2021
Solution File No.:	00719	File Date:	03/18/2021
Sequential File No.:	00719		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0290

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:45S	09:45D DN	
Control 1 EC	0.103%	08:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		
Control 2 EC	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 3 EC	0.101%	08:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S	09:50D DN	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Temperature Probe Serial Number: DDHH P2-094

Changed By:

Last Name: NAPOLITANO

First Name: DAVID

MI: M.

Signature:  7237

Badge No.: 7237

Date: 03/18/2021

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Dräger, Inc.

Model: ALCOTEST CU34

Model: MARK IIA

X-Cal 2000 (Alcosim)

Other: _____

Serial Number:

DDWES3-0188

Certification Date:

11.25.20

Technician:

am

Re-Certification Due Date:

11.25.21

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDHHP2-094

Certification Date:

11.25.20

Next Certification Due:

11.25.21

Probe Value:

105

Dräger, Inc.

am



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

SHEILA Y. OLIVER
Lt. Governor

GURBIR S. GREWAL
Attorney General

PATRICK J. CALLAHAN
Colonel

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 03/03/2020

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 20090

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1200 to 0.1218 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 19, 2022.

As Assistant Chief Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Assistant Chief Forensic Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 3 day of March, 2020.
Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 60110522
My Commission Expires 8/13/2024



An Internationally Accredited Agency

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger, Inc.

Model: ALCOTEST CU34

Model: MARK IIA

X-Cal 2000 (Alcosim)

Other: _____

Serial Number:

DDWES3-0188

Certification Date:

11.25.20

Technician:

am

Re-Certification Due Date:

11.25.21



Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe: Certification Date: Next Certification Due:

DDHHP2-094 11.25.20 11.25.21

Probe Value:

105

Draeger, Inc.

AM

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00705	Calib. Date:	11/13/2020
Certification File No.:	00706	Cert. Date:	11/13/2020
Linearity File No.:	00707	Lin. Date:	11/13/2020
Solution File No.:	00716	Soln. Date:	03/05/2021
Sequential File No.:	00716	File Date:	03/05/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0189
Control Solution %:	0.100%			Expires:	02/19/2022
Solution Control Lot:	20090			Bottle No.:	0232

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:10S		
Control 1 EC	0.102%	00:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	00:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11S		
Control 2 EC	0.101%	00:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:13S		
Control 3 EC	0.100%	00:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____

Sgt [Signature] 187

Badge No.: 16187

Date: 03/05/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE	Calib. No.:	00041
Calibration File No.:	00705	Calib. Date:	11/13/2020
Certification File No.:	00706	Cert. Date:	11/13/2020
Linearity File No.:	00707	Lin. Date:	11/13/2020
Solution File No.:	00714	Soln. Date:	02/05/2021
Sequential File No.:	00714	File Date:	02/05/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0520

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:29S		
Control 1 EC	0.104%	02:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:30S		
Control 2 EC	0.102%	02:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:32S		
Control 3 EC	0.101%	02:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 02/05/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00705	Calib. Date:	11/13/2020
Certification File No.:	00706	Cert. Date:	11/13/2020
Linearity File No.:	00707	Lin. Date:	11/13/2020
Solution File No.:	00713	Soln. Date:	01/07/2021
Sequential File No.:	00713	File Date:	01/07/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0189
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0236

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:22S		
Control 1 EC	0.101%	21:23S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:24S		
Control 2 EC	0.101%	21:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:25S		
Control 3 EC	0.100%	21:26S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____



Badge No.: 16187

Date: 01/07/2021