

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00816	Calib. Date:	03/22/2024
Certification File No.:	00817	Cert. Date:	03/22/2024
Linearity File No.:	00818	Lin. Date:	03/22/2024
Solution File No.:	00824	Soln. Date:	07/10/2024
Sequential File No.:	00824	File Date:	07/10/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1318

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:59D		
Control 1 EC	0.100%	19:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00D		
Control 2 EC	0.098%	20:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		
Control 3 EC	0.098%	20:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ABADIA

First Name: JOSEPH

MI: D

Signature:  #238

Badge No.: 16-238

Date: 07/10/2024

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00816	Calib. Date:	03/22/2024
Certification File No.:	00817	Cert. Date:	03/22/2024
Linearity File No.:	00818	Lin. Date:	03/22/2024
Solution File No.:	00823	Soln. Date:	06/15/2024
Sequential File No.:	00823	File Date:	06/15/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1296

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:51D		
Control 1 EC	0.101%	22:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:52D		
Control 2 EC	0.098%	22:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:54D		
Control 3 EC	0.098%	22:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ABADIA

First Name: JOSEPH

MI: D

Signature: _____

Handwritten Signature #238

Badge No.: 16-238

Date: 06/15/2024

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00816	Calib. Date: 03/22/2024
Certification File No.:	00817	Cert. No.: 00050
Linearity File No.:	00818	Cert. Date: 03/22/2024
Solution File No.:	00822	Lin. Date: 03/22/2024
Sequential File No.:	00822	Lin. No.: 00043
		Soln. Date: 05/24/2024
		Soln. No.: 00277
		File Date: 05/24/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0188
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 1282

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:12D		
Control 1 EC	0.102%	03:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:13D		
Control 2 EC	0.100%	03:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15D		
Control 3 EC	0.099%	03:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: 

Badge No.: 16187

Date: 05/24/2024

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00816	Calib. Date:	03/22/2024
Certification File No.:	00817	Cert. Date:	03/22/2024
Linearity File No.:	00818	Lin. Date:	03/22/2024
Solution File No.:	00820	Soln. Date:	04/18/2024
Sequential File No.:	00820	File Date:	04/18/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1306

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:52D		
Control 1 EC	0.102%	20:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:53D		
Control 2 EC	0.101%	20:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		
Control 3 EC	0.100%	20:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ABADIA

First Name: JOSEPH

MI: D

Signature: _____

PTM *AB* #238

Badge No.: 16-238

Date: 04/18/2024

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00804	Calib. Date:	11/22/2023
Certification File No.:	00805	Cert. Date:	11/22/2023
Linearity File No.:	00806	Lin. Date:	11/22/2023
Solution File No.:	00812	Soln. Date:	03/09/2024
Sequential File No.:	00812	File Date:	03/09/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1367

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:42S		
Control 1 EC	0.104%	02:43S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44S		
Control 2 EC	0.101%	02:44S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	02:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45S		
Control 3 EC	0.101%	02:46S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MITNICK

First Name: JENNIFER

MI: A

Signature: _____

Sgt JM 187

Badge No.: 16187

Date: 03/09/2024

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00804	Calib. Date:	11/22/2023
Certification File No.:	00805	Cert. Date:	11/22/2023
Linearity File No.:	00806	Lin. Date:	11/22/2023
Solution File No.:	00811	Soln. Date:	01/31/2024
Sequential File No.:	00811	File Date:	01/31/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1291

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:52S		
Control 1 EC	0.104%	19:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	19:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54S		
Control 2 EC	0.103%	19:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	19:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55S		
Control 3 EC	0.103%	19:56S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ABADIA

First Name: JOSEPH

MI: D

Signature: _____

Badge No.: 16238

Date: 01/31/2024

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00804 Calib. Date: 11/22/2023 Calib. No.: 00048
Certification File No.: 00805 Cert. Date: 11/22/2023 Cert. No.: 00041
Linearity File No.: 00806 Lin. Date: 11/22/2023 Lin. No.: 00041
Solution File No.: 00810 Soln. Date: 01/04/2024 Soln. No.: 00272
Sequential File No.: 00810 File Date: 01/04/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 01/11/2025
Solution Control Lot: 23010 Bottle No.: 1372

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:56S		
Control 1 EC	0.103%	19:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.104%	19:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:58S		
Control 2 EC	0.100%	19:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	19:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59S		
Control 3 EC	0.101%	20:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ABADIA

First Name: JOSEPH

MI: D

Signature:  #238

Badge No.: 16238

Date: 01/04/2024