

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: STRATFORD POLICE DEPT.  
Serial No.: ARNK-0039  
Calibration File No.: 00539  
Calib. Date: 09/15/2008  
Calib. No.: 00009  
Certification File No.: 00540  
Cert. Date: 09/15/2008  
Cert. No.: 00007  
Linearity File No.: 00541  
Lin. Date: 09/15/2008  
Lin. No.: 00006  
Solution File No.: 00597  
Soln. Date: 12/31/2008  
Soln. No.: 00069  
Sequential File No.: 00597  
File Date: 12/31/2008

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 08F053  
Model No.: CU-34  
Serial No.: DDUL S3-0163  
Expires: 06/09/2010  
Bottle No.: 0272

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:42S		
Control 1 EC	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		
Control 2 EC	0.099%	11:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S		
Control 3 EC	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  63200

Badge No.: 3200

Date: 12/31/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00539	Calib. Date: 09/15/2008
Certification File No.:	00540	Calib. No.: 00009
Linearity File No.:	00541	Cert. Date: 09/15/2008
Solution File No.:	00579	Cert. No.: 00007
Sequential File No.:	00579	Lin. Date: 09/15/2008
		Lin. No.: 00006
		Soln. Date: 12/06/2008
		Soln. No.: 00068
		File Date: 12/06/2008

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %:	0.100%		Expires: 06/09/2010
Solution Control Lot:	08F053		Bottle No.: 0274

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:30S		
Control 1 EC	0.103%	12:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32S		
Control 2 EC	0.101%	12:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:33S		
Control 3 EC	0.103%	12:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	12:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 12/06/2008

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.		
Calibration File No.: 00539	Calib. Date: 09/15/2008	Calib. No.: 00009
Certification File No.: 00540	Cert. Date: 09/15/2008	Cert. No.: 00007
Linearity File No.: 00541	Lin. Date: 09/15/2008	Lin. No.: 00006
Solution File No.: 00566	Soln. Date: 11/16/2008	Soln. No.: 00067
Sequential File No.: 00566	File Date: 11/16/2008	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 04/24/2009
Solution Control Lot: 07B046		Bottle No.: 0688

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:03S		
Control 1 EC	0.100%	14:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04S		
Control 2 EC	0.100%	14:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05S		
Control 3 EC	0.100%	14:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAGLIARDI

First Name: FRANK

MI: J

Signature: *Frank J. Gagliardi #3203*

Badge No.: 3203

Date: 11/16/2008

# Calibrating Unit

## New Standard Solution Report

### Equipment

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.		
Calibration File No.: 00539	Calib. Date: 09/15/2008	Calib. No.: 00009
Certification File No.: 00540	Cert. Date: 09/15/2008	Cert. No.: 00007
Linearity File No.: 00541	Lin. Date: 09/15/2008	Lin. No.: 00006
Solution File No.: 00558	Soln. Date: 11/01/2008	Soln. No.: 00066
Sequential File No.: 00558	File Date: 11/01/2008	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 04/24/2009
Solution Control Lot: 07B046		Bottle No.: 0676

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:32S		
Control 1 EC	0.100%	13:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33S		
Control 2 EC	0.100%	13:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34S		
Control 3 EC	0.101%	13:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

*[Handwritten Signature]* 63204

Badge No.: 3200

Date: 11/01/2008

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.		
Calibration File No.: 00539	Calib. Date: 09/15/2008	Calib. No.: 00009
Certification File No.: 00540	Cert. Date: 09/15/2008	Cert. No.: 00007
Linearity File No.: 00541	Lin. Date: 09/15/2008	Lin. No.: 00006
Solution File No.: 00548	Soln. Date: 10/07/2008	Soln. No.: 00065
Sequential File No.: 00548	File Date: 10/07/2008	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 04/24/2009
Solution Control Lot: 07B046		Bottle No.: 0689

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:24D		
Control 1 EC	0.101%	15:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25D		
Control 2 EC	0.101%	15:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27D		
Control 3 EC	0.100%	15:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: -

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 10/07/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00539	Calib. Date:	09/15/2008
Certification File No.:	00540	Calib. No.:	00009
Linearity File No.:	00541	Cert. Date:	09/15/2008
Solution File No.:	00542	Cert. No.:	00007
Sequential File No.:	00542	Lin. Date:	09/15/2008
		Lin. No.:	00006
		Soln. Date:	09/15/2008
		Soln. No.:	00064
		File Date:	09/15/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	07B046	Expires:	04/24/2009
		Bottle No.:	0687

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.099%	11:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 2 EC	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		
Control 3 EC	0.099%	11:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*TEMPERATURE PROBE SERIAL NUMBER: DDUJPA-030 NE*

**Changed By:**

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: \_\_\_\_\_

*TPR. I WILLIAM CROSS #5342*

Badge No.: 5342

Date: 09/15/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00509	Calib. Date: 07/18/2008
Certification File No.:	00510	Calib. No.: 00008
Linearity File No.:	00511	Cert. Date: 07/18/2008
Solution File No.:	00537	Cert. No.: 00006
Sequential File No.:	00537	Lin. Date: 07/18/2008
		Lin. No.: 00005
		Soln. Date: 09/04/2008
		Soln. No.: 00063
		File Date: 09/04/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	07B046	Expires: 04/24/2009
		Bottle No.: 0690

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:44D		
Control 1 EC	0.100%	17:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:45D		
Control 2 EC	0.100%	17:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:46D		
Control 3 EC	0.099%	17:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:47D		

All tests within acceptable tolerancce.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  \_\_\_\_\_

Badge No.: 3200

Date: 09/04/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00509	Calib. Date:	07/18/2008
Certification File No.:	00510	Cert. Date:	07/18/2008
Linearity File No.:	00511	Lin. Date:	07/18/2008
Solution File No.:	00525	Soln. Date:	08/12/2008
Sequential File No.:	00525	File Date:	08/12/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	07B046	Expires:	04/24/2009
		Bottle No.:	0691

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:17D		
Control 1 EC	0.098%	17:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 2 EC	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		
Control 3 EC	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

63200

Badge No.: 3200

Date: 08/12/2008



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00415	Calib. Date:	11/28/2007
Certification File No.:	00416	Cert. No.:	00007
Linearity File No.:	00417	Cert. Date:	11/28/2007
Solution File No.:	00507	Lin. No.:	00004
Sequential File No.:	00507	Lin. Date:	11/28/2007
		Soln. No.:	00060
		Soln. Date:	07/16/2008
		File Date:	07/16/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	07B046	Expires:	04/24/2009
		Bottle No.:	0681

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:17D		
Control 1 EC	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		
Control 2 EC	0.099%	17:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		
Control 3 EC	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

*Chris Morello* #3700

Badge No.: 3200

Date: 07/16/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00415	Calib. Date:	11/28/2007
Certification File No.:	00416	Cert. Date:	11/28/2007
Linearity File No.:	00417	Lin. Date:	11/28/2007
Solution File No.:	00506	Soln. Date:	07/16/2008
Sequential File No.:	00506	File Date:	07/16/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	07B046	Expires:	04/24/2009
		Bottle No.:	0686

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:28D		
Control 1 EC	----%	15:29D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	----%	15:29D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MOREIQO	First Name: RONALD	MI: M
Signature: <u><i>Ronald Moreiqo</i></u>	43200	Badge No.: 3200
		Date: 07/16/2008