

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02155	Calib. Date: 09/01/2015
Certification File No.:	02156	Calib. No.: 00032
Linearity File No.:	02157	Cert. Date: 09/01/2015
Solution File No.:	02228	Cert. No.: 00023
Sequential File No.:	02228	Lin. Date: 09/01/2015
		Lin. No.: 00022
		Soln. Date: 12/29/2015
		Soln. No.: 00235
		File Date: 12/29/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 0170

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:07S		
Control 1 EC	0.103%	16:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	16:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08S		
Control 2 EC	0.103%	16:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	16:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:10S		
Control 3 EC	0.102%	16:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

*[Handwritten Signature]* 63200

Badge No.: 3200

Date: 12/29/2015

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 02155 Calib. Date: 09/01/2015 Calib. No.: 00032  
Certification File No.: 02156 Ceri. Date: 09/01/2015 Cert. No.: 00023  
Linearity File No.: 02157 Lin. Date: 09/01/2015 Lin. No.: 00022  
Solution File No.: 02210 Soln. Date: 12/06/2015 Soln. No.: 00234  
Sequential File No.: 02210 File Date: 12/06/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133  
Control Solution %: 0.100% Expires: 04/29/2016  
Solution Control Lot: 14D129 Bottle No.: 1195

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:47S		
Control 1 EC	0.103%	21:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.105%	21:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48S		
Control 2 EC	0.102%	21:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	21:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:50S		
Control 3 EC	0.102%	21:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	21:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  6320

Badge No.: 3200

Date: 12/06/2015

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02155	Calib. Date: 09/01/2015
Certification File No.:	02156	Calib. No.: 00032
Linearity File No.:	02157	Cert. Date: 09/01/2015
Solution File No.:	02196	Cert. No.: 00023
Sequential File No.:	02196	Lin. Date: 09/01/2015
		Lin. No.: 00022
		Soln. Date: 11/12/2015
		Soln. No.: 00233
		File Date: 11/12/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 1199

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:16S		
Control 1 EC	0.103%	13:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.104%	13:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17S		
Control 2 EC	0.102%	13:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	13:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:19S		
Control 3 EC	0.102%	13:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	13:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20S		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO	First Name: RONALD	MI: M
--------------------	--------------------	-------


  

Signature: 	Badge No.: 3200
	Date: 11/12/2015

Signature: 	Badge No.: 3200
	Date: 11/12/2015

Signature: 	Badge No.: 3200
	Date: 11/12/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alconex 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT	
Calibration File No.:	02155	Calib. Date: 09/01/2015
Certification File No.:	02156	Cert. Date: 09/01/2015
Linearity File No.:	02157	Lin. Date: 09/01/2015
Solution File No.:	02186	Soln. Date: 10/19/2015
Sequential File No.:	02186	File Date: 10/19/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 1198

Function	Result	Time	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:05D		
Control 1 EC	0.102%	10:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	10:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		
Control 2 EC	0.101%	10:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		
Control 3 EC	0.101%	10:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alconex 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 10/19/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02155	Calib. Date:	09/01/2015
Certification File No.:	02156	Cert. Date:	09/01/2015
Linearity File No.:	02157	Lin. Date:	09/01/2015
Solution File No.:	02174	Soln. Date:	09/25/2015
Sequential File No.:	02174	File Date:	09/25/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	1194

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:53D		
Control 1 EC	0.103%	09:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	09:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54D		
Control 2 EC	0.102%	09:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	09:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 3 EC	0.101%	09:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	09:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 09/25/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.	Calib. No.:	00031
Calibration File No.:	02065	Calib. Date:	04/02/2015
Certification File No.:	02066	Cert. Date:	04/02/2015
Linearity File No.:	02067	Lin. Date:	04/02/2015
Solution File No.:	02143	Soln. Date:	08/06/2015
Sequential File No.:	02143	File Date:	08/06/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK 83-0133
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	1178

Function	Result	Time	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:54D		
Control 1 EC	0.101%	10:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		
Control 2 EC	0.100%	10:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		
Control 3 EC	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200-

Date: 08/06/2015

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.	ARNK 0039
<b>Location:</b>	STRATFORD POLICE DEPT		
<b>Calibration File No.:</b>	02065	<b>Calib. Date:</b>	04/02/2015
<b>Certification File No.:</b>	02066	<b>Cert. No.:</b>	00022
<b>Linearity File No.:</b>	02067	<b>Lin. Date:</b>	04/02/2015
<b>Solution File No.:</b>	02130	<b>Soln. No.:</b>	00228
<b>Sequential File No.:</b>	02130	<b>File Date:</b>	07/13/2015

<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34	<b>Serial No.:</b>	DDUK 53-0133
<b>Control Solution %:</b>	0.100%	<b>Expires:</b>	04/29/2016		
<b>Solution Control Lot:</b>	14D129	<b>Bottle No.:</b>	1177		

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:50D		
Control 1 EC	0.100%	12:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		
Control 2 EC	0.100%	12:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:53D		
Control 3 EC	0.100%	12:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist

**Changed By:**

Last Name: GAGLIARDI

First Name: FRANK

MI: 4

Signature:

*Frank J. Gagliardi # 3703*

Badge No: 3203

Date: 07/13/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02065	Calib. Date:	04/02/2015
Certification File No.:	02066	Cert. No.:	00031
Linearity File No.:	02067	Lin. No.:	00022
Solution File No.:	02123	Soln. Date:	04/02/2015
Sequential File No.:	02123	File Date:	06/23/2015
			00021
			00227
<b>Calibrating Unit:</b>	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0153
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	1184

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:49D		
Control 1 EC	0.101%	11:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		
Control 2 EC	0.100%	11:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		
Control 3 EC	0.100%	11:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: Δ Δ LEP1-0015 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W

Signature: TKA I KA W. #6704

Badge No.: 6704

Date: 06/23/2015





**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model ALCOTEST® C134
- Model MARK II
- Other: \_\_\_\_\_

Serial Number:

DDUKS3-0133

Certification Date

5-1-15

Technician

BC

Re-Certification Due Date

5-1-16

# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

# CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDLPI-0015

Certification date:

4-30-15

Next Certification date:

4-30-16

Probe Value

108

Dräger Safety Diagnostics, Inc.  
Technical Services Department

BC

DEPARTMENT OF  
**Traffic and Public Safety**  
*What is to certify that*

**KEVIN W. ALCOTT**  
**NEW JERSEY STATE POLICE**

IS QUALIFIED AND COMPETENT TO CONDUCT BREATH TESTS AND IS FURNISHING TO CHAPTER 17 OF  
 THE LAWS OF 1984 THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
 A METHOD TO DETERMINE INTOXICATION.

ISSUED UNDER MY HAND AT TRENTON, NEW JERSEY, ON 18th DAY OF APRIL

TWO THOUSAND AND **EIGHT**

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
8-25-10	BCPA	A. J. W.
1-2-12	Beverly Woods	C. J.
1-16-14	SAVANVILLE PD	R. J. W.
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 28.90 (Rev. 07/01)

DEPARTMENT OF  
**Traffic and Public Safety**  
*What is to certify that*

**Kevin W. Alcott**  
**Breath Test Coordinator/Instructor**

IS QUALIFIED AND COMPETENT TO CONDUCT BREATH TESTS AND IS FURNISHING TO CHAPTER 17 OF  
 THE LAWS OF 1984 THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
 A METHOD TO DETERMINE INTOXICATION.

ISSUED UNDER MY HAND AT TRENTON, NEW JERSEY, ON 21st DAY OF May

TWO THOUSAND AND **Eleven**

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 28.90 (Rev. 08/10)

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Aleotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02065	Calib. Date: 04/02/2015
Certification File No.:	02066	Calib. No.: 00031
Linearity File No.:	02067	Cert. Date: 04/02/2015
Solution File No.:	02111	Cert. No.: 00022
Sequential File No.:	02111	Lin. Date: 04/02/2015
		Lin. No.: 00021
		Soln. Date: 06/12/2015
		Soln. No.: 00226
		File Date: 06/12/2015
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 1193

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:10D		
Control 1 EC	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		
Control 2 EC	0.101%	10:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 3 EC	0.100%	10:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Aleotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

Mi: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 06/12/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02065	Calib. Date:	04/02/2015
Certification File No.:	02066	Cert. Date:	04/02/2015
Linearity File No.:	02067	Lin. Date:	04/02/2015
Solution File No.:	02094	Soln. Date:	05/21/2015
Sequential File No.:	02094	File Date:	05/21/2015
Calibrating Unit:	WE1	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	1189

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:16D		
Control 1 EC	0.103%	15:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17D		
Control 2 EC	0.103%	15:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.104%	15:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 3 EC	0.103%	15:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	15:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 05/21/2015

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.	Calib. No.:	00031
Calibration File No.:	02065	Cert. Date:	04/02/2015
Certification File No.:	02066	Lin. Date:	04/02/2015
Linearity File No.:	02067	Soln. Date:	04/26/2015
Solution File No.:	02083	File Date:	04/26/2015
Sequential File No.:	02083		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDVIL S3-0163
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	1183

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:07D		
Control 1 EC	0.102%	10:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	10:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		
Control 2 EC	0.101%	10:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 3 EC	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MUM

Signature: 

Badge No. : 3200

Date: 04/26/2015

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
 Location: STRATFORD POLICE DEPT.  
 Calibration File No.: 01957 Calib. Date: 10/20/2014 Calib. No.: 00030  
 Certification File No.: 01958 Cert. Date: 10/20/2014 Cert. No.: 00021  
 Linearity File No.: 01959 Lin. Date: 10/20/2014 Lin. No.: 00020  
 Solution File No.: 02061 Soln. Date: 03/28/2015 Soln. No.: 00222  
 Sequential File No.: 02061 File Date: 03/28/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163  
 Control Solution %: 0.100% Expires: 04/29/2016  
 Solution Control Lot: 14D129 Bottle No.: 1182


Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:30S		
Control 1 EC	0.098%	01:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:31S		
Control 2 EC	0.097%	01:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:33S		
Control 3 EC	0.097%	01:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:34S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**


Last Name: KELLY First Name: ROBERT MI: J

Signature:  # 3204 Badge No.: 3204  
 Date: 03/28/2015

Signature:  # 3204 Badge No.: 3204  
 Date: 03/28/2015

Signature:  # 3204 Badge No.: 3204  
 Date: 03/28/2015

Signature:  # 3204 Badge No.: 3204  
 Date: 03/28/2015

Signature:  # 3204 Badge No.: 3204  
 Date: 03/28/2015

# Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	01957	Calib. Date: 10/20/2014
Certification File No.:	01958	Cert. No.: 00030
Linearity File No.:	01959	Cert. Date: 10/20/2014
Solution File No.:	02050	Lin. No.: 00021
Sequential File No.:	02050	Lin. Date: 10/20/2014
		Soln. No.: 00221
		Soln. Date: 03/17/2015
		File Date: 03/17/2015

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 04/29/2016
Solution Control Lot: 14D129		Bottle No.: 1176


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:22S		
Control 1 EC	0.100%	23:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:24S		
Control 2 EC	0.098%	23:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:25S		
Control 3 EC	0.099%	23:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:27S		


All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**


Last Name: MCBRIDE      First Name: STEPHEN      MJ: J

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
<b>Location:</b>	STRATFORD POLICE DEPT.		
<b>Calibration File No.:</b>	01957	<b>Calib. Date:</b>	10/20/2014
<b>Certification File No.:</b>	01958	<b>Cert. Date:</b>	10/20/2014
<b>Lancarity File No.:</b>	01959	<b>Lin. Date:</b>	10/20/2014
<b>Solution File No.:</b>	02040	<b>Soln. Date:</b>	02/21/2015
<b>Sequential File No.:</b>	02040	<b>File Date:</b>	02/21/2015

<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34	<b>Serial No.:</b>	DDUL 53-0163
<b>Control Solution %:</b>	0.100%			<b>Expires:</b>	04/29/2016
<b>Solution Control Lot:</b>	14D129			<b>Bottle No.:</b>	1192

Function	Result	Time	Temperature	Comment(s)
	%BAC	MM MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:29S		
Control 1 EC	0.100%	10:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30S		
Control 2 EC	0.098%	10:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32S		
Control 3 EC	0.100%	10:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MJ, M

Signature: 

Badge No.: 3200

Date: 02/21/2015







State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

KIM GUADAGINO  
 Lt. Governor

JOHN J. HOFFMAN  
 Acting Attorney General

COLONEL JOSEPH R. PUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.

ANALYSIS DATE: 5/06/2014

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 14D129

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1208 to 0.1217 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 29, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.  
 Research Scientist  
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 6<sup>th</sup> day of May, 2014.

Kim Guadagno

Governor

KIM GUADAGINO  
 Lt. Governor

DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

Acting Attorney General

COLONEL JOSEPH R. PUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but

# Packing Slip

Customer id    Order No    Order date  
 150073901    10714884    07/21/2014

Packing Slip No    Ship date  
 80875365    07/21/2014

Please refer back to Invoice

Ship to

150073901  
 STRATFORD POLICE DEPARTMENT  
 ATM, SGT MCBRIDE  
 CAMDEN COUNTY  
 716 UNION AVENUE  
 STRATFORD, NJ 08084  
 USA

Ship to



Your Preferred Order Number

07/21/2014  
 072214

Any questions? Please contact  
 BRENDIA BROTHMAN

Customer

STRATFORD POLICE DEPARTMENT  
 CAMDEN COUNTY  
 716 UNION AVENUE  
 STRATFORD, NJ 08084  
 USA

Your sales person

Don Pouillot  
 Phone: 973-398-3228

Delivery terms

FA  
 FREIGHT ALLOWED

Page 1 / 2

Item#	QTY	Description
		Ship via: Fed Ex Ground
		NOTE: All claims for shortages/defects must be reported within 10 Business Days After receipt of order. All returns must have our stock clearly marked on the box. Please call Customer Service at 866-385-5909
6010	1 EA	4540813 JAR. SIMULATOR Ordered / Back ordered : 1 / 0 EA
0020	20 EA	4805700 MULTIPROB. ALCO 7510/1420r9639 PRO2 Ordered / Back ordered : 20 / 0 EA

Orsinger Safety Diagnostics Inc.  
 Accounting Address  
 507 Technology Drive  
 Pittsburgh, Pa 15225  
 Tel: (412) 788-5327  
 Fax: (412) 788-5366

Remit Wire Transfers To:  
 Citizens Bank  
 Acct # 4208426615  
 Acct Name: Orsinger Safety Diagnostics Inc  
 Transit Number: 026076150  
 Federal ID Number: 84-1600159

REMIT TO  
 Orsinger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150053903    10714984    07/21/2014

Packing slip no.    Ship date  
80875365    07/21/2014

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	25 EA	4407061 CERT SOLUTION, .10% NEW JERSEY CERTIFIED Ordered / Back ordered : 25 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	25 EA	4407061 CERT SOLUTION, .10% NEW JERSEY CERTIFIED Ordered / Back ordered : 25 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alister T140 MKIII-C	Serial No.	ARNE 0039
<b>Location</b>	STRATFORD POLICE DEPT		
Calibration File No.	08957	Calib. Start:	10/30/2014
Certification File No.	01908	Cal. Date:	10/29/2014
Laboratory File No.	01909	Lab. Date:	10/29/2014
Solution File No.	00018	Soln. Date:	01/10/2015
Sequential File No.	02018	File Date:	01/10/2015
Calibrating Unit	WEST	Model No.	CU 34
Current Solution No.	01008	Serial No.	DDLL 51-0161
Solution Control Lot	140029	Expires:	04/29/2016
		Batch No.	1101

Function	Result	Tare	Temperature	Comments (or Error)
	0.0000	11.535		
Ambient Air Blank	0.0000	11.535	34.0°C	*** TEST PASSED ***
Control 1 BC	0.0000	11.535	34.0°C	*** TEST PASSED ***
Control 1 BR	0.0000	11.535	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.0000	11.565		
Control 2 BC	0.0000	11.570	34.0°C	*** TEST PASSED ***
Control 2 BR	0.0000	11.570	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.0000	11.585		
Control 3 BC	0.0000	11.585	34.0°C	*** TEST PASSED ***
Control 3 BR	0.0000	11.585	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.0000	11.585		

All tests within acceptable tolerance.


On this date, I declared the above information "TRUE & ACCURATE" in accordance with Alister T140 operation training and procedures established by the (NIST) Chief Forensic Scientist.

**Changed By:**

Last Name: MORRILLO

First Name: RONALD

ML M

Signature: 

Badge No.: 1208

Date: 01/10/2015