

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02463	Calib. Date:	07/19/2017
Certification File No.:	02464	Cert. Date:	07/19/2017
Linearity File No.:	02465	Lin. Date:	07/19/2017
Solution File No.:	02514	Soln. Date:	12/02/2017
Sequential File No.:	02514	File Date:	12/02/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0133
Control Solution %:	0.100%			Expires:	06/07/2018
Solution Control Lot:	16140			Bottle No.:	0936

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:25S		
Control 1 EC	0.099%	16:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26S		
Control 2 EC	0.098%	16:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28S		
Control 3 EC	0.098%	16:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____

[Handwritten Signature]
63200

Badge No.: 3200

Date: 12/02/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02463 Calib. Date: 07/19/2017 Calib. No.: 00037
Certification File No.: 02464 Cert. Date: 07/19/2017 Cert. No.: 00027
Linearity File No.: 02465 Lin. Date: 07/19/2017 Lin. No.: 00026
Solution File No.: 02506 Soln. Date: 11/07/2017 Soln. No.: 00275
Sequential File No.: 02506 File Date: 11/07/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 1058

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:39S		
Control 1 EC	0.101%	12:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41S		
Control 2 EC	0.099%	12:41S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42S		
Control 3 EC	0.099%	12:43S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KELLY

First Name: ROBERT

MI: J

Signature: _____



Badge No.: 3201

Date: 11/07/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02463	Calib. Date:	07/19/2017
Certification File No.:	02464	Cert. Date:	07/19/2017
Linearity File No.:	02465	Lin. Date:	07/19/2017
Solution File No.:	02497	Soln. Date:	10/11/2017
Sequential File No.:	02497	File Date:	10/11/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0133
Control Solution %:	0.100%			Expires:	06/07/2018
Solution Control Lot:	16140			Bottle No.:	1174

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:07D		
Control 1 EC	0.099%	22:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:09D		
Control 2 EC	0.098%	22:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:10D		
Control 3 EC	0.098%	22:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAGLIARDI

First Name: FRANK

MI: J

Signature: _____

J. M. Gagliardi #3203

Badge No.: 3203

Date: 10/11/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02463	Calib. Date: 07/19/2017
Certification File No.:	02464	Calib. No.: 00037
Linearity File No.:	02465	Cert. Date: 07/19/2017
Solution File No.:	02489	Cert. No.: 00027
Sequential File No.:	02489	Lin. Date: 07/19/2017
		Lin. No.: 00026
		Soln. Date: 09/15/2017
		Soln. No.: 00273
		File Date: 09/15/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUK S3-0133
Control Solution %:	0.100%		Expires: 06/07/2018
Solution Control Lot:	16140		Bottle No.: 1271

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:01D		
Control 1 EC	0.099%	21:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02D		
Control 2 EC	0.098%	21:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04D		
Control 3 EC	0.099%	21:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____



Badge No.: 3200

Date: 09/15/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02463 Calib. Date: 07/19/2017 Calib. No.: 00037
Certification File No.: 02464 Cert. Date: 07/19/2017 Cert. No.: 00027
Linearity File No.: 02465 Lin. Date: 07/19/2017 Lin. No.: 00026
Solution File No.: 02479 Soln. Date: 08/19/2017 Soln. No.: 00272
Sequential File No.: 02479 File Date: 08/19/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 1487

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:24D		
Control 1 EC	0.099%	12:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		
Control 2 EC	0.098%	12:26D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		
Control 3 EC	0.099%	12:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 08/19/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02463	Calib. Date: 07/19/2017
Certification File No.:	02464	Calib. No.: 00037
Linearity File No.:	02465	Cert. Date: 07/19/2017
Solution File No.:	02478	Cert. No.: 00027
Sequential File No.:	02478	Lin. Date: 07/19/2017
		Lin. No.: 00026
		Soln. Date: 08/19/2017
		Soln. No.: 00271
		File Date: 08/19/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1155

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:09D		
Control 1 EC	-.---%	11:10D	N/A	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	11:10D	N/A	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____



Badge No.: 3200

Date: 08/19/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02463	Calib. Date: 07/19/2017
Certification File No.:	02464	Calib. No.: 00037
Linearity File No.:	02465	Cert. Date: 07/19/2017
Solution File No.:	02476	Cert. No.: 00027
Sequential File No.:	02476	Lin. Date: 07/19/2017
		Lin. No.: 00026
		Soln. Date: 08/15/2017
		Soln. No.: 00270
		File Date: 08/15/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1154

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:26D		
Control 1 EC	0.100%	10:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27D		
Control 2 EC	0.099%	10:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 3 EC	0.099%	10:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____

[Handwritten Signature] 6320

Badge No.: 3200

Date: 08/15/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02410	Calib. Date: 01/31/2017
Certification File No.:	02411	Calib. No.: 00036
Linearity File No.:	02412	Cert. Date: 01/31/2017
Solution File No.:	02461	Cert. No.: 00026
Sequential File No.:	02461	Lin. Date: 01/31/2017
		Lin. No.: 00025
		Soln. Date: 07/13/2017
		Soln. No.: 00268
		File Date: 07/13/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1152

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:54D		
Control 1 EC	0.099%	10:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		
Control 2 EC	0.099%	10:56D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		
Control 3 EC	0.098%	10:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____

[Handwritten Signature] 8320

Badge No.: 3200

Date: 07/13/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02410	Calib. Date: 01/31/2017
Certification File No.:	02411	Calib. No.: 00036
Linearity File No.:	02412	Cert. Date: 01/31/2017
Solution File No.:	02457	Cert. No.: 00026
Sequential File No.:	02457	Lin. Date: 01/31/2017
		Lin. No.: 00025
		Soln. Date: 06/19/2017
		Soln. No.: 00267
		File Date: 06/19/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1168

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:14D		
Control 1 EC	0.096%	03:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	03:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15D		
Control 2 EC	0.096%	03:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17D		
Control 3 EC	0.096%	03:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY

First Name: ROBERT

MI: J

Signature: _____



Badge No.: 3201

Date: 06/19/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02410	Calib. Date: 01/31/2017
Certification File No.:	02411	Calib. No.: 00036
Linearity File No.:	02412	Cert. Date: 01/31/2017
Solution File No.:	02445	Cert. No.: 00026
Sequential File No.:	02445	Lin. Date: 01/31/2017
		Lin. No.: 00025
		Soln. Date: 05/24/2017
		Soln. No.: 00266
		File Date: 05/24/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	16I40	Expires: 06/07/2018
		Bottle No.: 1449

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:26D		
Control 1 EC	0.098%	16:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27D		
Control 2 EC	0.098%	16:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28D		
Control 3 EC	0.098%	16:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____



Badge No.: 3200

Date: 05/24/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02410 Calib. Date: 01/31/2017 Calib. No.: 00036
Certification File No.: 02411 Cert. Date: 01/31/2017 Cert. No.: 00026
Linearity File No.: 02412 Lin. Date: 01/31/2017 Lin. No.: 00025
Solution File No.: 02439 Soln. Date: 04/28/2017 Soln. No.: 00265
Sequential File No.: 02439 File Date: 04/28/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 1446

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:55D		
Control 1 EC	0.099%	11:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 2 EC	0.099%	11:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		
Control 3 EC	0.099%	11:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KELLY

First Name: ROBERT

MI: J

Signature: _____



Badge No.: 3201

Date: 04/28/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02410	Calib. Date: 01/31/2017
Certification File No.:	02411	Calib. No.: 00036
Linearity File No.:	02412	Cert. Date: 01/31/2017
Solution File No.:	02435	Cert. No.: 00026
Sequential File No.:	02435	Lin. Date: 01/31/2017
		Lin. No.: 00025
		Soln. Date: 03/30/2017
		Soln. No.: 00264
		File Date: 03/30/2017

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 06/07/2018
Solution Control Lot: I6140		Bottle No.: 1456

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:10S		
Control 1 EC	0.099%	19:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:11S		
Control 2 EC	0.099%	19:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:13S		
Control 3 EC	0.099%	19:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____

[Handwritten Signature] 63200

Badge No.: 3200

Date: 03/30/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02410	Calib. Date: 01/31/2017
Certification File No.:	02411	Calib. No.: 00036
Linearity File No.:	02412	Cert. Date: 01/31/2017
Solution File No.:	02425	Cert. No.: 00026
Sequential File No.:	02425	Lin. Date: 01/31/2017
		Lin. No.: 00025
		Soln. Date: 03/04/2017
		Soln. No.: 00263
		File Date: 03/04/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 0941

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:24S		
Control 1 EC	0.097%	11:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25S		
Control 2 EC	0.097%	11:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27S		
Control 3 EC	0.097%	11:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____

[Handwritten Signature] *6320*

Badge No.: 3200

Date: 03/04/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02410 Calib. Date: 01/31/2017 Calib. No.: 00036
Certification File No.: 02411 Cert. Date: 01/31/2017 Cert. No.: 00026
Linearity File No.: 02412 Lin. Date: 01/31/2017 Lin. No.: 00025
Solution File No.: 02422 Soln. Date: 02/26/2017 Soln. No.: 00262
Sequential File No.: 02422 File Date: 02/26/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 1405

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:34S		
Control 1 EC	0.097%	00:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35S		
Control 2 EC	0.097%	00:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:36S		
Control 3 EC	0.097%	00:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KELLY

First Name: ROBERT

MI: J

Signature: _____



Badge No.: 3201

Date: 02/26/2017



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 682-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

CHRISTOPHER S. PORRINO
Acting Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 06/27/2016

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 16140

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1200 to 0.1212 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 07, 2018.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature of Ali M. Alaouie]

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 27th day of June, 2016.

[Handwritten signature of Mary E. McLaughlin]

MARY ELIZABETH MCLAUGHLIN
ID # 2052190
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"
New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable

