

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02938	Calib. Date: 06/29/2021
Certification File No.:	02939	Calib. No.: 00046
Linearity File No.:	02940	Cert. Date: 06/29/2021
Solution File No.:	02980	Cert. No.: 00036
Sequential File No.:	02980	Lin. Date: 06/29/2021
		Lin. No.: 00035
		Soln. Date: 11/16/2021
		Soln. No.: 00350
		File Date: 11/16/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1126

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:19S		
Control 1 EC	0.097%	19:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:20S		
Control 2 EC	0.096%	19:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:22S		
Control 3 EC	0.097%	19:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3205

Date: 11/16/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02938	Calib. Date: 06/29/2021
Certification File No.:	02939	Calib. No.: 00046
Linearity File No.:	02940	Cert. Date: 06/29/2021
Solution File No.:	02972	Cert. No.: 00036
Sequential File No.:	02972	Lin. Date: 06/29/2021
		Lin. No.: 00035
		Soln. Date: 10/24/2021
		Soln. No.: 00349
		File Date: 10/24/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0995

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:02D		
Control 1 EC	0.098%	20:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03D		
Control 2 EC	0.098%	20:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:05D		
Control 3 EC	0.098%	20:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3205

Date: 10/24/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02938	Calib. Date: 06/29/2021
Certification File No.:	02939	Calib. No.: 00046
Linearity File No.:	02940	Cert. Date: 06/29/2021
Solution File No.:	02967	Cert. No.: 00036
Sequential File No.:	02967	Lin. Date: 06/29/2021
		Lin. No.: 00035
		Soln. Date: 10/01/2021
		Soln. No.: 00348
		File Date: 10/01/2021

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0970

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:19D		
Control 1 EC	0.100%	20:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:20D		
Control 2 EC	0.098%	20:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:22D		
Control 3 EC	0.099%	20:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:23D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3208

Date: 10/01/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02938 Calib. Date: 06/29/2021 Calib. No.: 00046
Certification File No.: 02939 Cert. Date: 06/29/2021 Cert. No.: 00036
Linearity File No.: 02940 Lin. Date: 06/29/2021 Lin. No.: 00035
Solution File No.: 02959 Soln. Date: 09/05/2021 Soln. No.: 00347
Sequential File No.: 02959 File Date: 09/05/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 1091

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:58D		
Control 1 EC	0.098%	10:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 2 EC	0.097%	11:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		
Control 3 EC	0.098%	11:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 09/05/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02938 Calib. Date: 06/29/2021 Calib. No.: 00046
Certification File No.: 02939 Cert. Date: 06/29/2021 Cert. No.: 00036
Linearity File No.: 02940 Lin. Date: 06/29/2021 Lin. No.: 00035
Solution File No.: 02954 Soln. Date: 08/13/2021 Soln. No.: 00346
Sequential File No.: 02954 File Date: 08/13/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0881

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:35D		
Control 1 EC	0.097%	14:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37D		
Control 2 EC	0.097%	14:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		
Control 3 EC	0.097%	14:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 08/13/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02938	Calib. Date: 06/29/2021
Certification File No.:	02939	Calib. No.: 00046
Linearity File No.:	02940	Cert. Date: 06/29/2021
Solution File No.:	02947	Cert. No.: 00036
Sequential File No.:	02947	Lin. Date: 06/29/2021
		Lin. No.: 00035
		Soln. Date: 07/22/2021
		Soln. No.: 00345
		File Date: 07/22/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1036

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:16D		
Control 1 EC	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:17D		
Control 2 EC	0.099%	02:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:19D		
Control 3 EC	0.099%	02:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 07/22/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02893	Calib. Date:	01/21/2021
Certification File No.:	02894	Cert. No.:	00045
Linearity File No.:	02895	Cert. Date:	01/21/2021
Solution File No.:	02934	Lin. No.:	00035
Sequential File No.:	02934	Lin. Date:	01/21/2021
		Soln. No.:	00034
		Soln. Date:	06/18/2021
		File Date:	06/18/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0163
Control Solution %:	0.100%			Expires:	09/02/2022
Solution Control Lot:	20390			Bottle No.:	1041

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:55D		
Control 1 EC	0.102%	10:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		
Control 2 EC	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		
Control 3 EC	0.100%	10:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: _____

Badge No.: 3200

Date: 06/18/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02893	Calib. Date: 01/21/2021
Certification File No.:	02894	Calib. No.: 00045
Linearity File No.:	02895	Cert. Date: 01/21/2021
Solution File No.:	02926	Cert. No.: 00035
Sequential File No.:	02926	Lin. Date: 01/21/2021
		Lin. No.: 00034
		Soln. Date: 05/27/2021
		Soln. No.: 00342
		File Date: 05/27/2021

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1089

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:21D		
Control 1 EC	0.102%	06:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 2 EC	0.101%	06:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	06:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:24D		
Control 3 EC	0.101%	06:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	06:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 05/27/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02893 **Calib. Date:** 01/21/2021 **Calib. No.:** 00045
Certification File No.: 02894 **Cert. Date:** 01/21/2021 **Cert. No.:** 00035
Linearity File No.: 02895 **Lin. Date:** 01/21/2021 **Lin. No.:** 00034
Solution File No.: 02918 **Soln. Date:** 05/02/2021 **Soln. No.:** 00341
Sequential File No.: 02918 **File Date:** 05/02/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0163
Control Solution %: 0.100% **Expires:** 09/02/2022
Solution Control Lot: 20390 **Bottle No.:** 0908

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:02D		
Control 1 EC	0.103%	15:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03D		
Control 2 EC	0.101%	15:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05D		
Control 3 EC	0.101%	15:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 05/02/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02893	Calib. Date:	01/21/2021
Certification File No.:	02894	Cert. No.:	00045
Linearity File No.:	02895	Cert. Date:	01/21/2021
Solution File No.:	02914	Lin. No.:	00034
Sequential File No.:	02914	Soln. Date:	04/08/2021
		Soln. No.:	00340
		File Date:	04/08/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	0910

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:58D		
Control 1 EC	0.102%	10:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 2 EC	0.101%	11:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		
Control 3 EC	0.101%	11:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 04/08/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02893	Calib. Date: 01/21/2021
Certification File No.:	02894	Calib. No.: 00045
Linearity File No.:	02895	Cert. Date: 01/21/2021
Solution File No.:	02906	Cert. No.: 00035
Sequential File No.:	02906	Lin. Date: 01/21/2021
		Lin. No.: 00034
		Soln. Date: 03/11/2021
		Soln. No.: 00339
		File Date: 03/11/2021

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 09/02/2022
Solution Control Lot: 20390		Bottle No.: 0935

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:28S		
Control 1 EC	0.101%	18:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	18:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:29S		
Control 2 EC	0.100%	18:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	18:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:31S		
Control 3 EC	0.101%	18:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	18:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 03/11/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 02893 Calib. Date: 01/21/2021 Calib. No.: 00045
Certification File No.: 02894 Cert. Date: 01/21/2021 Cert. No.: 00035
Linearity File No.: 02895 Lin. Date: 01/21/2021 Lin. No.: 00034
Solution File No.: 02900 Soln. Date: 02/14/2021 Soln. No.: 00338
Sequential File No.: 02900 File Date: 02/14/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 1125

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:33S		
Control 1 EC	0.097%	13:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35S		
Control 2 EC	0.098%	13:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36S		
Control 3 EC	0.099%	13:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: 

Badge No.: 3208

Date: 02/14/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02839	Calib. Date: 07/22/2020
Certification File No.:	02840	Calib. No.: 00044
Linearity File No.:	02841	Cert. Date: 07/22/2020
Solution File No.:	02891	Cert. No.: 00034
Sequential File No.:	02891	Lin. Date: 07/22/2020
		Lin. No.: 00033
		Soln. Date: 01/09/2021
		Soln. No.: 00336
		File Date: 01/09/2021

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0158

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:10S		
Control 1 EC	0.098%	17:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	17:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:12S		
Control 2 EC	0.096%	17:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:13S		
Control 3 EC	0.097%	17:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3208

Date: 01/09/2021