

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	03137	Calib. Date:	08/01/2023
Certification File No.:	03138	Cert. Date:	08/01/2023
Linearity File No.:	03139	Lin. Date:	08/01/2023
Solution File No.:	03164	Soln. Date:	12/09/2023
Sequential File No.:	03164	File Date:	12/09/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0163
Control Solution %:	0.100%			Expires:	05/09/2024
Solution Control Lot:	22160			Bottle No.:	1204

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:42S		
Control 1 EC	0.100%	09:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43S		
Control 2 EC	0.099%	09:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		
Control 3 EC	0.100%	09:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	09:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JESS

First Name: VINCENT

MI: N

Signature: _____

Sat. VJ #3202

Badge No.: 3202

Date: 12/09/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03137	Calib. Date: 08/01/2023
Certification File No.:	03138	Cert. No.: 00051
Linearity File No.:	03139	Cert. Date: 08/01/2023
Solution File No.:	03158	Lin. No.: 00040
Sequential File No.:	03158	Soln. No.: 00384
		File Date: 11/17/2023

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 05/09/2024
Solution Control Lot: 22160		Bottle No.: 1417

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:20S		
Control 1 EC	0.099%	21:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:22S		
Control 2 EC	0.099%	21:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:23S		
Control 3 EC	0.099%	21:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3204

Date: 11/17/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03137	Calib. Date: 08/01/2023
Certification File No.:	03138	Calib. No.: 00051
Linearity File No.:	03139	Cert. Date: 08/01/2023
Solution File No.:	03153	Cert. No.: 00041
Sequential File No.:	03153	Lin. Date: 08/01/2023
		Lin. No.: 00040
		Soln. Date: 10/22/2023
		Soln. No.: 00383
		File Date: 10/22/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %:	0.100%		Expires: 05/09/2024
Solution Control Lot:	22160		Bottle No.: 1419

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:35D		
Control 1 EC	0.100%	13:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37D		
Control 2 EC	0.098%	13:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		
Control 3 EC	0.099%	13:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3204

Date: 10/22/2023

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03137 Calib. Date: 08/01/2023 Calib. No.: 00051
Certification File No.: 03138 Cert. Date: 08/01/2023 Cert. No.: 00041
Linearity File No.: 03139 Lin. Date: 08/01/2023 Lin. No.: 00040
Solution File No.: 03149 Soln. Date: 09/25/2023 Soln. No.: 00382
Sequential File No.: 03149 File Date: 09/25/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0966

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:40D		
Control 1 EC	0.099%	00:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:41D		
Control 2 EC	0.099%	00:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:43D		
Control 3 EC	0.099%	00:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3204

Date: 09/25/2023

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03137	Calib. Date: 08/01/2023
Certification File No.:	03138	Calib. No.: 00051
Linearity File No.:	03139	Cert. Date: 08/01/2023
Solution File No.:	03145	Cert. No.: 00041
Sequential File No.:	03145	Lin. Date: 08/01/2023
		Lin. No.: 00040
		Soln. Date: 08/30/2023
		Soln. No.: 00381
		File Date: 08/30/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0163
Control Solution %:	0.100%			Expires:	05/09/2024
Solution Control Lot:	22160			Bottle No.:	1295

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:01D		
Control 1 EC	0.099%	04:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:03D		
Control 2 EC	0.099%	04:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:04D		
Control 3 EC	0.099%	04:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	04:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 08/30/2023

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03107 Calib. Date: 03/14/2023 Calib. No.: 00050
Certification File No.: 03108 Cert. Date: 03/14/2023 Cert. No.: 00040
Linearity File No.: 03109 Lin. Date: 03/14/2023 Lin. No.: 00039
Solution File No.: 03136 Soln. Date: 06/26/2023 Soln. No.: 00379
Sequential File No.: 03136 File Date: 06/26/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0963

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:09D		
Control 1 EC	0.106%	12:10D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.103%	12:10D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.104%	12:11D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.105%	12:11D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.104%	12:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	12:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3230

Date: 06/26/2023

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03107 Calib. Date: 03/14/2023 Calib. No.: 00050
Certification File No.: 03108 Cert. Date: 03/14/2023 Cert. No.: 00040
Linearity File No.: 03109 Lin. Date: 03/14/2023 Lin. No.: 00039
Solution File No.: 03135 Soln. Date: 06/26/2023 Soln. No.: 00378
Sequential File No.: 03135 File Date: 06/26/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 1104

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:54D		
Control 1 EC	-.---%	10:55D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	10:55D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3230

Date: 06/26/2023

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C *Serial No.: ARNK-0039*
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03107 Calib. Date: 03/14/2023 Calib. No.: 00050
Certification File No.: 03108 Cert. Date: 03/14/2023 Cert. No.: 00040
Linearity File No.: 03109 Lin. Date: 03/14/2023 Lin. No.: 00039
Solution File No.: 03134 Soln. Date: 06/25/2023 Soln. No.: 00377
Sequential File No.: 03134 File Date: 06/25/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0955

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:39D		
Control 1 EC	0.106%	20:40D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.104%	20:40D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:41D		
Control 2 EC	0.104%	20:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	20:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:42D		
Control 3 EC	0.104%	20:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	20:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44D		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3230

Date: 06/25/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	03107	Calib. Date:	03/14/2023
Certification File No.:	03108	Cert. Date:	03/14/2023
Linearity File No.:	03109	Lin. Date:	03/14/2023
Solution File No.:	03133	Soln. Date:	06/24/2023
Sequential File No.:	03133	File Date:	06/24/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	22160	Expires:	05/09/2024
		Bottle No.:	0835

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:36D		
Control 1 EC	0.106%	23:36D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.103%	23:36D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:37D		
Control 2 EC	0.105%	23:38D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.104%	23:38D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:39D		
Control 3 EC	0.105%	23:39D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.103%	23:39D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:40D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JESS First Name: VINCENT MI: N

Signature: 

Badge No.: 3202
Date: 06/24/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03107	Calib. Date: 03/14/2023
Certification File No.:	03108	Calib. No.: 00050
Linearity File No.:	03109	Cert. Date: 03/14/2023
Solution File No.:	03132	Cert. No.: 00040
Sequential File No.:	03132	Lin. Date: 03/14/2023
		Lin. No.: 00039
		Soln. Date: 06/24/2023
		Soln. No.: 00375
		File Date: 06/24/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 0835

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:03D		
Control 1 EC	0.108%	22:03D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.102%	22:03D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	22:04D		
Control 2 EC	0.105%	22:05D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.103%	22:05D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	22:06D		
Control 3 EC	0.105%	22:06D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.103%	22:06D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	22:07D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JESS

First Name: VINCENT

MI: N

Signature: _____



Badge No.: 3202

Date: 06/24/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03107	Calib. Date: 03/14/2023
Certification File No.:	03108	Calib. No.: 00050
Linearity File No.:	03109	Cert. Date: 03/14/2023
Solution File No.:	03131	Cert. No.: 00040
Sequential File No.:	03131	Lin. Date: 03/14/2023
		Lin. No.: 00039
		Soln. Date: 06/24/2023
		Soln. No.: 00374
		File Date: 06/24/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 0833

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:21D		
Control 1 EC	0.107%	20:22D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.104%	20:22D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:22D		
Control 2 EC	0.106%	20:23D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.104%	20:23D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:24D		
Control 3 EC	0.105%	20:25D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.102%	20:25D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:25D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JESS

First Name: VINCENT

MI: N

Signature: _____



Badge No.: 3202

Date: 06/24/2023

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03107 Calib. Date: 03/14/2023 Calib. No.: 00050
Certification File No.: 03108 Cert. Date: 03/14/2023 Cert. No.: 00040
Linearity File No.: 03109 Lin. Date: 03/14/2023 Lin. No.: 00039
Solution File No.: 03129 Soln. Date: 06/05/2023 Soln. No.: 00373
Sequential File No.: 03129 File Date: 06/05/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0850

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:11D		
Control 1 EC	0.103%	02:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	02:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:13D		
Control 2 EC	0.102%	02:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	02:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:14D		
Control 3 EC	0.102%	02:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	02:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



3204

Badge No.: 3204

Date: 06/05/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03107	Calib. Date: 03/14/2023
Certification File No.:	03108	Calib. No.: 00050
Linearity File No.:	03109	Cert. Date: 03/14/2023
Solution File No.:	03119	Cert. No.: 00040
Sequential File No.:	03119	Lin. Date: 03/14/2023
		Lin. No.: 00039
		Soln. Date: 05/10/2023
		Soln. No.: 00372
		File Date: 05/10/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1376

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:09D		
Control 1 EC	0.104%	04:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	04:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:11D		
Control 2 EC	0.102%	04:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	04:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:12D		
Control 3 EC	0.102%	04:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	04:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 05/10/2023

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03107	Calib. Date: 03/14/2023
Certification File No.:	03108	Calib. No.: 00050
Linearity File No.:	03109	Cert. Date: 03/14/2023
Solution File No.:	03114	Cert. No.: 00040
Sequential File No.:	03114	Lin. Date: 03/14/2023
		Lin. No.: 00039
		Soln. Date: 04/13/2023
		Soln. No.: 00371
		File Date: 04/13/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUK S3-0133
Control Solution %:	0.100%		Expires: 05/09/2024
Solution Control Lot:	22160		Bottle No.: 1223

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:38D		
Control 1 EC	0.105%	20:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.105%	20:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40D		
Control 2 EC	0.104%	20:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.104%	20:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41D		
Control 3 EC	0.103%	20:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	20:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 04/13/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	03078	Calib. Date: 10/13/2022
Certification File No.:	03079	Cert. No.: 00049
Linearity File No.:	03080	Cert. Date: 10/13/2022
Solution File No.:	03104	Lin. No.: 00038
Sequential File No.:	03104	Soln. No.: 00369
		File Date: 02/14/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0163
Control Solution %:	0.100%			Expires:	05/09/2024
Solution Control Lot:	22160			Bottle No.:	1105

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:48S		
Control 1 EC	0.101%	22:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:50S		
Control 2 EC	0.100%	22:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:51S		
Control 3 EC	0.101%	22:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____



Badge No.: 3204

Date: 02/14/2023

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 03078 Calib. Date: 10/13/2022 Calib. No.: 00049
Certification File No.: 03079 Cert. Date: 10/13/2022 Cert. No.: 00039
Linearity File No.: 03080 Lin. Date: 10/13/2022 Lin. No.: 00038
Solution File No.: 03097 Soln. Date: 01/21/2023 Soln. No.: 00368
Sequential File No.: 03097 File Date: 01/21/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0848

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:07S		
Control 1 EC	0.102%	03:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:09S		
Control 2 EC	0.100%	03:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:10S		
Control 3 EC	0.101%	03:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: _____

Badge No.: 3204

Date: 01/21/2023