

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0024
Location:	BURLINGTON TOWNSHIP PD		
Calibration File No.:	00855	Calib. Date:	05/24/2013
Certification File No.:	00856	Cert. Date:	05/24/2013
Linearity File No.:	00857	Lin. Date:	05/24/2013
Solution File No.:	00895	Soln. Date:	11/01/2013
Sequential File No.:	00895	File Date:	11/01/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0166
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0984

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:51S		
Control 1 EC	0.100%	17:52S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:52S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52S		
Control 2 EC	0.099%	17:53S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:53S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:54S		
Control 3 EC	0.099%	17:54S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:54S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



84

Badge No.: 84
Date: 11/01/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00855	Calib. Date: 05/24/2013
Certification File No.:	00856	Calib. No.: 00020
Linearity File No.:	00857	Cert. Date: 05/24/2013
Solution File No.:	00891	Cert. No.: 00015
Sequential File No.:	00891	Lin. Date: 05/24/2013
		Lin. No.: 00015
		Soln. Date: 10/11/2013
		Soln. No.: 00133
		File Date: 10/11/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0166
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0977

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:45D		
Control 1 EC	0.099%	06:46D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:46D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		
Control 2 EC	0.099%	06:47D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:47D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:48D		
Control 3 EC	0.098%	06:49D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:49D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 10/11/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	Calib. No.:	00020
Calibration File No.:	00855	Cert. Date:	05/24/2013
Certification File No.:	00856	Lin. Date:	05/24/2013
Linearity File No.:	00857	Soln. Date:	09/12/2013
Solution File No.:	00883	File Date:	09/12/2013
Sequential File No.:	00883		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0166
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0985

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:24D		
Control 1 EC	0.099%	10:25D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:25D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25D		
Control 2 EC	0.099%	10:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27D		
Control 3 EC	0.099%	10:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		

All tests within acceptable tolerance.



On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____

Badge No.: 84

Date: 09/12/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00855 **Calib. Date:** 05/24/2013 **Calib. No.:** 00020
Cerification File No.: 00856 **Cert. Date:** 05/24/2013 **Cert. No.:** 00015
Linearity File No.: 00857 **Lin. Date:** 05/24/2013 **Lin. No.:** 00015
Solution File No.: 00876 **Soln. Date:** 08/16/2013 **Soln. No.:** 00131
Sequential File No.: 00876 **File Date:** 08/16/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0166
Control Solution %: 0.100% **Expires:** 12/26/2014
Solution Control Lot: 12L109 **Bottle No.:** 0978

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:53D		
Control 1 EC	0.100%	09:54D	34.1 °C	*** TEST PASSED ***
Control 1 IR	0.100%	09:54D	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54D		
Control 2 EC	0.099%	09:55D	34.1 °C	*** TEST PASSED ***
Control 2 IR	0.099%	09:55D	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 3 EC	0.099%	09:56D	34.1 °C	*** TEST PASSED ***
Control 3 IR	0.100%	09:56D	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SULLIVAN

First Name: JAMES

M1; B

Signature: 

Badge No.: 70

Date: 08/16/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00855 **Calib. Date:** 05/24/2013 **Calib. No.:** 00020
Certification File No.: 00856 **Cert. Date:** 05/24/2013 **Cert. No.:** 00015
Linearity File No.: 00857 **Lin. Date:** 05/24/2013 **Lin. No.:** 00015
Solution File No.: 00868 **Soln. Date:** 07/15/2013 **Soln. No.:** 00130
Sequential File No.: 00868 **File Date:** 07/15/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0166
Control Solution %: 0.100% **Expires:** 12/26/2014
Solution Control Lot: 12L109 **Bottle No.:** 0979

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:21D		
Control 1 EC	0.100%	04:21D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:21D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:22D		
Control 2 EC	0.099%	04:23D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:23D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23D		
Control 3 EC	0.100%	04:24D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:24D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 07/15/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00855 Calib. Date: 05/24/2013 Calib. No.: 00020
Certification File No.: 00856 Cert. Date: 05/24/2013 Cert. No.: 00015
Linearity File No.: 00857 Lin. Date: 05/24/2013 Lin. No.: 00015
Solution File No.: 00864 Soln. Date: 06/21/2013 Soln. No.: 00129
Sequential File No.: 00864 File Date: 06/21/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0166
Control Solution %: 0.100% Expires: 12/26/2014
Solution Control Lot: 12L109 Bottle No.: 0980

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:14D		
Control 1 EC	0.100%	06:15D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:15D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:16D		
Control 2 EC	0.099%	06:16D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:16D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:17D		
Control 3 EC	0.099%	06:17D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:17D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 06/21/2013

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00822	Calib. Date: 01/03/2013
Certification File No.:	00823	Calib. No.: 00019
Linearity File No.:	00824	Cert. Date: 01/03/2013
Solution File No.:	00848	Cert. No.: 00014
Sequential File No.:	00848	Lin. Date: 01/03/2013
		Lin. No.: 00014
		Soln. Date: 04/29/2013
		Soln. No.: 00127
		File Date: 04/29/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL 53-0166
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1085

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:33D		
Control 1 EC	0.101%	18:33D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:33D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:34D		
Control 2 EC	0.100%	18:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:35D		
Control 3 EC	0.101%	18:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LONG

First Name: DAVID

MI: A

Signature: _____

[Handwritten Signature] #77

Badge No.: 77

Date: 04/29/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00822	Calib. Date: 01/03/2013
Certification File No.:	00823	Cert. No.: 00019
Linearity File No.:	00824	Cert. Date: 01/03/2013
Solution File No.:	00838	Lin. No.: 00014
Sequential File No.:	00838	Soln. Date: 02/28/2013
		Soln. No.: 00125
		File Date: 02/28/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0166
Solution Control Lot:	11G093	Expires: 07/07/2013
		Bottle No.: 0871

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:16S		
Control 1 EC	0.101%	06:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:17S		
Control 2 EC	0.100%	06:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	06:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:19S		
Control 3 EC	0.101%	06:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____

Badge No.: 84

Date: 02/28/2013

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00822	Calib. Date: 01/03/2013
Certification File No.:	00823	Calib. No.: 00019
Linearity File No.:	00824	Cert. Date: 01/03/2013
Solution File No.:	00831	Cert. No.: 00014
Sequential File No.:	00831	Lin. Date: 01/03/2013
		Lin. No.: 00014
		Soln. Date: 02/01/2013
		Soln. No.: 00124
		File Date: 02/01/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0166
Control Solution %:	0.100%		Expires: 07/07/2013
Solution Control Lot:	11G093		Bottle No.: 0866

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:20S		
Control 1 EC	0.100%	15:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21S		
Control 2 EC	0.100%	15:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:23S		
Control 3 EC	0.100%	15:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____

Badge No.: 84

Date: 02/01/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00769	Calib. Date: 07/05/2012
Certification File No.:	00770	Cert. No.: 00018
Linearity File No.:	00771	Cert. Date: 07/05/2012
Solution File No.:	00820	Lin. No.: 00013
Sequential File No.:	00820	Lin. Date: 07/05/2012
		Soln. No.: 00122
		File Date: 01/02/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0167
Solution Control Lot:	11G093	Expires: 07/07/2013
		Bottle No.: 0872

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:06S		
Control 1 EC	0.098%	18:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07S		
Control 2 EC	0.098%	18:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:09S		
Control 3 EC	0.098%	18:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 01/02/2013