

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKII-C	Serial No.:	ARTL-0024
Location:	BURLINGTON TOWNSHIP PD		
Calibration File No.:	00971	Calib. Date:	09/10/2014
Certification File No.:	00972	Cert. Date:	09/10/2014
Linearity File No.:	00973	Lin. Date:	09/10/2014
Solution File No.:	00993	Soln. Date:	12/05/2014
Sequential File No.:	00993	File Date:	12/05/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0167
Solution Control Lot:	14H132	Expires:	08/22/2016
		Bottle No.:	0896

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:17S		
Control 1 EC	0.100%	17:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18S		
Control 2 EC	0.099%	17:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19S		
Control 3 EC	0.099%	17:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ORTIZ

First Name: JAVIER

MI: XX

Signature: _____



Badge No.: 81

Date: 12/05/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00971	Calib. Date: 09/10/2014
Certification File No.:	00972	Cert. No.: 00023
Linearity File No.:	00973	Cert. Date: 09/10/2014
Solution File No.:	00979	Lin. No.: 00018
Sequential File No.:	00979	Lin. Date: 09/10/2014
		Soln. No.: 00150
		Soln. Date: 10/05/2014
		File Date: 10/05/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0167
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0998

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:51D		
Control 1 EC	0.099%	05:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:53D		
Control 2 EC	0.100%	05:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:54D		
Control 3 EC	0.099%	05:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:55D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 10/05/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0024
Location:	BURLINGTON TOWNSHIP PD		
Calibration File No.:	00937	Calib. Date:	04/14/2014
Certification File No.:	00938	Cert. Date:	04/14/2014
Linearity File No.:	00939	Lin. Date:	04/14/2014
Solution File No.:	00967	Soln. Date:	07/28/2014
Sequential File No.:	00967	File Date:	07/28/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0167
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0992

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:20D		
Control 1 EC	0.099%	10:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22D		
Control 2 EC	0.099%	10:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23D		
Control 3 EC	0.098%	10:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 07/28/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00937	Calib. Date: 04/14/2014
Certification File No.:	00938	Calib. No.: 00022
Linearity File No.:	00939	Cert. Date: 04/14/2014
Solution File No.:	00965	Cert. No.: 00017
Sequential File No.:	00965	Lin. Date: 04/14/2014
		Lin. No.: 00017
		Soln. Date: 07/02/2014
		Soln. No.: 00145
		File Date: 07/02/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0167
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0993

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:19D		
Control 1 EC	0.099%	15:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		
Control 2 EC	0.099%	15:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		
Control 3 EC	0.099%	15:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 07/02/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00937 **Calib. Date:** 04/14/2014 **Calib. No.:** 00022
Certification File No.: 00938 **Cert. Date:** 04/14/2014 **Cert. No.:** 00017
Linearity File No.: 00939 **Lin. Date:** 04/14/2014 **Lin. No.:** 00017
Solution File No.: 00954 **Soln. Date:** 06/08/2014 **Soln. No.:** 00144
Sequential File No.: 00954 **File Date:** 06/08/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 12/26/2014
Solution Control Lot: 12L109 **Bottle No.:** 9940

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:19D		
Control 1 EC	0.099%	04:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20D		
Control 2 EC	0.099%	04:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21D		
Control 3 EC	0.099%	04:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROADWAY

First Name: JARROD

MI: D

Signature: 

Badge No.: 98

Date: 06/08/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00937	Calib. Date: 04/14/2014
Certification File No.:	00938	Calib. No.: 00022
Linearity File No.:	00939	Cert. Date: 04/14/2014
Solution File No.:	00952	Cert. No.: 00017
Sequential File No.:	00952	Lin. Date: 04/14/2014
		Lin. No.: 00017
		Soln. Date: 06/02/2014
		Soln. No.: 00143
		File Date: 06/02/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0167
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0995

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:51D		
Control 1 EC	0.099%	16:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52D		
Control 2 EC	0.099%	16:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:53D		
Control 3 EC	0.099%	16:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ORTIZ

First Name: JAVIER

MI: XX

Signature: _____



Badge No.: 81

Date: 06/02/2014

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00937 **Calib. Date:** 04/14/2014 **Calib. No.:** 00022
Certification File No.: 00938 **Cert. Date:** 04/14/2014 **Cert. No.:** 00017
Linearly File No.: 00939 **Lin. Date:** 04/14/2014 **Lin. No.:** 00017
Solution File No.: 00944 **Soln. Date:** 05/03/2014 **Soln. No.:** 00142
Sequential File No.: 00944 **File Date:** 05/03/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 12/26/2014
Solution Control Lot: 12L109 **Bottle No.:** 0986

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:33D		
Control 1 EC	0.099%	05:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:35D		
Control 2 EC	0.099%	05:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36D		
Control 3 EC	0.099%	05:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature:



Badge No.: 84

Date: 05/03/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00900	Calib. No.: 00021
Certification File No.:	00901	Cert. No.: 00016
Linearity File No.:	00902	Lin. No.: 00016
Solution File No.:	00933	Soln. No.: 00140
Sequential File No.:	00933	File Date: 04/06/2014

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0167
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0987

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:16D		
Control 1 EC	0.099%	04:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:17D		
Control 2 EC	0.099%	04:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		
Control 3 EC	0.100%	04:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LONG

First Name: DAVID

MI: A

Signature: _____



Badge No.: 77

Date: 04/06/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKII-C	Serial No.:	ARTL-0024
Location:	BURLINGTON TOWNSHIP PD		
Calibration File No.:	00900	Calib. Date:	11/20/2013
Certification File No.:	00901	Cert. Date:	11/20/2013
Linearity File No.:	00902	Lin. Date:	11/20/2013
Solution File No.:	00925	Soln. Date:	03/04/2014
Sequential File No.:	00925	File Date:	03/04/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0167
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0989

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:44S		
Control 1 EC	0.099%	10:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		
Control 2 EC	0.099%	10:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46S		
Control 3 EC	0.099%	10:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

M1: -

Signature: _____



Badge No.: 84

Date: 03/04/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00900 **Calib. Date:** 11/20/2013 **Calib. No.:** 00021
Certification File No.: 00901 **Cert. Date:** 11/20/2013 **Cert. No.:** 00016
Linearity File No.: 00902 **Lin. Date:** 11/20/2013 **Lin. No.:** 00016
Solution File No.: 00921 **Soln. Date:** 02/06/2014 **Soln. No.:** 00138
Sequential File No.: 00921 **File Date:** 02/06/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 12/26/2014
Solution Control Lot: 12L109 **Bottle No.:** 0981

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:55S		
Control 1 EC	0.099%	15:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57S		
Control 2 EC	0.099%	15:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58S		
Control 3 EC	0.099%	15:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 02/06/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0024
Location:	BURLINGTON TOWNSHIP PD	
Calibration File No.:	00900	Calib. Date: 11/20/2013
Certification File No.:	00901	Calib. No.: 00021
Linearity File No.:	00902	Cert. Date: 11/20/2013
Solution File No.:	00918	Cert. No.: 00016
Sequential File No.:	00918	Lin. Date: 11/20/2013
		Lin. No.: 00016
		Soln. Date: 01/11/2014
		Soln. No.: 00137
		File Date: 01/11/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0167
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0976

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:49S		
Control 1 EC	0.099%	12:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50S		
Control 2 EC	0.098%	12:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:51S		
Control 3 EC	0.098%	12:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BIRD

First Name: JOHN

MI: -

Signature: _____



Badge No.: 84

Date: 01/11/2014