

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|-------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARTL-0024 |
| Location: | BURLINGTON TOWNSHIP PD | Calib. No.: | 00025 |
| Calibration File No.: | 01038 | Cert. Date: | 09/22/2015 |
| Certification File No.: | 01039 | Lin. Date: | 09/22/2015 |
| Linearity File No.: | 01040 | Soln. Date: | 10/20/2015 |
| Solution File No.: | 01051 | File Date: | 10/20/2015 |
| Sequential File No.: | 01051 | | |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUL S3-0166 |
| Solution Control Lot: | 14H132 | Expires: | 08/22/2016 |
| | | Bottle No.: | 0884 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 05:58D | | |
| Control 1 EC | 0.099% | 05:59D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 05:59D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:59D | | |
| Control 2 EC | 0.099% | 06:00D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 06:00D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:01D | | |
| Control 3 EC | 0.099% | 06:01D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 06:01D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:02D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LONG

First Name: DAVID

MI: A

Signature:  # 77

Badge No.: 77

Date: 10/20/2015

Signature:  # 77

Badge No.: 77

Date: 10/20/2015

Calibrating Unit New Standard Solution Report

| | | |
|--------------------------------|------------------------|------------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0024 |
| Location: | BURLINGTON TOWNSHIP PD | |
| Calibration File No.: | 01008 | Calib. No.: 00024 |
| Certification File No.: | 01009 | Cert. No.: 00019 |
| Linearity File No.: | 01010 | Lin. No.: 00019 |
| Solution File No.: | 01030 | Soln. No.: 00162 |
| Sequential File No.: | 01030 | File Date: 07/10/2015 |


| | | |
|-------------------------------------|-------------------------|---------------------------------|
| Calibrating Unit: WET | Model No.: CU-34 | Serial No.: DDUL S3-0167 |
| Control Solution %: 0.100% | | Expires: 08/22/2016 |
| Solution Control Lot: 14H132 | | Bottle No.: 0891 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 15:15D | | |
| Control 1 EC | 0.100% | 15:16D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 15:16D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:16D | | |
| Control 2 EC | 0.100% | 15:17D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 15:17D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:18D | | |
| Control 3 EC | 0.100% | 15:18D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 15:18D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:19D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

| | | |
|--|---------------------|------------------|
| Last Name: ROBERSON | First Name: BRANDON | MI: C |
| Signature:  | | Badge No.: 55 |
| | | Date: 07/10/2015 |

| | | |
|--|--|------------------|
| Signature:  | | Badge No.: 55 |
| | | Date: 07/10/2015 |

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 01008 **Calib. Date:** 03/02/2015 **Calib. No.:** 00024
Certification File No.: 01009 **Cert. Date:** 03/02/2015 **Cert. No.:** 00019
Linearity File No.: 01010 **Lin. Date:** 03/02/2015 **Lin. No.:** 00019
Solution File No.: 01028 **Soln. Date:** 06/09/2015 **Soln. No.:** 00161
Sequential File No.: 01028 **File Date:** 06/09/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 08/22/2016
Solution Control Lot: 14H132 **Bottle No.:** 0889

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 01:39D | | |
| Control 1 EC | 0.101% | 01:40D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 01:40D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:40D | | |
| Control 2 EC | 0.100% | 01:41D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 01:41D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:42D | | |
| Control 3 EC | 0.101% | 01:42D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 01:42D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:43D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROBERSON

First Name: BRANDON

MI: C

Signature: Sgt B. Roberson

Badge No.: 55

Date: 06/09/2015

Last Name: ROBERSON

First Name: BRANDON

MI: C

Signature: Sgt B. Roberson

Badge No.: 55

Date: 06/09/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 01008 **Calib. Date:** 03/02/2015 **Calib. No.:** 00024
Certification File No.: 01009 **Cert. Date:** 03/02/2015 **Cert. No.:** 00019
Linearity File No.: 01010 **Lin. Date:** 03/02/2015 **Lin. No.:** 00019
Solution File No.: 01012 **Soln. Date:** 04/06/2015 **Soln. No.:** 00159
Sequential File No.: 01012 **File Date:** 04/06/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 08/22/2016
Solution Control Lot: 14H132 **Bottle No.:** 0893

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 02:04D | | |
| Control 1 EC | 0.101% | 02:04D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 02:04D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:05D | | |
| Control 2 EC | 0.100% | 02:06D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 02:06D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:07D | | |
| Control 3 EC | 0.100% | 02:07D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 02:07D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:08D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROBERSON

First Name: BRANDON

MI: C

Signature: Sgt B O Dr 55

Badge No.: 55

Date: 04/06/2015

LAST NAME: ROBERSON

FIRST NAME: BRANDON

MI: C

Signature: Sgt B O Dr 55

Badge No.: 55

Date: 04/06/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0024
Location: BURLINGTON TOWNSHIP PD
Calibration File No.: 00971 **Calib. Date:** 09/10/2014 **Calib. No.:** 00023
Certification File No.: 00972 **Cert. Date:** 09/10/2014 **Cert. No.:** 00018
Linearity File No.: 00973 **Lin. Date:** 09/10/2014 **Lin. No.:** 00018
Solution File No.: 01007 **Soln. Date:** 03/02/2015 **Soln. No.:** 00157
Sequential File No.: 01007 **File Date:** 03/02/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0167
Control Solution %: 0.100% **Expires:** 08/22/2016
Solution Control Lot: 14H132 **Bottle No.:** 0895


| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 06:48S | | |
| Control 1 EC | 0.100% | 06:49S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 06:49S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:49S | | |
| Control 2 EC | 0.099% | 06:50S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 06:50S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:51S | | |
| Control 3 EC | 0.099% | 06:52S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 06:52S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:53S | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LONG First Name: DAVID MI: A

Signature:  #77
Badge No.: 77
Date: 03/02/2015

Last Name: LONG First Name: DAVID MI: A

Signature:  #77
Badge No.: 77
Date: 03/02/2015

Calibrating Unit

New Standard Solution Report

| | | |
|--------------------------------------|--------------------------------|---------------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0024 |
| Location: | BURLINGTON TOWNSHIP PD | |
| Calibration File No.: 00971 | Calib. Date: 09/10/2014 | Calib. No.: 00023 |
| Certification File No.: 00972 | Cert. Date: 09/10/2014 | Cert. No.: 00018 |
| Linearity File No.: 00973 | Lin. Date: 09/10/2014 | Lin. No.: 00018 |
| Solution File No.: 01001 | Soln. Date: 02/06/2015 | Soln. No.: 00155 |
| Sequential File No.: 01001 | File Date: 02/06/2015 | |
| | | |
| Calibrating Unit: WET | Model No.: CU-34 | Serial No.: DDUL S3-0167 |
| Control Solution %: 0.100% | | Expires: 08/22/2016 |
| Solution Control Lot: 14H132 | | Bottle No.: 0898 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:01S | | |
| Control 1 EC | 0.098% | 10:01S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 10:01S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:02S | | |
| Control 2 EC | 0.098% | 10:03S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 10:03S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:04S | | |
| Control 3 EC | 0.099% | 10:04S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:04S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:05S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

| | | |
|---|-------------------|------------------|
| Last Name: LONG | First Name: DAVID | MI: A |
| Signature:  # 77 | | Badge No.: 77 |
| | | Date: 02/06/2015 |

| | | |
|---|-------------------|------------------|
| Last Name: LONG | First Name: DAVID | MI: A |
| Signature:  # 77 | | Badge No.: 77 |
| | | Date: 02/06/2015 |

Calibrating Unit

New Standard Solution Report


| | | |
|--------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0024 |
| Location: | BURLINGTON TOWNSHIP PD | |
| Calibration File No.: | 00971 | Calib. Date: 09/10/2014 |
| Certification File No.: | 00972 | Cert. No.: 00023 |
| Linearity File No.: | 00973 | Cert. Date: 09/10/2014 |
| Solution File No.: | 00997 | Lin. No.: 00018 |
| Sequential File No.: | 00997 | Soln. Date: 01/10/2015 |
| | | Soln. No.: 00154 |
| | | File Date: 01/10/2015 |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUL S3-0167 |
| Solution Control Lot: | 14H132 | Expires: 08/22/2016 |
| | | Bottle No.: 0899 |


| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 22:04S | | |
| Control 1 EC | 0.100% | 22:05S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 22:05S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:06S | | |
| Control 2 EC | 0.099% | 22:06S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 22:06S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:07S | | |
| Control 3 EC | 0.100% | 22:08S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 22:08S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:09S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

| | | |
|--|-------------------|------------------|
| Last Name: LONG | First Name: DAVID | MI: A |
| Signature:  | Badge No.: 77 | Date: 01/10/2015 |

| | | |
|--|-------------------|------------------|
| Last Name: LONG | First Name: DAVID | MI: A |
| Signature:  | Badge No.: 77 | Date: 01/10/2015 |