

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00794	Calib. Date: 09/15/2016
Certification File No.:	00795	Calib. No.: 00028
Linearity File No.:	00796	Cert. Date: 09/15/2016
Solution File No.:	00802	Cert. No.: 00022
Sequential File No.:	00802	Lin. Date: 09/15/2016
		Lin. No.: 00022
		Soln. Date: 12/31/2016
		Soln. No.: 00195
		File Date: 12/31/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 0538

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:14S		
Control 1 EC	0.100%	15:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16S		
Control 2 EC	0.099%	15:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17S		
Control 3 EC	0.099%	15:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 12/31/2016

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00794	Calib. Date: 09/15/2016
Certification File No.:	00795	Calib. No.: 00028
Linearity File No.:	00796	Cert. Date: 09/15/2016
Solution File No.:	00800	Cert. No.: 00022
Sequential File No.:	00800	Lin. Date: 09/15/2016
		Lin. No.: 00022
		Soln. Date: 11/30/2016
		Soln. No.: 00194
		File Date: 11/30/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1104

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:52S		
Control 1 EC	0.099%	20:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54S		
Control 2 EC	0.098%	20:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55S		
Control 3 EC	0.098%	20:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_



321

Badge No.: 321

Date: 11/30/2016

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00794	Calib. Date: 09/15/2016
Certification File No.:	00795	Calib. No.: 00028
Linearity File No.:	00796	Cert. Date: 09/15/2016
Solution File No.:	00798	Cert. No.: 00022
Sequential File No.:	00798	Lin. Date: 09/15/2016
		Lin. No.: 00022
		Soln. Date: 10/22/2016
		Soln. No.: 00193
		File Date: 10/22/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 0963

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:15D		
Control 1 EC	0.100%	16:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17D		
Control 2 EC	0.098%	16:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18D		
Control 3 EC	0.099%	16:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature:  321	Badge No.: 321	Date: 10/22/2016

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00779 Calib. Date: 04/08/2016 Calib. No.: 00027  
Certification File No.: 00780 Cert. Date: 04/08/2016 Cert. No.: 00021  
Linearity File No.: 00781 Lin. Date: 04/08/2016 Lin. No.: 00021  
Solution File No.: 00793 Soln. Date: 08/28/2016 Soln. No.: 00191  
Sequential File No.: 00793 File Date: 08/28/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0154  
Control Solution %: 0.100% Expires: 03/22/2018  
Solution Control Lot: 201603L Bottle No.: 0769

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:39D		
Control 1 EC	0.096%	03:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	03:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:40D		
Control 2 EC	0.095%	03:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	03:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:41D		
Control 3 EC	0.095%	03:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	03:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: SILVA

First Name: JAMES

MI: F

Signature: \_\_\_\_\_

*Sgt J Silva #319*

Badge No.: 319

Date: 08/28/2016

Signature: \_\_\_\_\_

*Sgt J Silva #319*

Badge No.: 319

Date: 08/28/2016

# Calibrating Unit New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00779	Calib. Date: 04/08/2016
Certification File No.:	00780	Calib. No.: 00027
Linearity File No.:	00781	Cert. Date: 04/08/2016
Solution File No.:	00791	Cert. No.: 00021
Sequential File No.:	00791	Lin. Date: 04/08/2016
		Lin. No.: 00021
		Soln. Date: 08/26/2016
		Soln. No.: 00190
		File Date: 08/26/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0770


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:41D		
Control 1 EC	0.099%	23:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:43D		
Control 2 EC	0.098%	23:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:44D		
Control 3 EC	0.098%	23:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature: <u> 321</u>	Badge No.: 321	Date: 08/26/2016

Signature: <u> 321</u>	Badge No.: 321 Date: 08/26/2016
---	------------------------------------

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE	Calib. Date:	04/08/2016
Calibration File No.:	00779	Calib. No.:	00027
Certification File No.:	00780	Cert. Date:	04/08/2016
Linearity File No.:	00781	Lin. Date:	04/08/2016
Solution File No.:	00789	Soln. Date:	08/09/2016
Sequential File No.:	00789	File Date:	08/09/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	201603L	Expires:	03/22/2018
		Bottle No.:	0771

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:22D		
Control 1 EC	0.098%	11:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 2 EC	0.097%	11:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 3 EC	0.097%	11:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINSOTTA

First Name: DAVID

MI: P

Signature:  321

Badge No.: 321

Date: 08/09/2016

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00779	Calib. Date:	04/08/2016
Certification File No.:	00780	Calib. No.:	00027
Linearity File No.:	00781	Cert. Date:	04/08/2016
Solution File No.:	00788	Cert. No.:	00021
Sequential File No.:	00788	Lin. Date:	04/08/2016
		Lin. No.:	00021
		Soln. Date:	07/09/2016
		Soln. No.:	00188
		File Date:	07/09/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0154
Control Solution %:	0.100%			Expires:	03/22/2018
Solution Control Lot:	201603L			Bottle No.:	0772

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:37D		
Control 1 EC	0.100%	02:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:38D		
Control 2 EC	0.099%	02:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D		
Control 3 EC	0.099%	02:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 07/09/2016

Solution Control Lot: 201603L

Bottle No.: 0772

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:37D		
Control 1 EC	0.100%	02:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:38D		
Control 2 EC	0.099%	02:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D		
Control 3 EC	0.099%	02:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:41D		

All tests within acceptable tolerance.

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00779	Calib. Date: 04/08/2016
Certification File No.:	00780	Calib. No.: 00027
Linearity File No.:	00781	Cert. Date: 04/08/2016
Solution File No.:	00786	Cert. No.: 00021
Sequential File No.:	00786	Lin. Date: 04/08/2016
		Lin. No.: 00021
		Soln. Date: 06/14/2016
		Soln. No.: 00187
		File Date: 06/14/2016

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %: 0.100%		Expires: 03/22/2018
Solution Control Lot: 201603L		Bottle No.: 0768

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:09D		
Control 1 EC	0.100%	22:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:11D		
Control 2 EC	0.099%	22:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12D		
Control 3 EC	0.099%	22:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINGOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

Badge No.: 321

Date: 06/14/2016

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:09D		
Control 1 EC	0.100%	22:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:11D		
Control 2 EC	0.099%	22:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12D		
Control 3 EC	0.099%	22:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13D		

All tests within acceptable tolerance.



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00779	Calib. Date: 04/08/2016
Certification File No.:	00780	Calib. No.: 00027
Linearity File No.:	00781	Cert. Date: 04/08/2016
Solution File No.:	00784	Cert. No.: 00021
Sequential File No.:	00784	Lin. Date: 04/08/2016
		Lin. No.: 00021
		Soln. Date: 05/11/2016
		Soln. No.: 00186
		File Date: 05/11/2016

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %: 0.100%		Expires: 03/22/2018
Solution Control Lot: 201603L		Bottle No.: 1442

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:10D		
Control 1 EC	0.101%	09:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		
Control 2 EC	0.100%	09:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13D		
Control 3 EC	0.100%	09:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15D		

A) tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

Badge No.: 321

Date: 05/11/2016

Solution Control Lot: 201603L

Bottle No.: 1442

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:10D		
Control 1 EC	0.101%	09:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		
Control 2 EC	0.100%	09:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13D		
Control 3 EC	0.100%	09:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15D		

All tests within acceptable tolerance.

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
**Location:** LONGPORT POLICE  
**Calibration File No.:** 00779 **Calib. Date:** 04/08/2016 **Calib. No.:** 00027  
**Certification File No.:** 00780 **Cert. Date:** 04/08/2016 **Cert. No.:** 00021  
**Linearity File No.:** 00781 **Lin. Date:** 04/08/2016 **Lin. No.:** 00021  
**Solution File No.:** 00782 **Soln. Date:** 04/08/2016 **Soln. No.:** 00185  
**Sequential File No.:** 00782 **File Date:** 04/08/2016

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDXC S3-0154  
**Control Solution %:** 0.100% **Expires:** 12/04/2016  
**Solution Control Lot:** 14L134 **Bottle No.:** 0168

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:46D		
Control 1 EC	0.100%	09:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		
Control 2 EC	0.099%	09:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49D		
Control 3 EC	0.100%	09:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔWAP2-177 (KA)

## Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I K W. #6704

Badge No.: 6704

Date: 04/08/2016

# Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00763	Calib. Date: 11/25/2015
Certification File No.:	00764	Calib. No.: 00026
Linearity File No.:	00765	Cert. Date: 11/25/2015
Solution File No.:	00774	Cert. No.: 00020
Sequential File No.:	00774	Lin. Date: 11/25/2015
		Lin. No.: 00020
		Soln. Date: 03/17/2016
		Soln. No.: 00184
		File Date: 03/17/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %:	0.100%		Expires: 12/31/2016
Solution Control Lot:	14L135		Bottle No.: 0913

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:16S		
Control 1 EC	0.099%	08:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		
Control 2 EC	0.098%	08:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19S		
Control 3 EC	0.098%	08:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BAUMGARDNER

First Name: JOSEPH

MI: T

Signature:

*Joseph T. Baumgardner #312*

Badge No.: 312

Date: 03/17/2016

3-17-2016

Solution Control Lot: 14L135

Bottle No.: 0913

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:16S		
Control 1 EC	0.099%	08:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		
Control 2 EC	0.098%	08:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19S		
Control 3 EC	0.098%	08:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20S		

All tests within acceptable tolerance.

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE	Calib. Date:	11/25/2015
Calibration File No.:	00763	Calib. No.:	00026
Certification File No.:	00764	Cert. Date:	11/25/2015
Linearity File No.:	00765	Cert. No.:	00020
Solution File No.:	00772	Lin. Date:	11/25/2015
Sequential File No.:	00772	Lin. No.:	00020
		Soln. Date:	02/18/2016
		Soln. No.:	00183
		File Date:	02/18/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	0914

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:12S		
Control 1 EC	0.101%	16:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14S		
Control 2 EC	0.100%	16:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15S		
Control 3 EC	0.100%	16:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BAUMGARDNER

First Name: JOSEPH

MI: T

Signature:

*Joseph T. Baumgardner #312*  
2/18/2016

Badge No.: 312

Date: 02/18/2016



State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

JOHN J. HOFFMAN  
 Acting Attorney General

KIM GUADAGNO  
 Lt. Governor

COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

**ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION:** Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

**MANUFACTURER:** Dräger Safety, Inc.

**ANALYSIS DATE:** 1/21/2015

**BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER:** 14L135

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1212 to 0.1219 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is December 31, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.  
 Research Scientist  
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 22nd day of January, 2015.

Notary  
**MARY ELIZABETH MCLAUGHLIN**  
 ID # 2062190  
 NOTARY PUBLIC  
 STATE OF NEW JERSEY  
 My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable





State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 Post Office Box 7068  
 West Trenton, NJ 08628-0068  
 (609) 593-2000

CHRIS CHRISTIE  
 Governor

KIM GUARDINO  
 Lt. Governor

JOHN J. HOFFMAN  
 Acting Attorney General

COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

**ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION:** Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

**MANUFACTURER:** Intoxecr Safety, Inc.

**ANALYSIS DATE:** 1/21/2015

**BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER:** 146136

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1212 to 0.1218 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is December 31, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alsouts, Ph.D.  
 Research Scientist  
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me on this 23rd day of JANUARY, 2015.

Notary  
**MARY ELIZABETH MCLAUGHLIN**  
 ID # 3088190  
 NOTARY PUBLIC  
 STATE OF NEW JERSEY  
 My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"

Only Jersey State Police Forensic Laboratory  
 Printed on Recycled Paper and Recyclable



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKJII-C	Serial No.: ARXA-0067
<b>Location:</b>	LONGPORT POLICE	
<b>Calibration File No.:</b>	00763	<b>Calib. Date:</b> 11/25/2015
<b>Certification File No.:</b>	00764	<b>Calib. No.:</b> 00026
<b>Linearity File No.:</b>	00765	<b>Cert. Date:</b> 11/25/2015
<b>Solution File No.:</b>	00771	<b>Cert. No.:</b> 00020
<b>Sequential File No.:</b>	00771	<b>Lin. Date:</b> 11/25/2015
		<b>Lin. No.:</b> 00020
		<b>Soln. Date:</b> 01/21/2016
		<b>Soln. No.:</b> 00182
		<b>File Date:</b> 01/21/2016

<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34	<b>Serial No.:</b> DDXC S3-0154
<b>Control Solution %:</b>	0.100%		<b>Expires:</b> 06/13/2017
<b>Solution Control Lot:</b>	15F139		<b>Bottle No.:</b> 0221

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:48S		
Control 1 EC	0.097%	17:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	17:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:49S		
Control 2 EC	0.096%	17:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	17:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51S		
Control 3 EC	0.096%	17:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	17:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SILVA

First Name: JAMES

MJ: F

Signature: \_\_\_\_\_



Badge No.: 319

Date: 01/21/2016

**Solution Control Lot:** 15F139

**Bottle No.:** 0221

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:48S		
Control 1 EC	0.097%	17:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	17:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:49S		
Control 2 EC	0.096%	17:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	17:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51S		
Control 3 EC	0.096%	17:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	17:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52S		

All tests within acceptable tolerance.