

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE	Calib. Date:	10/20/2017
Calibration File No.:	00827	Calib. No.:	00031
Certification File No.:	00828	Cert. Date:	10/20/2017
Linearity File No.:	00829	Lin. Date:	10/20/2017
Solution File No.:	00832	Soln. Date:	12/24/2017
Sequential File No.:	00832	File Date:	12/24/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	1316

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:57S		
Control 1 EC	0.097%	05:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:58S		
Control 2 EC	0.097%	05:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:59S		
Control 3 EC	0.097%	06:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SILVA

First Name: JAMES

MI: F

Signature: \_\_\_\_\_



Badge No.: 319

Date: 12/24/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00827	Calib. Date: 10/20/2017
Certification File No.:	00828	Calib. No.: 00031
Linearity File No.:	00829	Cert. Date: 10/20/2017
Solution File No.:	00831	Cert. No.: 00025
Sequential File No.:	00831	Lin. Date: 10/20/2017
		Lin. No.: 00025
		Soln. Date: 11/18/2017
		Soln. No.: 00206
		File Date: 11/18/2017

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0135
Control Solution %: 0.100%		Expires: 01/30/2019
Solution Control Lot: 17050		Bottle No.: 1198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:11S		
Control 1 EC	0.100%	08:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12S		
Control 2 EC	0.099%	08:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14S		
Control 3 EC	0.099%	08:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 11/18/2017

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00815	Calib. Date: 06/05/2017
Certification File No.:	00816	Calib. No.: 00030
Linearity File No.:	00817	Cert. Date: 06/05/2017
Solution File No.:	00826	Cert. No.: 00024
Sequential File No.:	00826	Lin. Date: 06/05/2017
		Lin. No.: 00024
		Soln. Date: 09/14/2017
		Soln. No.: 00204
		File Date: 09/14/2017

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %: 0.100%		Expires: 01/30/2019
Solution Control Lot: 17050		Bottle No.: 1214

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:31D		
Control 1 EC	0.100%	18:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:32D		
Control 2 EC	0.098%	18:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:33D		
Control 3 EC	0.099%	18:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 09/14/2017

# Calibrating Unit New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00815	Calib. Date:	06/05/2017
Certification File No.:	00816	Cert. No.:	00030
Linearity File No.:	00817	Lin. Date:	06/05/2017
Solution File No.:	00823	Lin. No.:	00024
Sequential File No.:	00823	Soln. Date:	08/06/2017
		File Date:	08/06/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	.0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	1322

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:14D		
Control 1 EC	0.099%	16:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15D		
Control 2 EC	0.098%	16:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17D		
Control 3 EC	0.098%	16:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SILVA	First Name: JAMES	MI: F
Signature: <u></u>	Badge No.: 319	Date: 08/06/2017



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00815	Calib. Date: 06/05/2017
Certification File No.:	00816	Calib. No.: 00030
Linearity File No.:	00817	Cert. Date: 06/05/2017
Solution File No.:	00821	Cert. No.: 00024
Sequential File No.:	00821	Lin. Date: 06/05/2017
		Lin. No.: 00024
		Soln. Date: 07/06/2017
		Soln. No.: 00202
		File Date: 07/06/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1203

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:11D		
Control 1 EC	0.098%	01:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13D		
Control 2 EC	0.097%	01:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:14D		
Control 3 EC	0.098%	01:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SILVA

First Name: JAMES

MI: F

Signature: \_\_\_\_\_

Sgt James 319

Badge No.: 319

Date: 07/06/2017

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00804	Calib. Date: 01/30/2017
Certification File No.:	00805	Calib. No.: 00029
Linearity File No.:	00806	Cert. Date: 01/30/2017
Solution File No.:	00814	Cert. No.: 00023
Sequential File No.:	00814	Lin. Date: 01/30/2017
		Lin. No.: 00023
		Soln. Date: 05/11/2017
		Soln. No.: 00200
		File Date: 05/11/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1052

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:06D		
Control 1 EC	0.099%	02:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08D		
Control 2 EC	0.098%	02:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09D		
Control 3 EC	0.098%	02:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta #321*

Badge No.: 321

Date: 05/11/2017

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00804	Calib. Date: 01/30/2017
Certification File No.:	00805	Calib. No.: 00029
Linearity File No.:	00806	Cert. Date: 01/30/2017
Solution File No.:	00812	Cert. No.: 00023
Sequential File No.:	00812	Lin. Date: 01/30/2017
		Lin. No.: 00023
		Soln. Date: 04/11/2017
		Soln. No.: 00199
		File Date: 04/11/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1215

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:44D		
Control 1 EC	0.097%	10:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 2 EC	0.097%	10:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		
Control 3 EC	0.097%	10:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 04/11/2017

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00804	Calib. Date: 01/30/2017
Certification File No.:	00805	Calib. No.: 00029
Linearity File No.:	00806	Cert. Date: 01/30/2017
Solution File No.:	00811	Cert. No.: 00023
Sequential File No.:	00811	Lin. Date: 01/30/2017
		Lin. No.: 00023
		Soln. Date: 03/12/2017
		Soln. No.: 00198
		File Date: 03/12/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1074

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:08S		
Control 1 EC	0.097%	22:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:10S		
Control 2 EC	0.096%	22:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:11S		
Control 3 EC	0.096%	22:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SILVA

First Name: JAMES

MI: F

Signature: \_\_\_\_\_



Badge No.: 319

Date: 03/12/2017



# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00804	Calib. Date: 01/30/2017
Certification File No.:	00805	Calib. No.: 00029
Linearity File No.:	00806	Cert. Date: 01/30/2017
Solution File No.:	00809	Cert. No.: 00023
Sequential File No.:	00809	Lin. Date: 01/30/2017
		Lin. No.: 00023
		Soln. Date: 02/14/2017
		Soln. No.: 00197
		File Date: 02/14/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 0960

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:51S		
Control 1 EC	0.097%	23:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	23:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:52S		
Control 2 EC	0.096%	23:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:54S		
Control 3 EC	0.096%	23:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	23:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINGOTTA	First Name: DAVID	MI: P
Signature: <u> 321</u>	Badge No.: 321	Date: 02/14/2017